

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q2473 SAS No.: Q2473 SDG NO.: Q2473

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2025 07/01/2025

Client Sample No.: CCAL02 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP073463.D Time Analyzed: 18:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.640	5.544	5.744	511.530	500.000	2.3
Aroclor-1016-2	5.662	5.566	5.766	527.050	500.000	5.4
Aroclor-1016-3	5.724	5.629	5.829	489.140	500.000	-2.2
Aroclor-1016-4	5.822	5.726	5.926	517.580	500.000	3.5
Aroclor-1016-5	6.114	6.018	6.218	501.770	500.000	0.4
Aroclor-1260-1	7.231	7.134	7.334	498.660	500.000	-0.3
Aroclor-1260-2	7.484	7.389	7.589	488.000	500.000	-2.4
Aroclor-1260-3	7.843	7.746	7.946	468.290	500.000	-6.3
Aroclor-1260-4	8.066	7.971	8.171	450.520	500.000	-9.9
Aroclor-1260-5	8.384	8.289	8.489	473.490	500.000	-5.3
Decachlorobiphenyl	10.181	10.086	10.286	48.200	50.000	-3.6
Tetrachloro-m-xylene	4.490	4.393	4.593	51.140	50.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q2473 SAS No.: Q2473 SDG NO.: Q2473

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2025 07/01/2025

Client Sample No.: CCAL02 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP073463.D Time Analyzed: 18:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.862	4.764	4.964	559.790	500.000	12.0
Aroclor-1016-2	4.880	4.782	4.982	561.770	500.000	12.4
Aroclor-1016-3	5.056	4.958	5.158	573.400	500.000	14.7
Aroclor-1016-4	5.098	5.000	5.200	575.300	500.000	15.1
Aroclor-1016-5	5.311	5.214	5.414	583.620	500.000	16.7
Aroclor-1260-1	6.342	6.244	6.444	528.760	500.000	5.8
Aroclor-1260-2	6.531	6.433	6.633	492.230	500.000	-1.6
Aroclor-1260-3	6.682	6.585	6.785	538.520	500.000	7.7
Aroclor-1260-4	7.152	7.054	7.254	518.690	500.000	3.7
Aroclor-1260-5	7.394	7.297	7.497	512.140	500.000	2.4
Decachlorobiphenyl	8.788	8.690	8.890	54.400	50.000	8.8
Tetrachloro-m-xylene	3.782	3.685	3.885	54.540	50.000	9.1