

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073411.D	01 Jul 2025 13:30		YP\AJ	Ok
2	I.BLK	I.BLK	PP073412.D	01 Jul 2025 13:46		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP073413.D	01 Jul 2025 14:04		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP073414.D	01 Jul 2025 14:21		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP073415.D	01 Jul 2025 14:37		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP073416.D	01 Jul 2025 14:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP073417.D	01 Jul 2025 15:10		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP073418.D	01 Jul 2025 15:26		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP073419.D	01 Jul 2025 15:43		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP073420.D	01 Jul 2025 15:59		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP073421.D	01 Jul 2025 16:16		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP073422.D	01 Jul 2025 16:33		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP073423.D	01 Jul 2025 16:49		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP073424.D	01 Jul 2025 17:05		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP073425.D	01 Jul 2025 17:22		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP073426.D	01 Jul 2025 17:39		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP073427.D	01 Jul 2025 17:55		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP073428.D	01 Jul 2025 18:12		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP073429.D	01 Jul 2025 18:28		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP073430.D	01 Jul 2025 18:45		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP073431.D	01 Jul 2025 19:01		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PP073432.D	01 Jul 2025 19:18		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP073433.D	01 Jul 2025 19:35		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP073434.D	01 Jul 2025 19:51		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP073435.D	01 Jul 2025 20:08		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP073436.D	01 Jul 2025 20:24		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP073437.D	01 Jul 2025 20:41		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP073438.D	01 Jul 2025 20:57		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP073439.D	01 Jul 2025 21:14		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP073440.D	01 Jul 2025 21:30		YPIAJ	Ok,M
31	PP070125ICV500	ICVPP070125	PP073441.D	01 Jul 2025 21:47		YPIAJ	Ok
32	AR1242ICV500	ICVPP070125AR1242	PP073442.D	01 Jul 2025 22:20		YPIAJ	Ok
33	AR1248ICV500	ICVPP070125AR1248	PP073443.D	01 Jul 2025 22:53		YPIAJ	Ok
34	AR1254ICV500	ICVPP070125AR1254	PP073444.D	01 Jul 2025 23:26		YPIAJ	Ok
35	AR1268ICV500	ICVPP070125AR1268	PP073445.D	02 Jul 2025 00:00		YPIAJ	Ok
36	DDT ANALOGUE	DDT ANALOGUE	PP073446.D	02 Jul 2025 07:34		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By		Supervise On	
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073447.D	02 Jul 2025 08:57		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP073448.D	02 Jul 2025 10:07		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP073449.D	02 Jul 2025 10:27		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP073450.D	02 Jul 2025 10:44		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP073451.D	02 Jul 2025 11:00		YP\AJ	Ok
6	I.BLK	I.BLK	PP073452.D	02 Jul 2025 11:17		YP\AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PP073453.D	02 Jul 2025 11:34		YP\AJ	Ok
8	PB168692BL	PB168692BL	PP073454.D	02 Jul 2025 15:14		YP\AJ	Ok
9	PB168692BS	PB168692BS	PP073455.D	02 Jul 2025 15:30		YP\AJ	Ok
10	Q2472-01	PN#1	PP073456.D	02 Jul 2025 15:46		YP\AJ	Ok
11	Q2473-01	PIT#1	PP073457.D	02 Jul 2025 16:03		YP\AJ	Ok,NS
12	Q2473-02	PIT#2	PP073458.D	02 Jul 2025 16:19		YP\AJ	Ok,NS
13	Q2473-03	PIT#3	PP073459.D	02 Jul 2025 16:36		YP\AJ	Ok,NS
14	Q2473-04	PIT#4	PP073460.D	02 Jul 2025 16:53		YP\AJ	Ok,NS
15	Q2473-04MS	PIT#4MS	PP073461.D	02 Jul 2025 17:09		YP\AJ	Ok,NS
16	Q2473-04MSD	PIT#4MSD	PP073462.D	02 Jul 2025 17:25		YP\AJ	Ok,NS
17	AR1660CCC500	AR1660CCC500	PP073463.D	02 Jul 2025 18:47		YP\AJ	Ok,NS
18	AR1242CCC500	AR1242CCC500	PP073464.D	02 Jul 2025 19:04		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By		Supervise On	
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP073465.D	02 Jul 2025 19:20		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP073466.D	02 Jul 2025 19:36		YPIAJ	Ok
21	I.BLK	I.BLK	PP073467.D	02 Jul 2025 19:53		YPIAJ	Ok
22	Q2484-01	TP-58	PP073468.D	02 Jul 2025 20:09		YPIAJ	Ok,NS
23	Q2484-02	TP-57	PP073469.D	02 Jul 2025 20:26		YPIAJ	Ok,NS
24	Q2484-03	TP-64	PP073470.D	02 Jul 2025 20:42		YPIAJ	Ok
25	Q2484-04	TP-107	PP073471.D	02 Jul 2025 20:58		YPIAJ	Ok
26	Q2484-05	TP-1006	PP073472.D	02 Jul 2025 21:15		YPIAJ	Ok
27	Q2484-06	TP-104	PP073473.D	02 Jul 2025 21:31		YPIAJ	Ok
28	Q2458-10	FB-06272025	PP073474.D	02 Jul 2025 21:47		YPIAJ	Ok
29	Q2477-01	50728	PP073475.D	02 Jul 2025 22:04		YPIAJ	Ok,NS
30	Q2481-01	CC0627-AL	PP073476.D	02 Jul 2025 22:20		YPIAJ	Ok,NS
31	Q2481-04	CC0627-AOXL	PP073477.D	02 Jul 2025 22:36		YPIAJ	Ok,NS
32	AR1660CCC500	AR1660CCC500	PP073478.D	02 Jul 2025 23:58		YPIAJ	Ok,NS
33	AR1242CCC500	AR1242CCC500	PP073479.D	03 Jul 2025 00:15	failing low in 1st column	YPIAJ	Not Ok
34	AR1248CCC500	AR1248CCC500	PP073480.D	03 Jul 2025 00:31		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PP073481.D	03 Jul 2025 00:47	failing low in 1st column	YPIAJ	Not Ok
36	I.BLK	I.BLK	PP073482.D	03 Jul 2025 01:04		YPIAJ	Ok

M : Manual Integration