

Surrogate Summary

SDG No.: Q2473

Client: JPCL Engineering

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD088990.D	PIBLK-PD088990.D	Decachlorobiphen	1	20	18.2	91		57	171
		Tetrachloro-m-xyl	1	20	16.1	80		61	148
		Decachlorobiphen	2	20	18.6	93		57	171
		Tetrachloro-m-xyl	2	20	17.7	88		61	148
I.BLK-PD089301.D	PIBLK-PD089301.D	Decachlorobiphen	1	20	18.9	95		57	171
		Tetrachloro-m-xyl	1	20	18.4	92		61	148
		Decachlorobiphen	2	20	18.8	94		57	171
		Tetrachloro-m-xyl	2	20	20.7	103		61	148
PB168693BL	PB168693BL	Decachlorobiphen	1	20	17.6	88		20	144
		Tetrachloro-m-xyl	1	20	16.9	85		19	148
		Decachlorobiphen	2	20	17.8	89		20	144
		Tetrachloro-m-xyl	2	20	18.3	92		19	148
PB168693BS	PB168693BS	Decachlorobiphen	1	20	20.6	103		20	144
		Tetrachloro-m-xyl	1	20	19.0	95		19	148
		Decachlorobiphen	2	20	21.1	105		20	144
		Tetrachloro-m-xyl	2	20	22.7	113		19	148
Q2475-01MS	SOIL-PILEMS	Decachlorobiphen	1	20	20.2	101		20	144
		Tetrachloro-m-xyl	1	20	13.5	68		19	148
		Decachlorobiphen	2	20	16.2	81		20	144
		Tetrachloro-m-xyl	2	20	15.3	77		19	148
Q2475-01MSD	SOIL-PILEMSD	Decachlorobiphen	1	20	20.1	101		20	144
		Tetrachloro-m-xyl	1	20	13.5	68		19	148
		Decachlorobiphen	2	20	16.0	80		20	144
		Tetrachloro-m-xyl	2	20	15.1	75		19	148
I.BLK-PD089310.D	PIBLK-PD089310.D	Decachlorobiphen	1	20	18.3	91		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	18.8	94		57	171
		Tetrachloro-m-xyl	2	20	20.6	103		61	148
Q2473-01	PIT#1	Decachlorobiphen	1	20	15.3	77		20	144
		Tetrachloro-m-xyl	1	20	12.8	64		19	148
		Decachlorobiphen	2	20	12.2	61		20	144
		Tetrachloro-m-xyl	2	20	15.1	75		19	148
Q2473-02	PIT#2	Decachlorobiphen	1	20	13.6	68		20	144
		Tetrachloro-m-xyl	1	20	12.9	65		19	148
		Decachlorobiphen	2	20	11.2	56		20	144
		Tetrachloro-m-xyl	2	20	14.5	72		19	148
Q2473-03	PIT#3	Decachlorobiphen	1	20	12.9	64		20	144
		Tetrachloro-m-xyl	1	20	12.2	61		19	148
		Decachlorobiphen	2	20	8.19	41		20	144
		Tetrachloro-m-xyl	2	20	14.4	72		19	148
Q2473-04	PIT#4	Decachlorobiphen	1	20	12.1	61		20	144

Surrogate Summary

SDG No.: Q2473

Client: JPCL Engineering

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2473-04	PIT#4	Tetrachloro-m-xyl	1	20	12.9	65		19	148
		Decachlorobiphen	2	20	9.36	47		20	144
		Tetrachloro-m-xyl	2	20	14.8	74		19	148
I.BLK-PD089323.D	PIBLK-PD089323.D	Decachlorobiphen	1	20	16.5	82		57	171
		Tetrachloro-m-xyl	1	20	19.0	95		61	148
		Decachlorobiphen	2	20	13.2	66		57	171
		Tetrachloro-m-xyl	2	20	21.3	107		61	148