

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: JPCL01
ProjectID: NOHO RFI No. SC64025 NYC
Lab Code: ACE SDG No.: Q2473
DataFile: FF016145.D Analyst Name: YP\AJ Analyst Date: 07-10-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	99480292	117036	106641	9.748