

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/02/2025 11:41</u>
Lab File ID:	<u>BF142964.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.175		-2.2	
Benzaldehyde	0.841	0.869		3.3	
Phenol-d6	1.417	1.364		-3.7	
Phenol	1.575	1.528		-3.0	20.0
bis(2-Chloroethyl)ether	1.126	1.072		-4.8	
2-Chlorophenol	1.296	1.252		-3.4	
2-Methylphenol	0.982	0.951		-3.2	
2,2-oxybis(1-Chloropropane)	1.809	1.643		-9.2	
Acetophenone	0.441	0.428		-2.9	
3+4-Methylphenols	1.224	1.220		-0.3	
n-Nitroso-di-n-propylamine	0.824	0.806	0.050	-2.2	
Nitrobenzene-d5	0.365	0.381		4.4	
Hexachloroethane	0.511	0.505		-1.2	
Nitrobenzene	0.330	0.320		-3.0	
Isophorone	0.585	0.565		-3.4	
2-Nitrophenol	0.180	0.179		-0.6	20.0
2,4-Dimethylphenol	0.311	0.301		-3.2	
bis(2-Chloroethoxy)methane	0.372	0.360		-3.2	
2,4-Dichlorophenol	0.282	0.275		-2.5	20.0
Naphthalene	0.979	0.947		-3.3	
4-Chloroaniline	0.398	0.390		-2.0	
Hexachlorobutadiene	0.190	0.186		-2.1	20.0
Caprolactam	0.075	0.075		0.0	
4-Chloro-3-methylphenol	0.281	0.275		-2.1	20.0
2-Methylnaphthalene	0.607	0.601		-1.0	
Hexachlorocyclopentadiene	0.335	0.265	0.050	-20.9	
2,4,6-Trichlorophenol	0.385	0.371		-3.6	20.0
2-Fluorobiphenyl	1.522	1.598		5.0	
2,4,5-Trichlorophenol	0.405	0.391		-3.5	
1,1-Biphenyl	1.580	1.544		-2.3	
2-Chloronaphthalene	1.186	1.140		-3.9	
2-Nitroaniline	0.332	0.326		-1.8	
Dimethylphthalate	1.295	1.269		-2.0	
Acenaphthylene	1.952	1.901		-2.6	
2,6-Dinitrotoluene	0.284	0.284		0.0	
3-Nitroaniline	0.313	0.313		0.0	
Acenaphthene	1.191	1.162		-2.4	20.0
2,4-Dinitrophenol	0.142	0.122	0.050	-14.1	
4-Nitrophenol	0.214	0.246	0.050	15.0	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/02/2025 11:41</u>
Lab File ID:	<u>BF142964.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.708	1.651		-3.3	
2,4-Dinitrotoluene	0.378	0.373		-1.3	
Diethylphthalate	1.269	1.281		0.9	
4-Chlorophenyl-phenylether	0.636	0.631		-0.8	
Fluorene	1.334	1.313		-1.6	
4-Nitroaniline	0.286	0.291		1.7	
4,6-Dinitro-2-methylphenol	0.121	0.113		-6.6	
n-Nitrosodiphenylamine	0.693	0.673		-2.9	20.0
2,4,6-Tribromophenol	0.206	0.203		-1.5	
4-Bromophenyl-phenylether	0.228	0.223		-2.2	
Hexachlorobenzene	0.253	0.240		-5.1	
Atrazine	0.179	0.188		5.0	
Pentachlorophenol	0.126	0.112		-11.1	20.0
Phenanthrene	1.083	1.038		-4.2	
Anthracene	1.111	1.071		-3.6	
Carbazole	0.959	0.921		-4.0	
Di-n-butylphthalate	0.999	1.051		5.2	
Fluoranthene	1.028	0.992		-3.5	20.0
Pyrene	1.875	1.887		0.6	
Terphenyl-d14	1.447	1.366		-5.6	
Butylbenzylphthalate	0.544	0.594		9.2	
3,3-Dichlorobenzidine	0.438	0.424		-3.2	
Benzo (a) anthracene	1.349	1.348		-0.1	
Chrysene	1.204	1.123		-6.7	
Bis(2-ethylhexyl)phthalate	0.782	0.847		8.3	
Di-n-octyl phthalate	1.475	1.461		-0.9	20.0
Benzo (b) fluoranthene	1.157	1.268		9.6	
Benzo (k) fluoranthene	1.083	1.006		-7.1	
Benzo (a) pyrene	1.118	1.116		-0.2	20.0
Indeno (1,2,3-cd) pyrene	1.523	1.408		-7.6	
Dibenzo (a,h) anthracene	1.238	1.141		-7.8	
Benzo (g,h,i) perylene	1.234	1.137		-7.9	
1,2,4,5-Tetrachlorobenzene	0.590	0.588		-0.3	
1,4-Dioxane	0.480	0.432		-10.0	20.0
2,3,4,6-Tetrachlorophenol	0.326	0.315		-3.4	

All other compounds must meet a minimum RRF of 0.010.