

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2473

Client ID : SSTDCCC040

Date Analyzed: 07/08/2025

Lab File ID: BF143025.D

Time Analyzed: 10:54

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	66813	6.875	257725	8.16	136330	9.92
UPPER LIMIT	133626	7.375	515450	8.663	272660	10.422
LOWER LIMIT	33406.5	6.375	128863	7.663	68165	9.422
EPA SAMPLE NO.						
01 PIT#1DL	55588	6.87	211740	8.16	106126	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.