

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/02/2025 10:37</u>
Lab File ID:	<u>VY022903.D</u>	Init. Calib. Date(s):	<u>06/23/2025 06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38 15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.411		-3.97	20
Chloromethane	0.816	0.897	0.1	9.93	20
Vinyl Chloride	1.020	1.044		2.35	20
Bromomethane	0.802	0.808		0.75	20
Chloroethane	0.686	0.718		4.66	20
Trichlorofluoromethane	1.129	1.131		0.18	20
1,1,2-Trichlorotrifluoroethane	0.516	0.529		2.52	20
1,1-Dichloroethene	0.506	0.524		3.56	20
Acetone	0.105	0.130		23.81	20
Carbon Disulfide	1.635	1.666		1.9	20
Methyl tert-butyl Ether	1.378	1.459		5.88	20
Methyl Acetate	0.349	0.342		-2.01	20
Methylene Chloride	0.666	0.654		-1.8	20
trans-1,2-Dichloroethene	0.578	0.601		3.98	20
1,1-Dichloroethane	1.044	1.127	0.1	7.95	20
Cyclohexane	0.959	1.013		5.63	20
2-Butanone	0.154	0.175		13.64	20
Carbon Tetrachloride	0.486	0.518		6.58	20
cis-1,2-Dichloroethene	0.672	0.696		3.57	20
Bromochloromethane	0.439	0.455		3.64	20
Chloroform	1.076	1.121		4.28	20
1,1,1-Trichloroethane	0.929	0.989		6.46	20
Methylcyclohexane	0.596	0.656		10.07	20
Benzene	1.417	1.521		7.34	20
1,2-Dichloroethane	0.388	0.415		6.96	20
Trichloroethene	0.356	0.379		6.46	20
1,2-Dichloropropane	0.332	0.363		9.34	20
Bromodichloromethane	0.485	0.513		5.77	20
4-Methyl-2-Pentanone	0.210	0.238		13.33	20
Toluene	0.894	0.967		8.17	20
t-1,3-Dichloropropene	0.437	0.480		9.84	20
cis-1,3-Dichloropropene	0.506	0.551		8.89	20
1,1,2-Trichloroethane	0.244	0.260		6.56	20
2-Hexanone	0.143	0.165		15.39	20
Dibromochloromethane	0.315	0.331		5.08	20
1,2-Dibromoethane	0.228	0.239		4.82	20
Tetrachloroethene	0.472	0.498		5.51	20
Chlorobenzene	1.099	1.178	0.3	7.19	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/02/2025</u> <u>10:37</u>
Lab File ID:	<u>VY022903.D</u>	Init. Calib. Date(s):	<u>06/23/2025</u> <u>06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38</u> <u>15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	2.164		12.12	20
m/p-Xylenes	0.746	0.834		11.8	20
o-Xylene	0.703	0.783		11.38	20
Styrene	1.178	1.313		11.46	20
Bromoform	0.207	0.220	0.1	6.28	20
Isopropylbenzene	3.698	4.114		11.25	20
1,1,2,2-Tetrachloroethane	0.596	0.635	0.3	6.54	20
1,3-Dichlorobenzene	1.683	1.822		8.26	20
1,4-Dichlorobenzene	1.672	1.793		7.24	20
1,2-Dichlorobenzene	1.483	1.577		6.34	20
1,2-Dibromo-3-Chloropropane	0.101	0.105		3.96	20
1,2,4-Trichlorobenzene	0.838	0.880		5.01	20
1,2,3-Trichlorobenzene	0.724	0.729		0.69	20
1,2-Dichloroethane-d4	0.558	0.550		-1.43	20
Dibromofluoromethane	0.304	0.303		-0.33	20
Toluene-d8	1.207	1.217		0.83	20
4-Bromofluorobenzene	0.388	0.389		0.26	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.