

LAB CHRONICLE

OrderID: Q2473	OrderDate: 7/1/2025 12:14:15 PM
Client: JPCL Engineering	Project: NOHO RFI No. SC64025 NYC
Contact: Paul Rotondi	Location: --Select--,A12,A43,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2473-01	PIT#1	SOIL	VOC-TCL	8260D	07/01/25		07/02/25	07/01/25
Q2473-02	PIT#2	SOIL	VOC-TCL	8260D	07/01/25		07/02/25	07/01/25
Q2473-03	PIT#3	SOIL	VOC-TCL	8260D	07/01/25		07/02/25	07/01/25
Q2473-04	PIT#4	SOIL	VOC-TCL	8260D	07/01/25		07/02/25	07/01/25
Q2473-04RE	PIT#4RE	SOIL	VOC-TCL	8260D	07/01/25		07/03/25	07/01/25

Hit Summary Sheet
SW-846

SDG No.: Q2473
Client: JPCL Engineering

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: Q2473-01	PIT#1 PIT#1	SOIL	Acetone	9.00	J	5.20	27.6	ug/Kg
			Total Voc :	9.00				
			Total Concentration:	9.00				
Client ID: Q2473-02	PIT#2 PIT#2	SOIL	Acetone	4.10	J	3.10	16.4	ug/Kg
			Total Voc :	4.10				
			Total Concentration:	4.10				
Client ID: Q2473-04	PIT#4 PIT#4	SOIL	Acetone	39.8		2.40	12.9	ug/Kg
Q2473-04	PIT#4	SOIL	2-Butanone	4.50	J	3.40	12.9	ug/Kg
			Total Voc :	44.3				
			Total Concentration:	44.3				
Client ID: Q2473-04RE	PIT#4RE PIT#4RE	SOIL	Acetone	39.2		3.60	19.0	ug/Kg
			Total Voc :	39.2				
			Total Concentration:	39.2				



SAMPLE DATA

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#1	SDG No.:	Q2473
Lab Sample ID:	Q2473-01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90
Sample Wt/Vol:	5.04 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022917.D	1	07/02/25 17:07	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.30	U	1.30	5.50	ug/Kg
74-87-3	Chloromethane	1.30	U	1.30	5.50	ug/Kg
75-01-4	Vinyl Chloride	0.87	U	0.87	5.50	ug/Kg
74-83-9	Bromomethane	1.20	U	1.20	5.50	ug/Kg
75-00-3	Chloroethane	1.40	U	1.40	5.50	ug/Kg
75-69-4	Trichlorofluoromethane	1.30	U	1.30	5.50	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.20	U	1.20	5.50	ug/Kg
75-35-4	1,1-Dichloroethene	1.10	U	1.10	5.50	ug/Kg
67-64-1	Acetone	9.00	J	5.20	27.6	ug/Kg
75-15-0	Carbon Disulfide	1.20	U	1.20	5.50	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.80	U	0.80	5.50	ug/Kg
79-20-9	Methyl Acetate	1.70	U	1.70	5.50	ug/Kg
75-09-2	Methylene Chloride	3.90	U	3.90	11.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.95	U	0.95	5.50	ug/Kg
75-34-3	1,1-Dichloroethane	0.88	U	0.88	5.50	ug/Kg
110-82-7	Cyclohexane	0.87	U	0.87	5.50	ug/Kg
78-93-3	2-Butanone	7.20	U	7.20	27.6	ug/Kg
56-23-5	Carbon Tetrachloride	1.10	U	1.10	5.50	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.83	U	0.83	5.50	ug/Kg
74-97-5	Bromochloromethane	1.30	U	1.30	5.50	ug/Kg
67-66-3	Chloroform	0.93	U	0.93	5.50	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.00	U	1.00	5.50	ug/Kg
108-87-2	Methylcyclohexane	1.00	U	1.00	5.50	ug/Kg
71-43-2	Benzene	0.87	U	0.87	5.50	ug/Kg
107-06-2	1,2-Dichloroethane	0.87	U	0.87	5.50	ug/Kg
79-01-6	Trichloroethene	0.89	U	0.89	5.50	ug/Kg
78-87-5	1,2-Dichloropropane	1.00	U	1.00	5.50	ug/Kg
75-27-4	Bromodichloromethane	0.86	U	0.86	5.50	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.90	U	3.90	27.6	ug/Kg
108-88-3	Toluene	0.86	U	0.86	5.50	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#1	SDG No.:	Q2473
Lab Sample ID:	Q2473-01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90
Sample Wt/Vol:	5.04 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022917.D	1	07/02/25 17:07	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.72	U	0.72	5.50	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.68	U	0.68	5.50	ug/Kg
79-00-5	1,1,2-Trichloroethane	1.00	U	1.00	5.50	ug/Kg
591-78-6	2-Hexanone	4.10	U	4.10	27.6	ug/Kg
124-48-1	Dibromochloromethane	0.96	U	0.96	5.50	ug/Kg
106-93-4	1,2-Dibromoethane	0.97	U	0.97	5.50	ug/Kg
127-18-4	Tetrachloroethene	1.20	U	1.20	5.50	ug/Kg
108-90-7	Chlorobenzene	1.00	U	1.00	5.50	ug/Kg
100-41-4	Ethyl Benzene	0.74	U	0.74	5.50	ug/Kg
179601-23-1	m/p-Xylenes	1.40	U	1.40	11.0	ug/Kg
95-47-6	o-Xylene	0.90	U	0.90	5.50	ug/Kg
100-42-5	Styrene	0.78	U	0.78	5.50	ug/Kg
75-25-2	Bromoform	0.95	U	0.95	5.50	ug/Kg
98-82-8	Isopropylbenzene	0.86	U	0.86	5.50	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.30	U	1.30	5.50	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.90	U	1.90	5.50	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.70	U	1.70	5.50	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.60	U	1.60	5.50	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	2.00	U	2.00	5.50	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.30	U	3.30	5.50	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.50	U	3.50	5.50	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.9		63 - 155	110%	SPK: 50
1868-53-7	Dibromofluoromethane	48.0		70 - 134	96%	SPK: 50
2037-26-5	Toluene-d8	51.5		74 - 123	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.6		17 - 146	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	279000	7.707			
540-36-3	1,4-Difluorobenzene	535000	8.615			
3114-55-4	Chlorobenzene-d5	530000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	216000	13.346			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#1	SDG No.:	Q2473
Lab Sample ID:	Q2473-01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90
Sample Wt/Vol:	5.04 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022917.D	1	07/02/25 17:07	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#2	SDG No.:	Q2473
Lab Sample ID:	Q2473-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.1
Sample Wt/Vol:	8.47 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022918.D	1	07/02/25 17:30	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.75	U	0.75	3.30	ug/Kg
74-87-3	Chloromethane	0.75	U	0.75	3.30	ug/Kg
75-01-4	Vinyl Chloride	0.52	U	0.52	3.30	ug/Kg
74-83-9	Bromomethane	0.70	U	0.70	3.30	ug/Kg
75-00-3	Chloroethane	0.83	U	0.83	3.30	ug/Kg
75-69-4	Trichlorofluoromethane	0.79	U	0.79	3.30	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.69	U	0.69	3.30	ug/Kg
75-35-4	1,1-Dichloroethene	0.66	U	0.66	3.30	ug/Kg
67-64-1	Acetone	4.10	J	3.10	16.4	ug/Kg
75-15-0	Carbon Disulfide	0.69	U	0.69	3.30	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.48	U	0.48	3.30	ug/Kg
79-20-9	Methyl Acetate	1.00	U	1.00	3.30	ug/Kg
75-09-2	Methylene Chloride	2.30	U	2.30	6.60	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.56	U	0.56	3.30	ug/Kg
75-34-3	1,1-Dichloroethane	0.52	U	0.52	3.30	ug/Kg
110-82-7	Cyclohexane	0.52	U	0.52	3.30	ug/Kg
78-93-3	2-Butanone	4.30	U	4.30	16.4	ug/Kg
56-23-5	Carbon Tetrachloride	0.64	U	0.64	3.30	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.49	U	0.49	3.30	ug/Kg
74-97-5	Bromochloromethane	0.75	U	0.75	3.30	ug/Kg
67-66-3	Chloroform	0.55	U	0.55	3.30	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.61	U	0.61	3.30	ug/Kg
108-87-2	Methylcyclohexane	0.60	U	0.60	3.30	ug/Kg
71-43-2	Benzene	0.52	U	0.52	3.30	ug/Kg
107-06-2	1,2-Dichloroethane	0.52	U	0.52	3.30	ug/Kg
79-01-6	Trichloroethene	0.53	U	0.53	3.30	ug/Kg
78-87-5	1,2-Dichloropropane	0.60	U	0.60	3.30	ug/Kg
75-27-4	Bromodichloromethane	0.51	U	0.51	3.30	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.30	U	2.30	16.4	ug/Kg
108-88-3	Toluene	0.51	U	0.51	3.30	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#2	SDG No.:	Q2473
Lab Sample ID:	Q2473-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.1
Sample Wt/Vol:	8.47 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022918.D	1	07/02/25 17:30	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.43	U	0.43	3.30	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.41	U	0.41	3.30	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.60	U	0.60	3.30	ug/Kg
591-78-6	2-Hexanone	2.40	U	2.40	16.4	ug/Kg
124-48-1	Dibromochloromethane	0.57	U	0.57	3.30	ug/Kg
106-93-4	1,2-Dibromoethane	0.58	U	0.58	3.30	ug/Kg
127-18-4	Tetrachloroethene	0.69	U	0.69	3.30	ug/Kg
108-90-7	Chlorobenzene	0.60	U	0.60	3.30	ug/Kg
100-41-4	Ethyl Benzene	0.44	U	0.44	3.30	ug/Kg
179601-23-1	m/p-Xylenes	0.81	U	0.81	6.60	ug/Kg
95-47-6	o-Xylene	0.54	U	0.54	3.30	ug/Kg
100-42-5	Styrene	0.47	U	0.47	3.30	ug/Kg
75-25-2	Bromoform	0.56	U	0.56	3.30	ug/Kg
98-82-8	Isopropylbenzene	0.51	U	0.51	3.30	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.79	U	0.79	3.30	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.10	U	1.10	3.30	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.00	U	1.00	3.30	ug/Kg
95-50-1	1,2-Dichlorobenzene	0.95	U	0.95	3.30	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.20	U	1.20	3.30	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	1.90	U	1.90	3.30	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.10	U	2.10	3.30	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.2		63 - 155	108%	SPK: 50
1868-53-7	Dibromofluoromethane	35.7		70 - 134	71%	SPK: 50
2037-26-5	Toluene-d8	50.8		74 - 123	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		17 - 146	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	300000	7.707			
540-36-3	1,4-Difluorobenzene	579000	8.616			
3114-55-4	Chlorobenzene-d5	556000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	220000	13.347			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#2	SDG No.:	Q2473
Lab Sample ID:	Q2473-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.1
Sample Wt/Vol:	8.47 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022918.D	1	07/02/25 17:30	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#3	SDG No.:	Q2473
Lab Sample ID:	Q2473-03	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	88.4
Sample Wt/Vol:	7.64 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022919.D	1	07/02/25 17:53	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.84	U	0.84	3.70	ug/Kg
74-87-3	Chloromethane	0.84	U	0.84	3.70	ug/Kg
75-01-4	Vinyl Chloride	0.58	U	0.58	3.70	ug/Kg
74-83-9	Bromomethane	0.79	U	0.79	3.70	ug/Kg
75-00-3	Chloroethane	0.93	U	0.93	3.70	ug/Kg
75-69-4	Trichlorofluoromethane	0.90	U	0.90	3.70	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.78	U	0.78	3.70	ug/Kg
75-35-4	1,1-Dichloroethene	0.74	U	0.74	3.70	ug/Kg
67-64-1	Acetone	3.50	U	3.50	18.5	ug/Kg
75-15-0	Carbon Disulfide	0.78	U	0.78	3.70	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.54	U	0.54	3.70	ug/Kg
79-20-9	Methyl Acetate	1.10	U	1.10	3.70	ug/Kg
75-09-2	Methylene Chloride	2.60	U	2.60	7.40	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.64	U	0.64	3.70	ug/Kg
75-34-3	1,1-Dichloroethane	0.59	U	0.59	3.70	ug/Kg
110-82-7	Cyclohexane	0.58	U	0.58	3.70	ug/Kg
78-93-3	2-Butanone	4.80	U	4.80	18.5	ug/Kg
56-23-5	Carbon Tetrachloride	0.72	U	0.72	3.70	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.56	U	0.56	3.70	ug/Kg
74-97-5	Bromochloromethane	0.85	U	0.85	3.70	ug/Kg
67-66-3	Chloroform	0.62	U	0.62	3.70	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.69	U	0.69	3.70	ug/Kg
108-87-2	Methylcyclohexane	0.67	U	0.67	3.70	ug/Kg
71-43-2	Benzene	0.58	U	0.58	3.70	ug/Kg
107-06-2	1,2-Dichloroethane	0.58	U	0.58	3.70	ug/Kg
79-01-6	Trichloroethene	0.60	U	0.60	3.70	ug/Kg
78-87-5	1,2-Dichloropropane	0.67	U	0.67	3.70	ug/Kg
75-27-4	Bromodichloromethane	0.58	U	0.58	3.70	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.70	U	2.70	18.5	ug/Kg
108-88-3	Toluene	0.58	U	0.58	3.70	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#3	SDG No.:	Q2473
Lab Sample ID:	Q2473-03	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	88.4
Sample Wt/Vol:	7.64 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022919.D	1	07/02/25 17:53	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.48	U	0.48	3.70	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.46	U	0.46	3.70	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.68	U	0.68	3.70	ug/Kg
591-78-6	2-Hexanone	2.70	U	2.70	18.5	ug/Kg
124-48-1	Dibromochloromethane	0.64	U	0.64	3.70	ug/Kg
106-93-4	1,2-Dibromoethane	0.65	U	0.65	3.70	ug/Kg
127-18-4	Tetrachloroethene	0.78	U	0.78	3.70	ug/Kg
108-90-7	Chlorobenzene	0.67	U	0.67	3.70	ug/Kg
100-41-4	Ethyl Benzene	0.50	U	0.50	3.70	ug/Kg
179601-23-1	m/p-Xylenes	0.92	U	0.92	7.40	ug/Kg
95-47-6	o-Xylene	0.61	U	0.61	3.70	ug/Kg
100-42-5	Styrene	0.53	U	0.53	3.70	ug/Kg
75-25-2	Bromoform	0.64	U	0.64	3.70	ug/Kg
98-82-8	Isopropylbenzene	0.58	U	0.58	3.70	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.90	U	0.90	3.70	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.30	U	1.30	3.70	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.20	U	1.20	3.70	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.10	U	1.10	3.70	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.40	U	1.40	3.70	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.20	U	2.20	3.70	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.40	U	2.40	3.70	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	35.9		63 - 155	72%	SPK: 50
1868-53-7	Dibromofluoromethane	38.2		70 - 134	76%	SPK: 50
2037-26-5	Toluene-d8	49.6		74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.6		17 - 146	89%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	310000	7.707			
540-36-3	1,4-Difluorobenzene	552000	8.616			
3114-55-4	Chlorobenzene-d5	476000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	182000	13.346			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#3	SDG No.:	Q2473
Lab Sample ID:	Q2473-03	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	88.4
Sample Wt/Vol:	7.64 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022919.D	1	07/02/25 17:53	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#4	SDG No.:	Q2473
Lab Sample ID:	Q2473-04	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.7
Sample Wt/Vol:	10.71 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022920.D	1	07/02/25 18:17	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.59	U	0.59	2.60	ug/Kg
74-87-3	Chloromethane	0.59	U	0.59	2.60	ug/Kg
75-01-4	Vinyl Chloride	0.41	U	0.41	2.60	ug/Kg
74-83-9	Bromomethane	0.55	U	0.55	2.60	ug/Kg
75-00-3	Chloroethane	0.65	U	0.65	2.60	ug/Kg
75-69-4	Trichlorofluoromethane	0.62	U	0.62	2.60	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.55	U	0.55	2.60	ug/Kg
75-35-4	1,1-Dichloroethene	0.51	U	0.51	2.60	ug/Kg
67-64-1	Acetone	39.8		2.40	12.9	ug/Kg
75-15-0	Carbon Disulfide	0.55	U	0.55	2.60	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.38	U	0.38	2.60	ug/Kg
79-20-9	Methyl Acetate	0.79	U	0.79	2.60	ug/Kg
75-09-2	Methylene Chloride	1.80	U	1.80	5.10	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.44	U	0.44	2.60	ug/Kg
75-34-3	1,1-Dichloroethane	0.41	U	0.41	2.60	ug/Kg
110-82-7	Cyclohexane	0.41	U	0.41	2.60	ug/Kg
78-93-3	2-Butanone	4.50	J	3.40	12.9	ug/Kg
56-23-5	Carbon Tetrachloride	0.50	U	0.50	2.60	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.39	U	0.39	2.60	ug/Kg
74-97-5	Bromochloromethane	0.59	U	0.59	2.60	ug/Kg
67-66-3	Chloroform	0.43	U	0.43	2.60	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.48	U	0.48	2.60	ug/Kg
108-87-2	Methylcyclohexane	0.47	U	0.47	2.60	ug/Kg
71-43-2	Benzene	0.41	U	0.41	2.60	ug/Kg
107-06-2	1,2-Dichloroethane	0.41	U	0.41	2.60	ug/Kg
79-01-6	Trichloroethene	0.42	U	0.42	2.60	ug/Kg
78-87-5	1,2-Dichloropropane	0.47	U	0.47	2.60	ug/Kg
75-27-4	Bromodichloromethane	0.40	U	0.40	2.60	ug/Kg
108-10-1	4-Methyl-2-Pentanone	1.80	U	1.80	12.9	ug/Kg
108-88-3	Toluene	0.40	U	0.40	2.60	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#4	SDG No.:	Q2473
Lab Sample ID:	Q2473-04	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.7
Sample Wt/Vol:	10.71 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022920.D	1	07/02/25 18:17	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.33	U	0.33	2.60	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.32	U	0.32	2.60	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.47	U	0.47	2.60	ug/Kg
591-78-6	2-Hexanone	1.90	U	1.90	12.9	ug/Kg
124-48-1	Dibromochloromethane	0.45	U	0.45	2.60	ug/Kg
106-93-4	1,2-Dibromoethane	0.45	U	0.45	2.60	ug/Kg
127-18-4	Tetrachloroethene	0.54	U	0.54	2.60	ug/Kg
108-90-7	Chlorobenzene	0.47	U	0.47	2.60	ug/Kg
100-41-4	Ethyl Benzene	0.34	U	0.34	2.60	ug/Kg
179601-23-1	m/p-Xylenes	0.64	U	0.64	5.10	ug/Kg
95-47-6	o-Xylene	0.42	U	0.42	2.60	ug/Kg
100-42-5	Styrene	0.37	U	0.37	2.60	ug/Kg
75-25-2	Bromoform	0.44	U	0.44	2.60	ug/Kg
98-82-8	Isopropylbenzene	0.40	U	0.40	2.60	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.62	U	0.62	2.60	ug/Kg
541-73-1	1,3-Dichlorobenzene	0.88	U	0.88	2.60	ug/Kg
106-46-7	1,4-Dichlorobenzene	0.80	U	0.80	2.60	ug/Kg
95-50-1	1,2-Dichlorobenzene	0.75	U	0.75	2.60	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	0.95	U	0.95	2.60	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	1.50	U	1.50	2.60	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	1.60	U	1.60	2.60	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	60.1		63 - 155	120%	SPK: 50
1868-53-7	Dibromofluoromethane	30.3	*	70 - 134	61%	SPK: 50
2037-26-5	Toluene-d8	51.5		74 - 123	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	59.4		17 - 146	119%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	269000	7.707			
540-36-3	1,4-Difluorobenzene	521000	8.616			
3114-55-4	Chlorobenzene-d5	540000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	245000	13.346			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#4	SDG No.:	Q2473
Lab Sample ID:	Q2473-04	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.7
Sample Wt/Vol:	10.71 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022920.D	1	07/02/25 18:17	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#4RE	SDG No.:	Q2473
Lab Sample ID:	Q2473-04RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.7
Sample Wt/Vol:	7.27 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022936.D	1	07/03/25 12:39	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.86	U	0.86	3.80	ug/Kg
74-87-3	Chloromethane	0.86	U	0.86	3.80	ug/Kg
75-01-4	Vinyl Chloride	0.60	U	0.60	3.80	ug/Kg
74-83-9	Bromomethane	0.81	U	0.81	3.80	ug/Kg
75-00-3	Chloroethane	0.96	U	0.96	3.80	ug/Kg
75-69-4	Trichlorofluoromethane	0.92	U	0.92	3.80	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.80	U	0.80	3.80	ug/Kg
75-35-4	1,1-Dichloroethene	0.76	U	0.76	3.80	ug/Kg
67-64-1	Acetone	39.2		3.60	19.0	ug/Kg
75-15-0	Carbon Disulfide	0.80	U	0.80	3.80	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.55	U	0.55	3.80	ug/Kg
79-20-9	Methyl Acetate	1.20	U	1.20	3.80	ug/Kg
75-09-2	Methylene Chloride	2.70	U	2.70	7.60	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.65	U	0.65	3.80	ug/Kg
75-34-3	1,1-Dichloroethane	0.61	U	0.61	3.80	ug/Kg
110-82-7	Cyclohexane	0.60	U	0.60	3.80	ug/Kg
78-93-3	2-Butanone	5.00	U	5.00	19.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.74	U	0.74	3.80	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.57	U	0.57	3.80	ug/Kg
74-97-5	Bromochloromethane	0.87	U	0.87	3.80	ug/Kg
67-66-3	Chloroform	0.64	U	0.64	3.80	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.71	U	0.71	3.80	ug/Kg
108-87-2	Methylcyclohexane	0.69	U	0.69	3.80	ug/Kg
71-43-2	Benzene	0.60	U	0.60	3.80	ug/Kg
107-06-2	1,2-Dichloroethane	0.60	U	0.60	3.80	ug/Kg
79-01-6	Trichloroethene	0.61	U	0.61	3.80	ug/Kg
78-87-5	1,2-Dichloropropane	0.69	U	0.69	3.80	ug/Kg
75-27-4	Bromodichloromethane	0.59	U	0.59	3.80	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.70	U	2.70	19.0	ug/Kg
108-88-3	Toluene	0.59	U	0.59	3.80	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#4RE	SDG No.:	Q2473
Lab Sample ID:	Q2473-04RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.7
Sample Wt/Vol:	7.27 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022936.D	1	07/03/25 12:39	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.49	U	0.49	3.80	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.47	U	0.47	3.80	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.70	U	0.70	3.80	ug/Kg
591-78-6	2-Hexanone	2.80	U	2.80	19.0	ug/Kg
124-48-1	Dibromochloromethane	0.66	U	0.66	3.80	ug/Kg
106-93-4	1,2-Dibromoethane	0.67	U	0.67	3.80	ug/Kg
127-18-4	Tetrachloroethene	0.80	U	0.80	3.80	ug/Kg
108-90-7	Chlorobenzene	0.69	U	0.69	3.80	ug/Kg
100-41-4	Ethyl Benzene	0.51	U	0.51	3.80	ug/Kg
179601-23-1	m/p-Xylenes	0.94	U	0.94	7.60	ug/Kg
95-47-6	o-Xylene	0.62	U	0.62	3.80	ug/Kg
100-42-5	Styrene	0.54	U	0.54	3.80	ug/Kg
75-25-2	Bromoform	0.65	U	0.65	3.80	ug/Kg
98-82-8	Isopropylbenzene	0.59	U	0.59	3.80	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.92	U	0.92	3.80	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.30	U	1.30	3.80	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.20	U	1.20	3.80	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.10	U	1.10	3.80	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.40	U	1.40	3.80	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.30	U	2.30	3.80	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.40	U	2.40	3.80	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.8		63 - 155	98%	SPK: 50
1868-53-7	Dibromofluoromethane	31.1	*	70 - 134	62%	SPK: 50
2037-26-5	Toluene-d8	49.0		74 - 123	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.3		17 - 146	93%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	95600	7.707			
540-36-3	1,4-Difluorobenzene	165000	8.616			
3114-55-4	Chlorobenzene-d5	149000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	56100	13.346			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	07/01/25
Project:	NOHO RFI No. SC64025 NYC	Date Received:	07/01/25
Client Sample ID:	PIT#4RE	SDG No.:	Q2473
Lab Sample ID:	Q2473-04RE	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.7
Sample Wt/Vol:	7.27 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022936.D	1	07/03/25 12:39	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q2473

Client: JPCL Engineering

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2473-01	PIT#1	1,2-Dichloroethane-d4	50	54.9	110		63	155
		Dibromofluoromethane	50	48.0	96		70	134
		Toluene-d8	50	51.5	103		74	123
		4-Bromofluorobenzene	50	53.6	107		17	146
Q2473-02	PIT#2	1,2-Dichloroethane-d4	50	54.2	108		63	155
		Dibromofluoromethane	50	35.7	71		70	134
		Toluene-d8	50	50.8	102		74	123
		4-Bromofluorobenzene	50	51.9	104		17	146
Q2473-03	PIT#3	1,2-Dichloroethane-d4	50	35.9	72		63	155
		Dibromofluoromethane	50	38.3	76		70	134
		Toluene-d8	50	49.6	99		74	123
		4-Bromofluorobenzene	50	44.6	89		17	146
Q2473-04	PIT#4	1,2-Dichloroethane-d4	50	60.1	120		63	155
		Dibromofluoromethane	50	30.3	61	*	70	134
		Toluene-d8	50	51.5	103		74	123
		4-Bromofluorobenzene	50	59.4	119		17	146
Q2473-04RE	PIT#4RE	1,2-Dichloroethane-d4	50	48.8	98		63	155
		Dibromofluoromethane	50	31.1	62	*	70	134
		Toluene-d8	50	49.0	98		74	123
		4-Bromofluorobenzene	50	46.3	93		17	146
VY0702SBL01	VY0702SBL01	1,2-Dichloroethane-d4	50	49.3	99		63	155
		Dibromofluoromethane	50	50.9	102		70	134
		Toluene-d8	50	51.3	103		74	123
		4-Bromofluorobenzene	50	55.3	111		17	146
VY0702SBS01	VY0702SBS01	1,2-Dichloroethane-d4	50	49.8	100		63	155
		Dibromofluoromethane	50	52.0	104		70	134
		Toluene-d8	50	51.9	104		74	123
		4-Bromofluorobenzene	50	48.9	98		17	146
VY0702SBSD01	VY0702SBSD01	1,2-Dichloroethane-d4	50	48.9	98		63	155
		Dibromofluoromethane	50	49.5	99		70	134
		Toluene-d8	50	49.2	98		74	123
		4-Bromofluorobenzene	50	47.2	94		17	146
VY0703SBL01	VY0703SBL01	1,2-Dichloroethane-d4	50	52.7	105		63	155
		Dibromofluoromethane	50	52.7	105		70	134
		Toluene-d8	50	52.1	104		74	123
		4-Bromofluorobenzene	50	55.3	111		17	146
VY0703SBS01	VY0703SBS01	1,2-Dichloroethane-d4	50	48.4	97		63	155
		Dibromofluoromethane	50	50.6	101		70	134
		Toluene-d8	50	51.1	102		74	123
		4-Bromofluorobenzene	50	47.0	94		17	146

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q2473	SW-846	Analytical Method:	SW8260D
Client:	JPCL Engineering		Datafile :	VY022905.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0702SBS01	Dichlorodifluoromethane	20	20.3	ug/Kg	102			64	136	
	Chloromethane	20	22.7	ug/Kg	114			52	151	
	Vinyl chloride	20	20.1	ug/Kg	101			56	148	
	Bromomethane	20	21.9	ug/Kg	110			58	141	
	Chloroethane	20	21.3	ug/Kg	106			69	130	
	Trichlorofluoromethane	20	19.4	ug/Kg	97			69	134	
	1,1,2-Trichlorotrifluoroethane	20	20.9	ug/Kg	104			81	123	
	1,1-Dichloroethene	20	20.7	ug/Kg	104			79	121	
	Acetone	100	110	ug/Kg	110			40	171	
	Carbon disulfide	20	20.6	ug/Kg	103			59	130	
	Methyl tert-butyl Ether	20	19.0	ug/Kg	95			77	129	
	Methyl Acetate	20	17.0	ug/Kg	85			69	149	
	Methylene Chloride	20	22.0	ug/Kg	110			72	131	
	trans-1,2-Dichloroethene	20	20.5	ug/Kg	103			80	123	
	1,1-Dichloroethane	20	21.0	ug/Kg	105			82	123	
	Cyclohexane	20	21.2	ug/Kg	106			76	122	
	2-Butanone	100	97.2	ug/Kg	97			69	131	
	Carbon Tetrachloride	20	20.5	ug/Kg	103			76	129	
	cis-1,2-Dichloroethene	20	20.2	ug/Kg	101			82	123	
	Bromochloromethane	20	20.2	ug/Kg	101			80	127	
	Chloroform	20	20.7	ug/Kg	104			82	125	
	1,1,1-Trichloroethane	20	20.8	ug/Kg	104			80	126	
	Methylcyclohexane	20	20.6	ug/Kg	103			77	123	
	Benzene	20	20.6	ug/Kg	103			84	121	
	1,2-Dichloroethane	20	20.0	ug/Kg	100			81	126	
	Trichloroethene	20	19.9	ug/Kg	100			83	122	
	1,2-Dichloropropane	20	20.8	ug/Kg	104			83	122	
	Bromodichloromethane	20	19.8	ug/Kg	99			82	123	
	4-Methyl-2-Pentanone	100	89.3	ug/Kg	89			70	135	
	Toluene	20	20.1	ug/Kg	101			83	122	
	t-1,3-Dichloropropene	20	19.3	ug/Kg	97			78	124	
	cis-1,3-Dichloropropene	20	20.2	ug/Kg	101			81	122	
	1,1,2-Trichloroethane	20	19.2	ug/Kg	96			82	125	
	2-Hexanone	100	90.1	ug/Kg	90			66	138	
	Dibromochloromethane	20	18.8	ug/Kg	94			79	125	
	1,2-Dibromoethane	20	18.3	ug/Kg	92			80	125	
	Tetrachloroethene	20	19.7	ug/Kg	99			83	125	
	Chlorobenzene	20	20.4	ug/Kg	102			84	122	
	Ethyl Benzene	20	20.6	ug/Kg	103			82	124	
	m/p-Xylenes	40	40.6	ug/Kg	102			83	124	
	o-Xylene	20	20.0	ug/Kg	100			83	123	
	Styrene	20	19.7	ug/Kg	99			82	124	
	Bromoform	20	18.4	ug/Kg	92			75	127	
	Isopropylbenzene	20	21.8	ug/Kg	109			82	124	
	1,1,2,2-Tetrachloroethane	20	20.8	ug/Kg	104			77	127	
	1,3-Dichlorobenzene	20	20.4	ug/Kg	102			83	122	
	1,4-Dichlorobenzene	20	20.5	ug/Kg	103			84	121	
	1,2-Dichlorobenzene	20	20.1	ug/Kg	101			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2473</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>JPCL Engineering</u>	Datafile :	<u>VY022905.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VY0702SBS01	1,2-Dibromo-3-Chloropropane	20	18.3	ug/Kg	92			66	134	
	1,2,4-Trichlorobenzene	20	18.4	ug/Kg	92			78	127	
	1,2,3-Trichlorobenzene	20	17.9	ug/Kg	90			70	137	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q2473	SW-846	Analytical Method:	SW8260D
Client:	JPCL Engineering		Datafile :	VY022906.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0702SBSD01	Dichlorodifluoromethane	20	21.1	ug/Kg	106	4		64	136	20
	Chloromethane	20	20.5	ug/Kg	103	10		52	151	20
	Vinyl chloride	20	19.5	ug/Kg	98	3		56	148	20
	Bromomethane	20	21.3	ug/Kg	106	4		58	141	20
	Chloroethane	20	20.5	ug/Kg	103	3		69	130	20
	Trichlorofluoromethane	20	19.3	ug/Kg	97	0		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	21.3	ug/Kg	106	2		81	123	20
	1,1-Dichloroethene	20	21.4	ug/Kg	107	3		79	121	20
	Acetone	100	120	ug/Kg	120	9		40	171	20
	Carbon disulfide	20	21.0	ug/Kg	105	2		59	130	20
	Methyl tert-butyl Ether	20	20.3	ug/Kg	102	7		77	129	20
	Methyl Acetate	20	19.1	ug/Kg	96	12		69	149	20
	Methylene Chloride	20	24.1	ug/Kg	121	10		72	131	20
	trans-1,2-Dichloroethene	20	21.3	ug/Kg	106	3		80	123	20
	1,1-Dichloroethane	20	22.0	ug/Kg	110	5		82	123	20
	Cyclohexane	20	21.9	ug/Kg	110	4		76	122	20
	2-Butanone	100	110	ug/Kg	110	13		69	131	20
	Carbon Tetrachloride	20	20.9	ug/Kg	104	1		76	129	20
	cis-1,2-Dichloroethene	20	20.6	ug/Kg	103	2		82	123	20
	Bromochloromethane	20	21.4	ug/Kg	107	6		80	127	20
	Chloroform	20	21.3	ug/Kg	106	2		82	125	20
	1,1,1-Trichloroethane	20	21.6	ug/Kg	108	4		80	126	20
	Methylcyclohexane	20	20.6	ug/Kg	103	0		77	123	20
	Benzene	20	21.2	ug/Kg	106	3		84	121	20
	1,2-Dichloroethane	20	20.8	ug/Kg	104	4		81	126	20
	Trichloroethene	20	20.6	ug/Kg	103	3		83	122	20
	1,2-Dichloropropane	20	21.6	ug/Kg	108	4		83	122	20
	Bromodichloromethane	20	20.6	ug/Kg	103	4		82	123	20
	4-Methyl-2-Pentanone	100	97.1	ug/Kg	97	9		70	135	20
	Toluene	20	20.8	ug/Kg	104	3		83	122	20
	t-1,3-Dichloropropene	20	20.1	ug/Kg	101	4		78	124	20
	cis-1,3-Dichloropropene	20	20.9	ug/Kg	104	3		81	122	20
	1,1,2-Trichloroethane	20	19.9	ug/Kg	100	4		82	125	20
	2-Hexanone	100	98.9	ug/Kg	99	10		66	138	20
	Dibromochloromethane	20	20.1	ug/Kg	101	7		79	125	20
	1,2-Dibromoethane	20	19.2	ug/Kg	96	4		80	125	20
	Tetrachloroethene	20	20.3	ug/Kg	102	3		83	125	20
	Chlorobenzene	20	21.2	ug/Kg	106	4		84	122	20
	Ethyl Benzene	20	21.1	ug/Kg	106	3		82	124	20
	m/p-Xylenes	40	41.8	ug/Kg	104	2		83	124	20
	o-Xylene	20	20.9	ug/Kg	104	4		83	123	20
	Styrene	20	20.5	ug/Kg	103	4		82	124	20
	Bromoform	20	19.3	ug/Kg	97	5		75	127	20
	Isopropylbenzene	20	22.1	ug/Kg	111	2		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.4	ug/Kg	107	3		77	127	20
	1,3-Dichlorobenzene	20	21.3	ug/Kg	106	4		83	122	20
	1,4-Dichlorobenzene	20	21.0	ug/Kg	105	2		84	121	20
	1,2-Dichlorobenzene	20	20.9	ug/Kg	104	3		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2473</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>JPCL Engineering</u>	Datafile :	<u>VY022906.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0702SBSD01	1,2-Dibromo-3-Chloropropane	20	19.3	ug/Kg	97	5		66	134	20
	1,2,4-Trichlorobenzene	20	19.7	ug/Kg	99	7		78	127	20
	1,2,3-Trichlorobenzene	20	19.2	ug/Kg	96	6		70	137	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q2473	SW-846	Analytical Method:	SW8260D
Client:	JPCL Engineering		Datafile :	VY022931.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0703SBS01	Dichlorodifluoromethane	20	21.3	ug/Kg	106			64	136	
	Chloromethane	20	20.1	ug/Kg	101			52	151	
	Vinyl chloride	20	20.1	ug/Kg	101			56	148	
	Bromomethane	20	19.6	ug/Kg	98			58	141	
	Chloroethane	20	20.3	ug/Kg	102			69	130	
	Trichlorofluoromethane	20	19.3	ug/Kg	97			69	134	
	1,1,2-Trichlorotrifluoroethane	20	21.8	ug/Kg	109			81	123	
	1,1-Dichloroethene	20	21.2	ug/Kg	106			79	121	
	Acetone	100	140	ug/Kg	140			40	171	
	Carbon disulfide	20	21.1	ug/Kg	106			59	130	
	Methyl tert-butyl Ether	20	18.3	ug/Kg	92			77	129	
	Methyl Acetate	20	16.2	ug/Kg	81			69	149	
	Methylene Chloride	20	24.0	ug/Kg	120			72	131	
	trans-1,2-Dichloroethene	20	20.9	ug/Kg	104			80	123	
	1,1-Dichloroethane	20	21.6	ug/Kg	108			82	123	
	Cyclohexane	20	21.2	ug/Kg	106			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	21.1	ug/Kg	106			76	129	
	cis-1,2-Dichloroethene	20	20.5	ug/Kg	103			82	123	
	Bromochloromethane	20	20.3	ug/Kg	102			80	127	
	Chloroform	20	21.0	ug/Kg	105			82	125	
	1,1,1-Trichloroethane	20	21.2	ug/Kg	106			80	126	
	Methylcyclohexane	20	20.6	ug/Kg	103			77	123	
	Benzene	20	21.4	ug/Kg	107			84	121	
	1,2-Dichloroethane	20	20.6	ug/Kg	103			81	126	
	Trichloroethene	20	21.2	ug/Kg	106			83	122	
	1,2-Dichloropropane	20	21.4	ug/Kg	107			83	122	
	Bromodichloromethane	20	20.4	ug/Kg	102			82	123	
	4-Methyl-2-Pentanone	100	88.7	ug/Kg	89			70	135	
	Toluene	20	20.6	ug/Kg	103			83	122	
	t-1,3-Dichloropropene	20	19.2	ug/Kg	96			78	124	
	cis-1,3-Dichloropropene	20	20.4	ug/Kg	102			81	122	
	1,1,2-Trichloroethane	20	19.4	ug/Kg	97			82	125	
	2-Hexanone	100	97.3	ug/Kg	97			66	138	
	Dibromochloromethane	20	19.1	ug/Kg	96			79	125	
	1,2-Dibromoethane	20	18.2	ug/Kg	91			80	125	
	Tetrachloroethene	20	22.3	ug/Kg	112			83	125	
	Chlorobenzene	20	20.9	ug/Kg	104			84	122	
	Ethyl Benzene	20	21.4	ug/Kg	107			82	124	
	m/p-Xylenes	40	42.4	ug/Kg	106			83	124	
	o-Xylene	20	20.1	ug/Kg	101			83	123	
	Styrene	20	20.2	ug/Kg	101			82	124	
	Bromoform	20	17.7	ug/Kg	89			75	127	
	Isopropylbenzene	20	21.8	ug/Kg	109			82	124	
	1,1,2,2-Tetrachloroethane	20	18.9	ug/Kg	95			77	127	
	1,3-Dichlorobenzene	20	20.7	ug/Kg	104			83	122	
	1,4-Dichlorobenzene	20	20.5	ug/Kg	103			84	121	
	1,2-Dichlorobenzene	20	19.6	ug/Kg	98			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.: Q2473 SW-846 Analytical Method: SW8260D
Client: JPCL Engineering Datafile : VY022931.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0703SBS01	1,2-Dibromo-3-Chloropropane	20	17.8	ug/Kg	89			66	134	
	1,2,4-Trichlorobenzene	20	17.7	ug/Kg	89			78	127	
	1,2,3-Trichlorobenzene	20	16.9	ug/Kg	85			70	137	

VOLATILE METHOD BLANK SUMMARY

Client ID

VY0702SBL01

Lab Name: Alliance

Contract: JPCL01

Lab Code: ACE

SDG NO.: Q2473

Lab File ID: VY022904.D

Lab Sample ID: VY0702SBL01

Date Analyzed: 07/02/2025

Time Analyzed: 11:42

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0702SBS01	VY0702SBS01	VY022905.D	07/02/2025
VY0702SBSD01	VY0702SBSD01	VY022906.D	07/02/2025
PIT#1	Q2473-01	VY022917.D	07/02/2025
PIT#2	Q2473-02	VY022918.D	07/02/2025
PIT#3	Q2473-03	VY022919.D	07/02/2025
PIT#4	Q2473-04	VY022920.D	07/02/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

Client ID

VY0703SBL01

Lab Name: Alliance

Contract: JPCL01

Lab Code: ACE

SDG NO.: Q2473

Lab File ID: VY022930.D

Lab Sample ID: VY0703SBL01

Date Analyzed: 07/03/2025

Time Analyzed: 09:57

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0703SBS01	VY0703SBS01	VY022931.D	07/03/2025
PIT#4RE	Q2473-04RE	VY022936.D	07/03/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JPCL01

Lab Code: ACE

SDG NO.: Q2473

Lab File ID: VY022775.D

BFB Injection Date: 06/23/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 10:17

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.8
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	81.9
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	78.2 (95.5) 1
177	5.0 - 9.0% of mass 176	5.1 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY022776.D	06/23/2025	13:38
VSTDIC010	VSTDIC010	VY022777.D	06/23/2025	14:00
VSTDIC020	VSTDIC020	VY022778.D	06/23/2025	14:23
VSTDIC050	VSTDIC050	VY022779.D	06/23/2025	14:46
VSTDIC100	VSTDIC100	VY022780.D	06/23/2025	15:08
VSTDIC150	VSTDIC150	VY022781.D	06/23/2025	15:31

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JPCL01

Lab Code: ACE

SDG NO.: Q2473

Lab File ID: VY022902.D

BFB Injection Date: 07/02/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:51

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	55.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	83.6
175	5.0 - 9.0% of mass 174	6.1 (7.3) 1
176	95.0 - 101.0% of mass 174	81.3 (97.3) 1
177	5.0 - 9.0% of mass 176	5.4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022903.D	07/02/2025	10:37
VY0702SBL01	VY0702SBL01	VY022904.D	07/02/2025	11:42
VY0702SBS01	VY0702SBS01	VY022905.D	07/02/2025	12:12
VY0702SBSD01	VY0702SBSD01	VY022906.D	07/02/2025	12:35
PIT#1	Q2473-01	VY022917.D	07/02/2025	17:07
PIT#2	Q2473-02	VY022918.D	07/02/2025	17:30
PIT#3	Q2473-03	VY022919.D	07/02/2025	17:53
PIT#4	Q2473-04	VY022920.D	07/02/2025	18:17

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JPCL01

Lab Code: ACE

SDG NO.: Q2473

Lab File ID: VY022928.D

BFB Injection Date: 07/03/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:03

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23
75	30.0 - 60.0% of mass 95	56.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1.1 (1.3) 1
174	50.0 - 100.0% of mass 95	85.2
175	5.0 - 9.0% of mass 174	6.3 (7.3) 1
176	95.0 - 101.0% of mass 174	81 (95.1) 1
177	5.0 - 9.0% of mass 176	5.3 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022929.D	07/03/2025	09:26
VY0703SBL01	VY0703SBL01	VY022930.D	07/03/2025	09:57
VY0703SBS01	VY0703SBS01	VY022931.D	07/03/2025	10:29
PIT#4RE	Q2473-04RE	VY022936.D	07/03/2025	12:39

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JPCL01
 Lab Code: ACE SDG NO.: Q2473
 Lab File ID: VY022903.D Date Analyzed: 07/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	453239	7.71	750175	8.62	644334	11.41
UPPER LIMIT	906478	8.213	1500350	9.116	1288670	11.914
LOWER LIMIT	226620	7.213	375088	8.116	322167	10.914
EPA SAMPLE NO.						
PIT#1	279065	7.71	534952	8.62	530195	11.41
PIT#2	300149	7.71	579446	8.62	556091	11.41
PIT#3	309766	7.71	551696	8.62	475549	11.41
PIT#4	269260	7.71	521273	8.62	540209	11.41
VY0702SBL01	327303	7.71	608180	8.62	599654	11.42
VY0702SBS01	424569	7.71	714057	8.62	598160	11.41
VY0702SBSD01	405936	7.71	690466	8.62	580305	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2473</u>
Lab File ID:	<u>VY022903.D</u>	Date Analyzed:	<u>07/02/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>10:37</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	321198	13.347				
UPPER LIMIT	642396	13.847				
LOWER LIMIT	160599	12.847				
EPA SAMPLE NO.						
PIT#1	216048	13.35				
PIT#2	220138	13.35				
PIT#3	181660	13.35				
PIT#4	244944	13.35				
VY0702SBL01	251367	13.35				
VY0702SBS01	274634	13.35				
VY0702SBSD01	268831	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JPCL01
Lab Code: ACE SDG NO.: Q2473
Lab File ID: VY022929.D Date Analyzed: 07/03/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:26
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	402903	7.71	653588	8.62	556811	11.41
UPPER LIMIT	805806	8.207	1307180	9.116	1113620	11.914
LOWER LIMIT	201452	7.207	326794	8.116	278406	10.914
EPA SAMPLE NO.						
PIT#4RE	95586 *	7.71	165245 *	8.62	149379 *	11.41
VY0703SBL01	297416	7.71	558302	8.62	553906	11.41
VY0703SBS01	387913	7.71	648160	8.62	537972	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2473</u>
Lab File ID:	<u>VY022929.D</u>	Date Analyzed:	<u>07/03/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>09:26</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	273118	13.347				
UPPER LIMIT	546236	13.847				
LOWER LIMIT	136559	12.847				
EPA SAMPLE NO.						
PIT#4RE	56106 *	13.35				
VY0703SBL01	230762	13.35				
VY0703SBS01	252496	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0702SBL01	SDG No.:	Q2473
Lab Sample ID:	VY0702SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022904.D	1	07/02/25 11:42	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	JPCL Engineering		Date Collected:	
Project:	NOHO RFI No. SC64025 NYC		Date Received:	
Client Sample ID:	VY0702SBL01		SDG No.:	Q2473
Lab Sample ID:	VY0702SBL01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCL
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022904.D	1	07/02/25 11:42	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.3		63 - 155	99%	SPK: 50
1868-53-7	Dibromofluoromethane	50.9		70 - 134	102%	SPK: 50
2037-26-5	Toluene-d8	51.3		74 - 123	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.3		17 - 146	111%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	327000	7.713			
540-36-3	1,4-Difluorobenzene	608000	8.616			
3114-55-4	Chlorobenzene-d5	600000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	251000	13.346			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0702SBL01	SDG No.:	Q2473
Lab Sample ID:	VY0702SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022904.D	1	07/02/25 11:42	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering		Date Collected:	
Project:	NOHO RFI No. SC64025 NYC		Date Received:	
Client Sample ID:	VY0703SBL01		SDG No.:	Q2473
Lab Sample ID:	VY0703SBL01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCL
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022930.D	1	07/03/25 09:57	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0703SBL01	SDG No.:	Q2473
Lab Sample ID:	VY0703SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022930.D	1	07/03/25 09:57	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.7		63 - 155	105%	SPK: 50
1868-53-7	Dibromofluoromethane	52.7		70 - 134	105%	SPK: 50
2037-26-5	Toluene-d8	52.1		74 - 123	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.3		17 - 146	111%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	297000	7.707			
540-36-3	1,4-Difluorobenzene	558000	8.615			
3114-55-4	Chlorobenzene-d5	554000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	231000	13.346			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0703SBL01	SDG No.:	Q2473
Lab Sample ID:	VY0703SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022930.D	1	07/03/25 09:57	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0702SBS01	SDG No.:	Q2473
Lab Sample ID:	VY0702SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022905.D	1	07/02/25 12:12	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	20.3		1.10	5.00	ug/Kg
74-87-3	Chloromethane	22.7		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	20.1		0.79	5.00	ug/Kg
74-83-9	Bromomethane	21.9		1.10	5.00	ug/Kg
75-00-3	Chloroethane	21.3		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	19.4		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	20.9		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	20.7		1.00	5.00	ug/Kg
67-64-1	Acetone	110		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	20.6		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	19.0		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	17.0		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	22.0		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	20.5		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	21.0		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	21.2		0.79	5.00	ug/Kg
78-93-3	2-Butanone	97.2		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	20.5		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	20.2		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	20.2		1.20	5.00	ug/Kg
67-66-3	Chloroform	20.7		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.8		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	20.6		0.91	5.00	ug/Kg
71-43-2	Benzene	20.6		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	20.0		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	19.9		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	20.8		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	19.8		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	89.3		3.60	25.0	ug/Kg
108-88-3	Toluene	20.1		0.78	5.00	ug/Kg

Report of Analysis

Client:	JPCL Engineering		Date Collected:	
Project:	NOHO RFI No. SC64025 NYC		Date Received:	
Client Sample ID:	VY0702SBS01		SDG No.:	Q2473
Lab Sample ID:	VY0702SBS01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCL
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022905.D	1	07/02/25 12:12	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	19.3		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	20.2		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	19.2		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	90.1		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	18.8		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	18.3		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	19.7		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	20.4		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.6		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	40.6		1.20	10.0	ug/Kg
95-47-6	o-Xylene	20.0		0.82	5.00	ug/Kg
100-42-5	Styrene	19.7		0.71	5.00	ug/Kg
75-25-2	Bromoform	18.4		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.8		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	20.8		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	20.4		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	20.5		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.1		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	18.3		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	18.4		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	17.9		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.8		63 - 155	100%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		70 - 134	104%	SPK: 50
2037-26-5	Toluene-d8	51.9		74 - 123	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.9		17 - 146	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	425000	7.707			
540-36-3	1,4-Difluorobenzene	714000	8.616			
3114-55-4	Chlorobenzene-d5	598000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	275000	13.346			

Report of Analysis

Client:	JPCL Engineering		Date Collected:	
Project:	NOHO RFI No. SC64025 NYC		Date Received:	
Client Sample ID:	VY0702SBS01		SDG No.:	Q2473
Lab Sample ID:	VY0702SBS01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCL
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022905.D	1	07/02/25 12:12	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0703SBS01	SDG No.:	Q2473
Lab Sample ID:	VY0703SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022931.D	1	07/03/25 10:29	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	21.3		1.10	5.00	ug/Kg
74-87-3	Chloromethane	20.1		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	20.1		0.79	5.00	ug/Kg
74-83-9	Bromomethane	19.6		1.10	5.00	ug/Kg
75-00-3	Chloroethane	20.3		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	19.3		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	21.8		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	21.2		1.00	5.00	ug/Kg
67-64-1	Acetone	140		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	21.1		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	18.3		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	16.2		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	24.0		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	20.9		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	21.6		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	21.2		0.79	5.00	ug/Kg
78-93-3	2-Butanone	110		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	21.1		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	20.5		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	20.3		1.20	5.00	ug/Kg
67-66-3	Chloroform	21.0		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	21.2		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	20.6		0.91	5.00	ug/Kg
71-43-2	Benzene	21.4		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	20.6		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	21.2		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	21.4		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.4		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	88.7		3.60	25.0	ug/Kg
108-88-3	Toluene	20.6		0.78	5.00	ug/Kg

Report of Analysis

Client:	JPCL Engineering		Date Collected:	
Project:	NOHO RFI No. SC64025 NYC		Date Received:	
Client Sample ID:	VY0703SBS01		SDG No.:	Q2473
Lab Sample ID:	VY0703SBS01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCL
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022931.D	1	07/03/25 10:29	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	19.2		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	20.4		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	19.4		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	97.3		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	19.1		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	18.2		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	22.3		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	20.9		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	21.4		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	42.4		1.20	10.0	ug/Kg
95-47-6	o-Xylene	20.1		0.82	5.00	ug/Kg
100-42-5	Styrene	20.2		0.71	5.00	ug/Kg
75-25-2	Bromoform	17.7		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.8		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	18.9		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	20.7		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	20.5		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	19.6		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	17.8		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	17.7		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	16.9		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.4		63 - 155	97%	SPK: 50
1868-53-7	Dibromofluoromethane	50.6		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	51.1		74 - 123	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.0		17 - 146	94%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	388000	7.707			
540-36-3	1,4-Difluorobenzene	648000	8.616			
3114-55-4	Chlorobenzene-d5	538000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	252000	13.347			

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0703SBS01	SDG No.:	Q2473
Lab Sample ID:	VY0703SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022931.D	1	07/03/25 10:29	VY070325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JPCL Engineering	Date Collected:	
Project:	NOHO RFI No. SC64025 NYC	Date Received:	
Client Sample ID:	VY0702SBSD01	SDG No.:	Q2473
Lab Sample ID:	VY0702SBSD01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCL
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022906.D	1	07/02/25 12:35	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	21.1		1.10	5.00	ug/Kg
74-87-3	Chloromethane	20.5		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	19.5		0.79	5.00	ug/Kg
74-83-9	Bromomethane	21.3		1.10	5.00	ug/Kg
75-00-3	Chloroethane	20.5		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	19.3		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	21.3		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	21.4		1.00	5.00	ug/Kg
67-64-1	Acetone	120		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	21.0		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	20.3		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	19.1		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	24.1		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	21.3		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	22.0		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	21.9		0.79	5.00	ug/Kg
78-93-3	2-Butanone	110		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	20.9		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	20.6		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	21.4		1.20	5.00	ug/Kg
67-66-3	Chloroform	21.3		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	21.6		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	20.6		0.91	5.00	ug/Kg
71-43-2	Benzene	21.2		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	20.8		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.6		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	21.6		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.6		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	97.1		3.60	25.0	ug/Kg
108-88-3	Toluene	20.8		0.78	5.00	ug/Kg

Report of Analysis

Client:	JPCL Engineering		Date Collected:	
Project:	NOHO RFI No. SC64025 NYC		Date Received:	
Client Sample ID:	VY0702SBSD01	SDG No.:	Q2473	
Lab Sample ID:	VY0702SBSD01	Matrix:	SOIL	
Analytical Method:	8260D	% Solid:	100	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCL
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022906.D	1	07/02/25 12:35	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	20.1		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	20.9		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	19.9		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	98.9		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	20.1		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	19.2		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	20.3		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	21.2		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	21.1		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	41.8		1.20	10.0	ug/Kg
95-47-6	o-Xylene	20.9		0.82	5.00	ug/Kg
100-42-5	Styrene	20.5		0.71	5.00	ug/Kg
75-25-2	Bromoform	19.3		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	22.1		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	21.4		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	21.3		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	21.0		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.9		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	19.3		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	19.7		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	19.2		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.9		63 - 155	98%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		70 - 134	99%	SPK: 50
2037-26-5	Toluene-d8	49.2		74 - 123	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		17 - 146	94%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	406000	7.713			
540-36-3	1,4-Difluorobenzene	690000	8.615			
3114-55-4	Chlorobenzene-d5	580000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	269000	13.346			

Report of Analysis

Client:	JPCL Engineering		Date Collected:	
Project:	NOHO RFI No. SC64025 NYC		Date Received:	
Client Sample ID:	VY0702SBSD01		SDG No.:	Q2473
Lab Sample ID:	VY0702SBSD01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCL
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY022906.D	1	07/02/25 12:35	VY070225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCI01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date(s):	<u>06/23/2025</u> <u>06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Calibration Time(s):	<u>13:38</u> <u>15:31</u>
GC Column:	<u>RXI-624</u> ID: <u>0.25</u> (mm)		

LAB FILE ID:		RRF005 = VY022776.D		RRF010 = VY022777.D		RRF020 = VY022778.D		
		RRF050 = VY022779.D		RRF100 = VY022780.D		RRF150 = VY022781.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.424	0.456	0.474	0.424	0.404	0.384	0.428	7.7
Chloromethane	0.837	0.921	0.865	0.793	0.758	0.724	0.816	8.9
Vinyl Chloride	0.934	1.099	1.091	1.045	0.993	0.958	1.020	6.8
Bromomethane	0.784	0.885	0.854	0.771	0.760	0.756	0.802	6.8
Chloroethane	0.649	0.736	0.722	0.694	0.673	0.640	0.686	5.6
Trichlorofluoromethane	0.999	1.180	1.219	1.166	1.127	1.085	1.129	7
1,1,2-Trichlorotrifluoroethane	0.508	0.560	0.547	0.515	0.492	0.474	0.516	6.3
1,1-Dichloroethene	0.478	0.539	0.524	0.514	0.500	0.483	0.506	4.7
Acetone	0.117	0.124	0.114	0.095	0.096	0.087	0.105	13.9
Carbon Disulfide	1.516	1.705	1.731	1.667	1.625	1.566	1.635	5.1
Methyl tert-butyl Ether	1.173	1.398	1.396	1.435	1.460	1.405	1.378	7.5
Methyl Acetate	0.272	0.358	0.440	0.351	0.353	0.322	0.349	15.7
Methylene Chloride	0.840	0.777	0.664	0.590	0.578	0.548	0.666	17.7
trans-1,2-Dichloroethene	0.521	0.604	0.597	0.592	0.581	0.575	0.578	5.2
1,1-Dichloroethane	0.949	1.075	1.079	1.077	1.055	1.030	1.044	4.8
Cyclohexane	0.998	1.021	0.988	0.946	0.905	0.894	0.959	5.4
2-Butanone	0.145	0.160	0.160	0.153	0.156	0.147	0.154	4.4
Carbon Tetrachloride	0.439	0.498	0.507	0.491	0.492	0.491	0.486	5
cis-1,2-Dichloroethene	0.606	0.689	0.687	0.685	0.687	0.678	0.672	4.8
Bromochloromethane	0.437	0.431	0.437	0.459	0.443	0.427	0.439	2.6
Chloroform	0.986	1.130	1.099	1.096	1.084	1.059	1.076	4.6
1,1,1-Trichloroethane	0.847	0.945	0.973	0.950	0.939	0.923	0.929	4.7
Methylcyclohexane	0.543	0.589	0.610	0.618	0.608	0.611	0.596	4.7
Benzene	1.248	1.433	1.451	1.464	1.467	1.440	1.417	5.9
1,2-Dichloroethane	0.335	0.397	0.402	0.400	0.404	0.392	0.388	6.8
Trichloroethene	0.305	0.364	0.382	0.372	0.360	0.350	0.356	7.6
1,2-Dichloropropane	0.289	0.339	0.345	0.339	0.341	0.337	0.332	6.4
Bromodichloromethane	0.422	0.495	0.496	0.498	0.504	0.498	0.485	6.4
4-Methyl-2-Pentanone	0.168	0.201	0.215	0.226	0.230	0.221	0.210	10.9
Toluene	0.747	0.873	0.908	0.926	0.955	0.954	0.894	8.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>JPCL01</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q2473</u>
Instrument ID: <u>MSVOA_Y</u>	Calibration Date(s): <u>06/23/2025</u> <u>06/23/2025</u>
Heated Purge: (Y/N) <u>Y</u>	Calibration Time(s): <u>13:38</u> <u>15:31</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:								
RRF005 = VY022776.D			RRF010 = VY022777.D			RRF020 = VY022778.D		
RRF050 = VY022779.D			RRF100 = VY022780.D			RRF150 = VY022781.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.355	0.430	0.438	0.451	0.473	0.473	0.437	10
cis-1,3-Dichloropropene	0.412	0.503	0.523	0.524	0.540	0.538	0.506	9.6
1,1,2-Trichloroethane	0.207	0.249	0.249	0.253	0.255	0.250	0.244	7.4
2-Hexanone	0.115	0.140	0.145	0.151	0.157	0.149	0.143	10.5
Dibromochloromethane	0.260	0.315	0.321	0.329	0.336	0.329	0.315	8.8
1,2-Dibromoethane	0.193	0.231	0.229	0.237	0.244	0.236	0.228	7.8
Tetrachloroethene	0.399	0.465	0.535	0.515	0.473	0.446	0.472	10.3
Chlorobenzene	0.981	1.110	1.131	1.126	1.130	1.114	1.099	5.3
Ethyl Benzene	1.644	1.881	1.971	2.029	2.040	2.018	1.930	7.9
m/p-Xylenes	0.624	0.722	0.759	0.782	0.800	0.791	0.746	8.8
o-Xylene	0.578	0.674	0.708	0.734	0.759	0.765	0.703	10
Styrene	0.926	1.108	1.165	1.249	1.309	1.309	1.178	12.5
Bromoform	0.178	0.204	0.203	0.212	0.225	0.220	0.207	8
Isopropylbenzene	3.354	3.764	3.823	3.778	3.709	3.759	3.698	4.7
1,1,2,2-Tetrachloroethane	0.597	0.659	0.566	0.567	0.594	0.593	0.596	5.6
1,3-Dichlorobenzene	1.546	1.660	1.692	1.708	1.750	1.744	1.683	4.5
1,4-Dichlorobenzene	1.564	1.740	1.688	1.685	1.690	1.666	1.672	3.5
1,2-Dichlorobenzene	1.395	1.488	1.502	1.499	1.515	1.502	1.483	3
1,2-Dibromo-3-Chloropropane	0.102	0.101	0.103	0.103	0.102	0.096	0.101	2.7
1,2,4-Trichlorobenzene	0.778	0.841	0.848	0.843	0.871	0.845	0.838	3.7
1,2,3-Trichlorobenzene	0.679	0.723	0.735	0.728	0.751	0.727	0.724	3.3
1,2-Dichloroethane-d4	0.568	0.550	0.557	0.559	0.571	0.545	0.558	1.8
Dibromofluoromethane	0.306	0.297	0.295	0.304	0.314	0.308	0.304	2.3
Toluene-d8	1.182	1.148	1.186	1.215	1.262	1.247	1.207	3.6
4-Bromofluorobenzene	0.368	0.362	0.370	0.385	0.423	0.421	0.388	7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/02/2025</u> <u>10:37</u>
Lab File ID:	<u>VY022903.D</u>	Init. Calib. Date(s):	<u>06/23/2025</u> <u>06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38</u> <u>15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.411		-3.97	20
Chloromethane	0.816	0.897	0.1	9.93	20
Vinyl Chloride	1.020	1.044		2.35	20
Bromomethane	0.802	0.808		0.75	20
Chloroethane	0.686	0.718		4.66	20
Trichlorofluoromethane	1.129	1.131		0.18	20
1,1,2-Trichlorotrifluoroethane	0.516	0.529		2.52	20
1,1-Dichloroethene	0.506	0.524		3.56	20
Acetone	0.105	0.130		23.81	20
Carbon Disulfide	1.635	1.666		1.9	20
Methyl tert-butyl Ether	1.378	1.459		5.88	20
Methyl Acetate	0.349	0.342		-2.01	20
Methylene Chloride	0.666	0.654		-1.8	20
trans-1,2-Dichloroethene	0.578	0.601		3.98	20
1,1-Dichloroethane	1.044	1.127	0.1	7.95	20
Cyclohexane	0.959	1.013		5.63	20
2-Butanone	0.154	0.175		13.64	20
Carbon Tetrachloride	0.486	0.518		6.58	20
cis-1,2-Dichloroethene	0.672	0.696		3.57	20
Bromochloromethane	0.439	0.455		3.64	20
Chloroform	1.076	1.121		4.28	20
1,1,1-Trichloroethane	0.929	0.989		6.46	20
Methylcyclohexane	0.596	0.656		10.07	20
Benzene	1.417	1.521		7.34	20
1,2-Dichloroethane	0.388	0.415		6.96	20
Trichloroethene	0.356	0.379		6.46	20
1,2-Dichloropropane	0.332	0.363		9.34	20
Bromodichloromethane	0.485	0.513		5.77	20
4-Methyl-2-Pentanone	0.210	0.238		13.33	20
Toluene	0.894	0.967		8.17	20
t-1,3-Dichloropropene	0.437	0.480		9.84	20
cis-1,3-Dichloropropene	0.506	0.551		8.89	20
1,1,2-Trichloroethane	0.244	0.260		6.56	20
2-Hexanone	0.143	0.165		15.39	20
Dibromochloromethane	0.315	0.331		5.08	20
1,2-Dibromoethane	0.228	0.239		4.82	20
Tetrachloroethene	0.472	0.498		5.51	20
Chlorobenzene	1.099	1.178	0.3	7.19	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/02/2025</u> <u>10:37</u>
Lab File ID:	<u>VY022903.D</u>	Init. Calib. Date(s):	<u>06/23/2025</u> <u>06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38</u> <u>15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	2.164		12.12	20
m/p-Xylenes	0.746	0.834		11.8	20
o-Xylene	0.703	0.783		11.38	20
Styrene	1.178	1.313		11.46	20
Bromoform	0.207	0.220	0.1	6.28	20
Isopropylbenzene	3.698	4.114		11.25	20
1,1,2,2-Tetrachloroethane	0.596	0.635	0.3	6.54	20
1,3-Dichlorobenzene	1.683	1.822		8.26	20
1,4-Dichlorobenzene	1.672	1.793		7.24	20
1,2-Dichlorobenzene	1.483	1.577		6.34	20
1,2-Dibromo-3-Chloropropane	0.101	0.105		3.96	20
1,2,4-Trichlorobenzene	0.838	0.880		5.01	20
1,2,3-Trichlorobenzene	0.724	0.729		0.69	20
1,2-Dichloroethane-d4	0.558	0.550		-1.43	20
Dibromofluoromethane	0.304	0.303		-0.33	20
Toluene-d8	1.207	1.217		0.83	20
4-Bromofluorobenzene	0.388	0.389		0.26	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/03/2025 09:26</u>
Lab File ID:	<u>VY022929.D</u>	Init. Calib. Date(s):	<u>06/23/2025 06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38 15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.395		-7.71	20
Chloromethane	0.816	0.853	0.1	4.53	20
Vinyl Chloride	1.020	1.001		-1.86	20
Bromomethane	0.802	0.783		-2.37	20
Chloroethane	0.686	0.704		2.62	20
Trichlorofluoromethane	1.129	1.112		-1.51	20
1,1,2-Trichlorotrifluoroethane	0.516	0.521		0.97	20
1,1-Dichloroethene	0.506	0.499		-1.38	20
Acetone	0.105	0.122		16.19	20
Carbon Disulfide	1.635	1.601		-2.08	20
Methyl tert-butyl Ether	1.378	1.350		-2.03	20
Methyl Acetate	0.349	0.306		-12.32	20
Methylene Chloride	0.666	0.618		-7.21	20
trans-1,2-Dichloroethene	0.578	0.573		-0.87	20
1,1-Dichloroethane	1.044	1.086	0.1	4.02	20
Cyclohexane	0.959	0.963		0.42	20
2-Butanone	0.154	0.158		2.6	20
Carbon Tetrachloride	0.486	0.511		5.14	20
cis-1,2-Dichloroethene	0.672	0.674		0.3	20
Bromochloromethane	0.439	0.476		8.43	20
Chloroform	1.076	1.085		0.93	20
1,1,1-Trichloroethane	0.929	0.949		2.15	20
Methylcyclohexane	0.596	0.639		7.22	20
Benzene	1.417	1.504		6.14	20
1,2-Dichloroethane	0.388	0.408		5.16	20
Trichloroethene	0.356	0.371		4.21	20
1,2-Dichloropropane	0.332	0.354		6.63	20
Bromodichloromethane	0.485	0.508		4.74	20
4-Methyl-2-Pentanone	0.210	0.223		6.19	20
Toluene	0.894	0.947		5.93	20
t-1,3-Dichloropropene	0.437	0.460		5.26	20
cis-1,3-Dichloropropene	0.506	0.534		5.53	20
1,1,2-Trichloroethane	0.244	0.251		2.87	20
2-Hexanone	0.143	0.154		7.69	20
Dibromochloromethane	0.315	0.319		1.27	20
1,2-Dibromoethane	0.228	0.231		1.32	20
Tetrachloroethene	0.472	0.517		9.53	20
Chlorobenzene	1.099	1.171	0.3	6.55	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JPCL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2473</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>07/03/2025</u> <u>09:26</u>
Lab File ID:	<u>VY022929.D</u>	Init. Calib. Date(s):	<u>06/23/2025</u> <u>06/23/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>13:38</u> <u>15:31</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	2.129		10.31	20
m/p-Xylenes	0.746	0.813		8.98	20
o-Xylene	0.703	0.753		7.11	20
Styrene	1.178	1.277		8.4	20
Bromoform	0.207	0.207	0.1	0	20
Isopropylbenzene	3.698	4.069		10.03	20
1,1,2,2-Tetrachloroethane	0.596	0.596	0.3	0	20
1,3-Dichlorobenzene	1.683	1.801		7.01	20
1,4-Dichlorobenzene	1.672	1.727		3.29	20
1,2-Dichlorobenzene	1.483	1.511		1.89	20
1,2-Dibromo-3-Chloropropane	0.101	0.094		-6.93	20
1,2,4-Trichlorobenzene	0.838	0.802		-4.3	20
1,2,3-Trichlorobenzene	0.724	0.673		-7.04	20
1,2-Dichloroethane-d4	0.558	0.550		-1.43	20
Dibromofluoromethane	0.304	0.318		4.61	20
Toluene-d8	1.207	1.258		4.22	20
4-Bromofluorobenzene	0.388	0.393		1.29	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.