

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JPCL01
 Lab Code: ACE SDG NO.: Q2473
 Lab File ID: VY022903.D Date Analyzed: 07/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	321198	13.347				
UPPER LIMIT	642396	13.847				
LOWER LIMIT	160599	12.847				
EPA SAMPLE NO.						
PIT#1	216048	13.35				
PIT#2	220138	13.35				
PIT#3	181660	13.35				
PIT#4	244944	13.35				
VY0702SBL01	251367	13.35				
VY0702SBS01	274634	13.35				
VY0702SBSD01	268831	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.