

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JPCL01

Lab Code: ACE

SDG NO.: Q2473

Lab File ID: VY022902.D

BFB Injection Date: 07/02/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:51

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	55.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	83.6
175	5.0 - 9.0% of mass 174	6.1 (7.3) 1
176	95.0 - 101.0% of mass 174	81.3 (97.3) 1
177	5.0 - 9.0% of mass 176	5.4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022903.D	07/02/2025	10:37
VY0702SBL01	VY0702SBL01	VY022904.D	07/02/2025	11:42
VY0702SBS01	VY0702SBS01	VY022905.D	07/02/2025	12:12
VY0702SBSD01	VY0702SBSD01	VY022906.D	07/02/2025	12:35
PIT#1	Q2473-01	VY022917.D	07/02/2025	17:07
PIT#2	Q2473-02	VY022918.D	07/02/2025	17:30
PIT#3	Q2473-03	VY022919.D	07/02/2025	17:53
PIT#4	Q2473-04	VY022920.D	07/02/2025	18:17