

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/07/25 12:25		Jaswal	OK
2	S1	S1	CAL2	07/07/25 12:30		Jaswal	OK
3	S2	S2	CAL3	07/07/25 12:34		Jaswal	OK
4	S3	S3	CAL4	07/07/25 12:38		Jaswal	OK
5	S4	S4	CAL5	07/07/25 12:42		Jaswal	OK
6	S5	S5	CAL6	07/07/25 12:47		Jaswal	OK
7	ICV01	ICV01	ICV	07/07/25 13:22		Jaswal	OK
8	LLICV01	LLICV01	LLICV	07/07/25 13:41		Jaswal	OK
9	ICB01	ICB01	ICB	07/07/25 13:45		Jaswal	OK
10	CRI01	CRI01	CRDL	07/07/25 13:49		Jaswal	OK
11	ICSA01	ICSA01	ICSA	07/07/25 13:53		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	07/07/25 14:10		Jaswal	OK
13	ICSADL	ICSADL	ICSA	07/07/25 14:15		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	07/07/25 14:19		Jaswal	OK
15	CCV01	CCV01	CCV	07/07/25 14:38		Jaswal	OK
16	CCB01	CCB01	CCB	07/07/25 14:42		Jaswal	OK
17	Q2491-01	EO-1-070225	SAM	07/07/25 14:46		Jaswal	OK
18	Q2472-01DL	PN#1DL	SAM	07/07/25 14:50	Straight 5x for all elements	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

19	Q2478-01	WC-1	SAM	07/07/25 14:55		Jaswal	OK
20	Q2381-08	VNJ-207	SAM	07/07/25 14:59		Jaswal	OK
21	Q2398-09	M00-25-0179	SAM	07/07/25 15:03	Confirm TCLP (Pb)	Jaswal	Not Ok
22	Q2472-02	PN#1	SAM	07/07/25 15:07		Jaswal	OK
23	Q2474-01	FO-1	SAM	07/07/25 15:12		Jaswal	OK
24	Q2478-04	WC-1	SAM	07/07/25 15:16	Confirm TCLP (Ba)	Jaswal	Not Ok
25	Q2473-05	PIT#1	SAM	07/07/25 15:21		Jaswal	OK
26	Q2473-06	PIT#2	SAM	07/07/25 15:25		Jaswal	OK
27	CCV02	CCV02	CCV	07/07/25 15:30		Jaswal	OK
28	CCB02	CCB02	CCB	07/07/25 15:34		Jaswal	OK
29	Q2472-01DL2	PN#1DL2	SAM	07/07/25 15:38	NOT USE	Jaswal	Not Ok
30	Q2507-01	SU-04-7.3-2025	SAM	07/07/25 15:42		Jaswal	OK
31	Q2493-04	WC-11	SAM	07/07/25 15:46		Jaswal	OK
32	Q2493-04DUP	WC-11DUP	DUP	07/07/25 15:51		Jaswal	OK
33	Q2493-04L	WC-11L	SD	07/07/25 15:55		Jaswal	OK
34	Q2493-04MS	WC-11MS	MS	07/07/25 15:59		Jaswal	OK
35	Q2493-04MSD	WC-11MSD	MSD	07/07/25 16:04		Jaswal	OK
36	Q2493-04A	WC-11A	PS	07/07/25 16:08		Jaswal	OK
37	Q2477-01	50728	SAM	07/07/25 16:16	NOT USE	Jaswal	Not Ok
38	CCV03	CCV03	CCV	07/07/25 16:33		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

39	CCB03	CCB03	CCB	07/07/25 16:37		Jaswal	OK
40	PB168711BL	PB168711BL	MB	07/07/25 16:41		Jaswal	OK
41	PB168713BL	PB168713BL	MB	07/07/25 16:45		Jaswal	OK
42	PB168714BL	PB168714BL	MB	07/07/25 16:50		Jaswal	OK
43	PB168713BS	PB168713BS	LCS	07/07/25 17:01		Jaswal	OK
44	PB168714BS	PB168714BS	LCS	07/07/25 17:05		Jaswal	OK
45	Q2473-01	PIT#1	SAM	07/07/25 17:09		Jaswal	OK
46	PB168711BS	PB168711BS	LCS	07/07/25 17:14		Jaswal	OK
47	Q2473-03	PIT#3	SAM	07/07/25 17:18		Jaswal	OK
48	Q2473-04	PIT#4	SAM	07/07/25 17:22		Jaswal	OK
49	Q2484-01	TP-58	SAM	07/07/25 17:26		Jaswal	OK
50	CCV04	CCV04	CCV	07/07/25 17:31		Jaswal	OK
51	CCB04	CCB04	CCB	07/07/25 17:35		Jaswal	OK
52	Q2484-02	TP-57	SAM	07/07/25 17:39	CCV fail for TI	Jaswal	Not Ok
53	Q2484-03	TP-64	SAM	07/07/25 17:43	CCV fail for TI	Jaswal	Not Ok
54	Q2484-04	TP-107	SAM	07/07/25 17:47	CCV fail for TI	Jaswal	Not Ok
55	Q2484-05	TP-106	SAM	07/07/25 17:51	CCV fail for TI	Jaswal	Not Ok
56	Q2484-06	TP-104	SAM	07/08/25 01:19	CCV fail for TI	Jaswal	Not Ok
57	Q2484-06DUP	TP-104DUP	DUP	07/08/25 01:23	CCV fail for TI	Jaswal	Not Ok
58	Q2484-06L	TP-104L	SD	07/08/25 01:27	CCV fail for TI	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

59	Q2484-06MS	TP-104MS	MS	07/08/25 01:31	CCV fail for TI	Jaswal	Not Ok
60	Q2484-06MSD	TP-104MSD	MSD	07/08/25 01:35	CCV fail for TI	Jaswal	Not Ok
61	Q2484-06A	TP-104A	PS	07/08/25 01:39	CCV fail for TI	Jaswal	Not Ok
62	CCV05	CCV05	CCV	07/08/25 01:43	CCV fail for TI	Jaswal	OK
63	CCB05	CCB05	CCB	07/08/25 01:47		Jaswal	OK
64	Q2487-09	G4(0-6)	SAM	07/08/25 01:51	CCV fail for K,TI	Jaswal	Not Ok
65	Q2487-10	G4(6-12)	SAM	07/08/25 01:56	CCV fail for K,TI	Jaswal	Not Ok
66	Q2487-11	G3(0-6)	SAM	07/08/25 02:00	CCV fail for K,TI	Jaswal	Not Ok
67	Q2487-12	G3(6-12)	SAM	07/08/25 02:05	CCV fail for K,TI	Jaswal	Not Ok
68	Q2487-13	G2(0-6)	SAM	07/08/25 02:09	CCV fail for K,TI	Jaswal	Not Ok
69	Q2487-14	G2(6-12)	SAM	07/08/25 02:13	CCV fail for K,TI	Jaswal	Not Ok
70	Q2487-15	G1(0-6)	SAM	07/08/25 02:18	CCV fail for K,TI	Jaswal	Not Ok
71	Q2487-16	G1(6-12)	SAM	07/08/25 02:22	CCV fail for K,TI	Jaswal	Not Ok
72	Q2493-01	WC-11	SAM	07/08/25 02:41	CCV fail for K,TI	Jaswal	Not Ok
73	Q2493-01DUP	WC-11DUP	DUP	07/08/25 02:45	CCV fail for K,TI	Jaswal	Not Ok
74	CCV06	CCV06	CCV	07/08/25 02:49	CCV fail for K,TI	Jaswal	OK
75	CCB06	CCB06	CCB	07/08/25 02:53		Jaswal	OK
76	Q2493-01L	WC-11L	SD	07/08/25 02:58	CCV fail for K,TI	Jaswal	Not Ok
77	Q2493-01MS	WC-11MS	MS	07/08/25 03:02	CCV fail for K,TI	Jaswal	Not Ok
78	Q2493-01MSD	WC-11MSD	MSD	07/08/25 03:06	CCV fail for K,TI	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

79	Q2493-01A	WC-11A	PS	07/08/25 03:10	CCV fail for K,TI	Jaswal	Not Ok
80	Q2473-07	PIT#3	SAM	07/08/25 03:14		Jaswal	OK
81	Q2473-08	PIT#4	SAM	07/08/25 03:18		Jaswal	OK
82	PB168689TB	PB168689TB	MB	07/08/25 03:23		Jaswal	OK
83	Q2473-02	PIT#2	SAM	07/08/25 03:36	CCV fail for K,TI	Jaswal	Not Ok
84	CCV07	CCV07	CCV	07/08/25 03:40		Jaswal	OK
85	CCB07	CCB07	CCB	07/08/25 03:44		Jaswal	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136391

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP86288,MP86289,MP86290,MP86291,MP86292,MP86293
ICV Standard	MP86294
CCV Standard	MP86296
ICSA Standard	
CRI Standard	MP86298
LCS Standard	
Chk Standard	MP86295,MP86297,MP8699,MP86304

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/08/25 08:35		MOHAN	OK
2	S0.2	S0.2	CAL2	07/08/25 08:37		MOHAN	OK
3	S2.5	S2.5	CAL3	07/08/25 08:39		MOHAN	OK
4	S5	S5	CAL4	07/08/25 08:42		MOHAN	OK
5	S7.5	S7.5	CAL5	07/08/25 08:44		MOHAN	OK
6	S10	S10	CAL6	07/08/25 08:52		MOHAN	OK
7	ICV19	ICV19	ICV	07/08/25 08:59		MOHAN	OK
8	ICB19	ICB19	ICB	07/08/25 09:04		MOHAN	OK
9	CCV61	CCV61	CCV	07/08/25 09:07		MOHAN	OK
10	CCB61	CCB61	CCB	07/08/25 09:12		MOHAN	OK
11	CRA	CRA	CRDL	07/08/25 09:14		MOHAN	OK
12	HighStd	HighStd	HIGH STD	07/08/25 09:19		MOHAN	OK
13	ChkStd	ChkStd	SAM	07/08/25 09:21		MOHAN	OK
14	PB168741BL	PB168741BL	MB	07/08/25 09:24		MOHAN	OK
15	PB168741BS	PB168741BS	LCS	07/08/25 09:30		MOHAN	OK
16	Q2473-05	PIT#1	SAM	07/08/25 09:33		MOHAN	OK
17	Q2473-06	PIT#2	SAM	07/08/25 09:35		MOHAN	OK
18	Q2473-07	PIT#3	SAM	07/08/25 09:37		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136391

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP86288,MP86289,MP86290,MP86291,MP86292,MP86293
ICV Standard	MP86294
CCV Standard	MP86296
ICSA Standard	
CRI Standard	MP86298
LCS Standard	
Chk Standard	MP86295,MP86297,MP8699,MP86304

19	Q2473-08	PIT#4	SAM	07/08/25 09:40		MOHAN	OK
20	Q2474-01	FO-1	SAM	07/08/25 09:42		MOHAN	OK
21	CCV62	CCV62	CCV	07/08/25 09:44		MOHAN	OK
22	CCB62	CCB62	CCB	07/08/25 09:50		MOHAN	OK
23	Q2478-04	WC-1	SAM	07/08/25 09:52		MOHAN	OK
24	Q2493-04	WC-11	SAM	07/08/25 09:55		MOHAN	OK
25	Q2493-04DUP	WC-11DUP	DUP	07/08/25 09:57		MOHAN	OK
26	Q2493-04MS	WC-11MS	MS	07/08/25 10:05		MOHAN	OK
27	Q2493-04MSD	WC-11MSD	MSD	07/08/25 10:07		MOHAN	OK
28	PB168742BL	PB168742BL	MB	07/08/25 10:12		MOHAN	OK
29	PB168742BS	PB168742BS	LCS	07/08/25 10:15		MOHAN	OK
30	Q2481-01	CC0627-AL	SAM	07/08/25 10:17		MOHAN	OK
31	Q2481-01DUP	CC0627-ALDUP	DUP	07/08/25 10:19		MOHAN	OK
32	Q2481-01MS	CC0627-ALMS	MS	07/08/25 10:21		MOHAN	OK
33	CCV63	CCV63	CCV	07/08/25 10:24		MOHAN	OK
34	CCB63	CCB63	CCB	07/08/25 10:26		MOHAN	OK
35	Q2481-01MSD	CC0627-ALMSD	MSD	07/08/25 10:28		MOHAN	OK
36	Q2481-02DL	CC0627-CLOXPLDL	SAM	07/08/25 10:32	Report Straight 10X	MOHAN	OK
37	Q2481-03DL	CC0625-OBXBLDL	SAM	07/08/25 10:34	Report Straight 10X	MOHAN	OK
38	Q2481-04DL	CC0627-AOXLDL	SAM	07/08/25 10:37	Report Straight 10X	MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136391

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP86288,MP86289,MP86290,MP86291,MP86292,MP86293
ICV Standard	MP86294
CCV Standard	MP86296
ICSA Standard	
CRI Standard	MP86298
LCS Standard	
Chk Standard	MP86295,MP86297,MP8699,MP86304

39	Q2481-05	CC0625-NL	SAM	07/08/25 10:39		MOHAN	OK
40	Q2481-06DL	CC0267-OXPLDL	SAM	07/08/25 10:42	Report Straight 10X	MOHAN	OK
41	Q2481-07DL	CC0627-OXLDL	SAM	07/08/25 10:44	Report Straight 10X	MOHAN	OK
42	Q2481-08	CC0627-CLOXAL	SAM	07/08/25 10:46		MOHAN	OK
43	Q2481-09DL	CC0627-BLDL	SAM	07/08/25 10:49	Report Straight 10X	MOHAN	OK
44	Q2481-10	CC0627-SFBL	SAM	07/08/25 10:51		MOHAN	OK
45	CCV64	CCV64	CCV	07/08/25 10:53		MOHAN	OK
46	CCB64	CCB64	CCB	07/08/25 10:56		MOHAN	OK
47	PB168689TB	PB168689TB	MB	07/08/25 10:58		MOHAN	OK
48	PB168705TB	PB168705TB	MB	07/08/25 11:03		MOHAN	OK
49	Q2493-04L	WC-11L	SD	07/08/25 11:05		MOHAN	OK
50	Q2493-04A	WC-11A	PS	07/08/25 11:07		MOHAN	OK
51	Q2481-01L	CC0627-ALL	SD	07/08/25 11:10		MOHAN	OK
52	Q2481-01A	CC0627-ALA	PS	07/08/25 11:12		MOHAN	OK
53	CCV65	CCV65	CCV	07/08/25 11:14		MOHAN	OK
54	CCB65	CCB65	CCB	07/08/25 11:17		MOHAN	OK