



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Parkway Generation Operating LLC

SDG No.: Q2474

Contract: PSEG04

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	23.0	20.0	115	65 - 135	P	07/07/2025	13:49	LB136390
	Barium	97.5	100	98	65 - 135	P	07/07/2025	13:49	LB136390
	Cadmium	7.08	6.0	118	65 - 135	P	07/07/2025	13:49	LB136390
	Chromium	11.8	10.0	118	65 - 135	P	07/07/2025	13:49	LB136390
	Lead	13.7	12.0	114	65 - 135	P	07/07/2025	13:49	LB136390
	Selenium	19.2	20.0	96	65 - 135	P	07/07/2025	13:49	LB136390
	Silver	10.6	10.0	106	65 - 135	P	07/07/2025	13:49	LB136390
CRA	Mercury	0.25	0.2	125	70 - 130	CV	07/08/2025	09:14	LB136391