



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q2474

Contract: PSEG04

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB19	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	09:04	LB136391

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q2474

Contract: PSEG04

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	09:12	LB136391
CCB62	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	09:50	LB136391
CCB63	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	10:26	LB136391
CCB64	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	10:56	LB136391
CCB65	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	11:17	LB136391

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q2474

Contract: PSEG04

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	07/07/2025	13:45	LB136390
	Barium	14.6	+/-50	U	100	P	07/07/2025	13:45	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/07/2025	13:45	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/07/2025	13:45	LB136390
	Lead	2.30	+/-6	U	12.0	P	07/07/2025	13:45	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/07/2025	13:45	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/07/2025	13:45	LB136390

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q2474

Contract: PSEG04

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	07/07/2025	14:42	LB136390
	Barium	14.6	+/-50	U	100	P	07/07/2025	14:42	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/07/2025	14:42	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/07/2025	14:42	LB136390
	Lead	2.30	+/-6	U	12.0	P	07/07/2025	14:42	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/07/2025	14:42	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/07/2025	14:42	LB136390
CCB02	Arsenic	5.12	+/-10	U	20.0	P	07/07/2025	15:34	LB136390
	Barium	14.6	+/-50	U	100	P	07/07/2025	15:34	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/07/2025	15:34	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/07/2025	15:34	LB136390
	Lead	2.30	+/-6	U	12.0	P	07/07/2025	15:34	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/07/2025	15:34	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/07/2025	15:34	LB136390
CCB03	Arsenic	5.12	+/-10	U	20.0	P	07/07/2025	16:37	LB136390
	Barium	14.6	+/-50	U	100	P	07/07/2025	16:37	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/07/2025	16:37	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/07/2025	16:37	LB136390
	Lead	2.30	+/-6	U	12.0	P	07/07/2025	16:37	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/07/2025	16:37	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/07/2025	16:37	LB136390
CCB04	Arsenic	5.12	+/-10	U	20.0	P	07/07/2025	17:35	LB136390
	Barium	14.6	+/-50	U	100	P	07/07/2025	17:35	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/07/2025	17:35	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/07/2025	17:35	LB136390
	Lead	2.30	+/-6	U	12.0	P	07/07/2025	17:35	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/07/2025	17:35	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/07/2025	17:35	LB136390
CCB05	Arsenic	5.12	+/-10	U	20.0	P	07/08/2025	01:47	LB136390
	Barium	14.6	+/-50	U	100	P	07/08/2025	01:47	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/08/2025	01:47	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/08/2025	01:47	LB136390
	Lead	2.30	+/-6	U	12.0	P	07/08/2025	01:47	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/08/2025	01:47	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/08/2025	01:47	LB136390
CCB06	Arsenic	5.12	+/-10	U	20.0	P	07/08/2025	02:53	LB136390
	Barium	14.6	+/-50	U	100	P	07/08/2025	02:53	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/08/2025	02:53	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/08/2025	02:53	LB136390

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q2474

Contract: PSEG04

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	07/08/2025	02:53	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/08/2025	02:53	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/08/2025	02:53	LB136390
CCB07	Arsenic	5.12	+/-10	U	20.0	P	07/08/2025	03:44	LB136390
	Barium	14.6	+/-50	U	100	P	07/08/2025	03:44	LB136390
	Cadmium	0.50	+/-3	U	6.00	P	07/08/2025	03:44	LB136390
	Chromium	2.12	+/-5	U	10.0	P	07/08/2025	03:44	LB136390
	Lead	2.30	+/-6	U	12.0	P	07/08/2025	03:44	LB136390
	Selenium	9.64	+/-10	U	20.0	P	07/08/2025	03:44	LB136390
	Silver	1.62	+/-5	U	10.0	P	07/08/2025	03:44	LB136390