

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC

SDG No.: Q2474

Contract: PSEG04

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-4.19			-20	20	07/07/2025	13:53	LB136390
	Barium	-3.69	6.0	62	-94	106	07/07/2025	13:53	LB136390
	Cadmium	-0.042	1.0	4	-5	7	07/07/2025	13:53	LB136390
	Chromium	53.6	52.0	103	42	62	07/07/2025	13:53	LB136390
	Lead	-5.45			-12	12	07/07/2025	13:53	LB136390
	Selenium	8.20			-20	20	07/07/2025	13:53	LB136390
	Silver	-6.96			-10	10	07/07/2025	13:53	LB136390
ICSAB01	Arsenic	99.1	104	95	88.4	120	07/07/2025	14:10	LB136390
	Barium	455	537	85	437	637	07/07/2025	14:10	LB136390
	Cadmium	992	972	102	826	1120	07/07/2025	14:10	LB136390
	Chromium	558	542	103	460	624	07/07/2025	14:10	LB136390
	Lead	41.5	49.0	85	37	61	07/07/2025	14:10	LB136390
	Selenium	53.0	46.0	115	26	66	07/07/2025	14:10	LB136390
	Silver	208	201	104	170	232	07/07/2025	14:10	LB136390