

Continuing Calibration Report for SequenceID : FC070225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069355.D

Aliphatic C9-C12	9339051.000	60.000	3.304	6.604	155650.850	146679.627	-6.116
Aliphatic C12-C16	6424914.000	40.000	6.605	10.008	160622.850	159210.532	-0.887
Aliphatic C16-C21	9216722.000	60.000	10.009	13.377	153612.033	154434.323	0.532
Aliphatic C21-C28	11012797.000	80.000	13.378	17.043	137659.963	133966.368	-2.757
Aliphatic C28-C40	12367304.000	120.000	17.044	22.026	103060.867	92124.417	-11.871
Aliphatic EPH	48360788.000	360.000	3.304	22.026	134335.522	128354.160	-4.660