

Continuing Calibration Report for SequenceID : FC070225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069365.D

Aliphatic C9-C12	8532118.000	60.000	3.304	6.604	142201.967	146679.627	3.053
Aliphatic C12-C16	5790506.000	40.000	6.605	10.008	144762.650	159210.532	9.075
Aliphatic C16-C21	8092528.000	60.000	10.009	13.377	134875.467	154434.323	12.665
Aliphatic C21-C28	9462211.000	80.000	13.378	17.043	118277.638	133966.368	11.711
Aliphatic C28-C40	10961952.000	120.000	17.044	22.026	91349.600	92124.417	0.841
Aliphatic EPH	42839315.000	360.000	3.304	22.026	118998.097	128354.160	7.289