

Continuing Calibration Report for SequenceID : FE070225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054681.D

Aliphatic C9-C12	9050829.000	60.000	3.319	6.951	150847.150	136055.079	-10.872
Aliphatic C12-C16	6254036.000	40.000	6.952	10.402	156350.900	140568.128	-11.228
Aliphatic C16-C21	9238520.000	60.000	10.403	13.777	153975.333	144393.643	-6.636
Aliphatic C21-C28	12968372.000	80.000	13.778	17.448	162104.650	144651.717	-12.065
Aliphatic C28-C40	16599930.000	120.000	17.449	22.462	138332.750	138647.492	0.227
Aliphatic EPH	54111687.000	360.000	3.319	22.462	150310.242	140720.791	-6.815