

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	-0.01
2 T	Dichlorodifluoromethane	0.428	0.393	8.2	95	0.00
3 P	Chloromethane	0.816	0.799	2.1	104	0.00
4 C	Vinyl Chloride	1.020	1.000	2.0#	99	0.00
5 T	Bromomethane	0.802	0.776	3.2	104	-0.01
6 T	Chloroethane	0.686	0.661	3.6	98	-0.01
7 T	Trichlorofluoromethane	1.129	1.081	4.3	96	-0.01
8 T	Diethyl Ether	0.279	0.268	3.9	97	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.516	0.488	5.4	97	-0.01
10 T	Methyl Iodide	0.562	0.527	6.2	87	-0.01
11 T	Tert butyl alcohol	0.037	0.038	-2.7	102	-0.02
12 CM	1,1-Dichloroethene	0.506	0.485	4.2#	97	-0.02
13 T	Acrolein	0.050	0.042	16.0	86	-0.01
14 T	Allyl chloride	0.778	0.745	4.2	96	-0.01
15 T	Acrylonitrile	0.116	0.116	0.0	99	-0.02
16 T	Acetone	0.105	0.091	13.3	98	-0.02
17 T	Carbon Disulfide	1.635	1.570	4.0	97	-0.01
18 T	Methyl Acetate	0.349	0.353	-1.1	104	-0.01
19 T	Methyl tert-butyl Ether	1.378	1.360	1.3	98	-0.02
20 T	Methylene Chloride	0.666	0.581	12.8	101	-0.01
21 T	trans-1,2-Dichloroethene	0.578	0.556	3.8	97	-0.01
22 T	Diisopropyl ether	1.729	1.668	3.5	96	-0.01
23 T	Vinyl Acetate	0.955	0.943	1.3	96	-0.02
24 P	1,1-Dichloroethane	1.044	1.014	2.9	97	-0.01
25 T	2-Butanone	0.154	0.147	4.5	99	-0.01
26 T	2,2-Dichloropropane	0.875	0.808	7.7	94	-0.01
27 T	cis-1,2-Dichloroethene	0.672	0.643	4.3	97	-0.01
28 T	Bromochloromethane	0.439	0.440	-0.2	99	0.00
29 T	Tetrahydrofuran	0.097	0.098	-1.0	99	-0.01
30 C	Chloroform	1.076	1.043	3.1#	98	-0.01
31 T	Cyclohexane	0.959	0.880	8.2	96	-0.01
32 T	1,1,1-Trichloroethane	0.929	0.899	3.2	97	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.559	-0.2	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.304	0.306	-0.7	103	-0.01
36 T	1,1-Dichloropropene	0.461	0.440	4.6	96	-0.01
37 T	Ethyl Acetate	0.199	0.198	0.5	99	-0.01
38 T	Carbon Tetrachloride	0.486	0.463	4.7	97	0.00
39 T	Methylcyclohexane	0.596	0.575	3.5	96	-0.01
40 TM	Benzene	1.417	1.371	3.2	96	-0.01
41 T	Methacrylonitrile	0.122	0.132	-8.2	113	-0.01
42 TM	1,2-Dichloroethane	0.388	0.381	1.8	98	-0.01
43 T	Isopropyl Acetate	0.413	0.413	0.0	97	0.00
44 TM	Trichloroethene	0.356	0.341	4.2	94	0.00
45 C	1,2-Dichloropropane	0.332	0.322	3.0#	97	0.00
46 T	Dibromomethane	0.188	0.181	3.7	97	0.00
47 T	Bromodichloromethane	0.485	0.475	2.1	98	0.00
48 T	Methyl methacrylate	0.199	0.204	-2.5	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	-0.02
50 S	Toluene-d8	1.207	1.201	0.5	101	0.00
51 T	4-Methyl-2-Pentanone	0.210	0.215	-2.4	98	0.00
52 CM	Toluene	0.894	0.874	2.2#	97	0.00
53 T	t-1,3-Dichloropropene	0.437	0.432	1.1	98	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.502	0.8	98	0.00
55 T	1,1,2-Trichloroethane	0.244	0.241	1.2	98	0.00
56 T	Ethyl methacrylate	0.328	0.335	-2.1	97	-0.01
57 T	1,3-Dichloropropane	0.425	0.422	0.7	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.169	-16.6	102	0.00
59 T	2-Hexanone	0.143	0.144	-0.7	98	0.00
60 T	Dibromochloromethane	0.315	0.315	0.0	98	0.00
61 T	1,2-Dibromoethane	0.228	0.225	1.3	98	0.00
62 S	4-Bromofluorobenzene	0.388	0.391	-0.8	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.472	0.428	9.3	86	0.00
65 PM	Chlorobenzene	1.099	1.057	3.8	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.357	4.0	96	0.00
67 C	Ethyl Benzene	1.930	1.904	1.3#	98	0.00
68 T	m/p-Xylenes	0.746	0.735	1.5	98	-0.01
69 T	o-Xylene	0.703	0.698	0.7	99	0.00
70 T	Styrene	1.178	1.175	0.3	98	0.00
71 P	Bromoform	0.207	0.200	3.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.698	3.519	4.8	97	0.00
74 T	N-amyl acetate	0.830	0.834	-0.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	0.596	0.594	0.3	109	0.00
76 T	1,2,3-Trichloropropane	0.510	0.478	6.3	98	0.00
77 T	Bromobenzene	0.838	0.801	4.4	98	0.00
78 T	n-propylbenzene	4.465	4.296	3.8	98	0.00
79 T	2-Chlorotoluene	2.522	2.426	3.8	98	0.00
80 T	1,3,5-Trimethylbenzene	2.985	2.928	1.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.184	8.9	97	0.00
82 T	4-Chlorotoluene	2.650	2.558	3.5	99	0.00
83 T	tert-Butylbenzene	2.633	2.553	3.0	98	0.00
84 T	1,2,4-Trimethylbenzene	2.989	2.946	1.4	99	0.00
85 T	sec-Butylbenzene	3.961	3.835	3.2	98	0.00
86 T	p-Isopropyltoluene	3.296	3.222	2.2	98	0.00
87 T	1,3-Dichlorobenzene	1.683	1.626	3.4	99	0.00
88 T	1,4-Dichlorobenzene	1.672	1.600	4.3	99	0.00
89 T	n-Butylbenzene	3.101	3.015	2.8	98	0.00
90 T	Hexachloroethane	0.657	0.636	3.2	98	0.00
91 T	1,2-Dichlorobenzene	1.483	1.425	3.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.099	2.0	100	0.00
93 T	1,2,4-Trichlorobenzene	0.838	0.818	2.4	101	0.00
94 T	Hexachlorobutadiene	0.474	0.437	7.8	98	0.00
95 T	Naphthalene	1.515	1.555	-2.6	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022783.D
Acq On : 23 Jun 2025 16:17
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY062325

Quant Time: Jun 24 03:36:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 03:08:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.724	0.702	3.0	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6