

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	103	-0.01
2 T	Dichlorodifluoromethane	50.000	45.936	8.1	95	0.00
3 P	Chloromethane	50.000	48.905	2.2	104	0.00
4 C	Vinyl Chloride	50.000	49.034	1.9#	99	0.00
5 T	Bromomethane	50.000	48.375	3.3	104	-0.01
6 T	Chloroethane	50.000	48.206	3.6	98	-0.01
7 T	Trichlorofluoromethane	50.000	47.875	4.3	96	-0.01
8 T	Diethyl Ether	50.000	48.097	3.8	97	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.248	5.5	97	-0.01
10 T	Methyl Iodide	50.000	46.847	6.3	87	-0.01
11 T	Tert butyl alcohol	250.000	253.906	-1.6	102	-0.02
12 CM	1,1-Dichloroethene	50.000	47.868	4.3#	97	-0.02
13 T	Acrolein	250.000	206.396	17.4	86	-0.01
14 T	Allyl chloride	50.000	47.825	4.3	96	-0.01
15 T	Acrylonitrile	250.000	248.823	0.5	99	-0.02
16 T	Acetone	250.000	214.694	14.1	98	-0.02
17 T	Carbon Disulfide	50.000	48.003	4.0	97	-0.01
18 T	Methyl Acetate	50.000	50.527	-1.1	104	-0.01
19 T	Methyl tert-butyl Ether	50.000	49.329	1.3	98	-0.02
20 T	Methylene Chloride	50.000	43.635	12.7	101	-0.01
21 T	trans-1,2-Dichloroethene	50.000	48.066	3.9	97	-0.01
22 T	Diisopropyl ether	50.000	48.237	3.5	96	-0.01
23 T	Vinyl Acetate	250.000	247.008	1.2	96	-0.02
24 P	1,1-Dichloroethane	50.000	48.573	2.9	97	-0.01
25 T	2-Butanone	250.000	239.723	4.1	99	-0.01
26 T	2,2-Dichloropropane	50.000	46.171	7.7	94	-0.01
27 T	cis-1,2-Dichloroethene	50.000	47.867	4.3	97	-0.01
28 T	Bromochloromethane	50.000	50.087	-0.2	99	0.00
29 T	Tetrahydrofuran	250.000	252.693	-1.1	99	-0.01
30 C	Chloroform	50.000	48.491	3.0#	98	-0.01
31 T	Cyclohexane	50.000	45.926	8.1	96	-0.01
32 T	1,1,1-Trichloroethane	50.000	48.388	3.2	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.109	-0.2	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	50.304	-0.6	103	-0.01
36 T	1,1-Dichloropropene	50.000	47.735	4.5	96	-0.01
37 T	Ethyl Acetate	50.000	49.894	0.2	99	-0.01
38 T	Carbon Tetrachloride	50.000	47.632	4.7	97	0.00
39 T	Methylcyclohexane	50.000	48.226	3.5	96	-0.01
40 TM	Benzene	50.000	48.376	3.2	96	-0.01
41 T	Methacrylonitrile	50.000	53.720	-7.4	113	-0.01
42 TM	1,2-Dichloroethane	50.000	49.108	1.8	98	-0.01
43 T	Isopropyl Acetate	50.000	49.983	0.0	97	0.00
44 TM	Trichloroethene	50.000	47.904	4.2	94	0.00
45 C	1,2-Dichloropropane	50.000	48.510	3.0#	97	0.00
46 T	Dibromomethane	50.000	48.070	3.9	97	0.00
47 T	Bromodichloromethane	50.000	48.872	2.3	98	0.00
48 T	Methyl methacrylate	50.000	51.285	-2.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY062325

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	983.335	1.7	97	-0.02
50 S	Toluene-d8	50.000	49.755	0.5	101	0.00
51 T	4-Methyl-2-Pentanone	250.000	256.103	-2.4	98	0.00
52 CM	Toluene	50.000	48.894	2.2#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	49.428	1.1	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.547	0.9	98	0.00
55 T	1,1,2-Trichloroethane	50.000	49.380	1.2	98	0.00
56 T	Ethyl methacrylate	50.000	51.054	-2.1	97	-0.01
57 T	1,3-Dichloropropane	50.000	49.568	0.9	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	245.249	1.9	102	0.00
59 T	2-Hexanone	250.000	251.999	-0.8	98	0.00
60 T	Dibromochloromethane	50.000	49.955	0.1	98	0.00
61 T	1,2-Dibromoethane	50.000	49.399	1.2	98	0.00
62 S	4-Bromofluorobenzene	50.000	50.342	-0.7	104	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	104	0.00
64 T	Tetrachloroethene	50.000	45.257	9.5	86	0.00
65 PM	Chlorobenzene	50.000	48.095	3.8	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.973	4.1	96	0.00
67 C	Ethyl Benzene	50.000	49.316	1.4#	98	0.00
68 T	m/p-Xylenes	100.000	98.506	1.5	98	-0.01
69 T	o-Xylene	50.000	49.631	0.7	99	0.00
70 T	Styrene	50.000	49.897	0.2	98	0.00
71 P	Bromoform	50.000	48.380	3.2	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	47.581	4.8	97	0.00
74 T	N-amyl acetate	50.000	50.258	-0.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.808	0.4	109	0.00
76 T	1,2,3-Trichloropropane	50.000	46.775	6.5	98	0.00
77 T	Bromobenzene	50.000	47.796	4.4	98	0.00
78 T	n-propylbenzene	50.000	48.108	3.8	98	0.00
79 T	2-Chlorotoluene	50.000	48.080	3.8	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.044	1.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.554	8.9	97	0.00
82 T	4-Chlorotoluene	50.000	48.276	3.4	99	0.00
83 T	tert-Butylbenzene	50.000	48.481	3.0	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.278	1.4	99	0.00
85 T	sec-Butylbenzene	50.000	48.405	3.2	98	0.00
86 T	p-Isopropyltoluene	50.000	48.873	2.3	98	0.00
87 T	1,3-Dichlorobenzene	50.000	48.286	3.4	99	0.00
88 T	1,4-Dichlorobenzene	50.000	47.844	4.3	99	0.00
89 T	n-Butylbenzene	50.000	48.623	2.8	98	0.00
90 T	Hexachloroethane	50.000	48.434	3.1	98	0.00
91 T	1,2-Dichlorobenzene	50.000	48.036	3.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.145	1.7	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.842	2.3	101	0.00
94 T	Hexachlorobutadiene	50.000	46.178	7.6	98	0.00
95 T	Naphthalene	50.000	51.347	-2.7	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022783.D
Acq On : 23 Jun 2025 16:17
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY062325

Quant Time: Jun 24 03:36:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 03:08:29 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.503	3.0	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6