

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2480</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/14/2025 09:16</u>
Lab File ID:	<u>VX046961.D</u>	Init. Calib. Date(s):	<u>07/02/2025 07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:11 14:31</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.478	0.487		1.88	20
Chloromethane	0.522	0.477	0.1	-8.62	20
Vinyl Chloride	0.569	0.565		-0.7	20
Bromomethane	0.366	0.365		-0.27	20
Chloroethane	0.376	0.357		-5.05	20
Trichlorofluoromethane	0.880	0.919		4.43	20
1,1,2-Trichlorotrifluoroethane	0.561	0.585		4.28	20
1,1-Dichloroethene	0.542	0.553		2.03	20
Acetone	0.205	0.234		14.15	20
Carbon Disulfide	1.422	1.298		-8.72	20
Methyl tert-butyl Ether	1.531	1.868		22.01	20
Methyl Acetate	0.541	0.769		42.14	20
Methylene Chloride	0.607	0.629		3.62	20
trans-1,2-Dichloroethene	0.555	0.557		0.36	20
1,1-Dichloroethane	1.064	1.127	0.1	5.92	20
Cyclohexane	0.944	0.919		-2.65	20
2-Butanone	0.274	0.340		24.09	20
Carbon Tetrachloride	0.484	0.506		4.55	20
cis-1,2-Dichloroethene	0.677	0.712		5.17	20
Bromochloromethane	0.525	0.551		4.95	20
Chloroform	1.087	1.141		4.97	20
1,1,1-Trichloroethane	0.908	0.987		8.7	20
Methylcyclohexane	0.573	0.556		-2.97	20
Benzene	1.385	1.391		0.43	20
1,2-Dichloroethane	0.470	0.497		5.74	20
Trichloroethene	0.361	0.355		-1.66	20
1,2-Dichloropropane	0.350	0.364		4	20
Bromodichloromethane	0.518	0.552		6.56	20
4-Methyl-2-Pentanone	0.353	0.436		23.51	20
Toluene	0.858	0.875		1.98	20
t-1,3-Dichloropropene	0.462	0.534		15.58	20
cis-1,3-Dichloropropene	0.527	0.581		10.25	20
1,1,2-Trichloroethane	0.320	0.348		8.75	20
2-Hexanone	0.234	0.300		28.2	20
Dibromochloromethane	0.383	0.413		7.83	20
1,2-Dibromoethane	0.321	0.347		8.1	20
Tetrachloroethene	0.333	0.320		-3.9	20
Chlorobenzene	1.081	1.090	0.3	0.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2480</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/14/2025</u> <u>09:16</u>
Lab File ID:	<u>VX046961.D</u>	Init. Calib. Date(s):	<u>07/02/2025</u> <u>07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:11</u> <u>14:31</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.887		1.95	20
m/p-Xylenes	0.698	0.709		1.58	20
o-Xylene	0.674	0.690		2.37	20
Styrene	1.153	1.208		4.77	20
Bromoform	0.270	0.301	0.1	11.48	20
Isopropylbenzene	3.515	3.763		7.05	20
1,1,2,2-Tetrachloroethane	0.966	1.107	0.3	14.6	20
1,3-Dichlorobenzene	1.637	1.719		5.01	20
1,4-Dichlorobenzene	1.704	1.738		2	20
1,2-Dichlorobenzene	1.587	1.657		4.41	20
1,2-Dibromo-3-Chloropropane	0.169	0.217		28.4	20
1,2,4-Trichlorobenzene	1.078	1.145		6.22	20
1,2,3-Trichlorobenzene	1.019	1.115		9.42	20
1,2-Dichloroethane-d4	0.661	0.674		1.97	20
Dibromofluoromethane	0.339	0.336		-0.88	20
Toluene-d8	1.197	1.136		-5.1	20
4-Bromofluorobenzene	0.455	0.448		-1.54	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.