

Method Path : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\
 Method File : 82Y062325S.M
 Title : SW846 8260
 Last Update : Tue Jun 24 08:29:52 2025
 Response Via : Initial Calibration

Calibration Files

5 =VY022776.D 10 =VY022777.D 20 =VY022778.D 50 =VY022779.D 100 =VY022780.D 150 =VY022781.D

	Compound	5	10	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.424	0.456	0.474	0.424	0.404	0.384	0.428	7.74
3) P	Chloromethane	0.837	0.921	0.865	0.793	0.758	0.724	0.816	8.89
4) C	Vinyl Chloride	0.934	1.099	1.091	1.045	0.993	0.958	1.020	6.78#
5) T	Bromomethane	0.784	0.885	0.854	0.771	0.760	0.756	0.802	6.77
6) T	Chloroethane	0.649	0.736	0.722	0.694	0.673	0.640	0.686	5.64
7) T	Trichlorofluor...	0.999	1.180	1.219	1.166	1.127	1.085	1.129	6.96
8) T	Diethyl Ether	0.260	0.289	0.287	0.285	0.281	0.271	0.279	3.93
9) T	1,1,2-Trichlor...	0.508	0.560	0.547	0.515	0.492	0.474	0.516	6.35
10) T	Methyl Iodide	0.451	0.542	0.567	0.626	0.611	0.575	0.562	11.12
11) T	Tert butyl alc...	0.034	0.037	0.038	0.038	0.039	0.036	0.037	4.97
12) CM	1,1-Dichloroet...	0.478	0.539	0.524	0.514	0.500	0.483	0.506	4.68#
13) T	Acrolein	0.052	0.051	0.053	0.049	0.049	0.047	0.050	4.22
14) T	Allyl chloride	0.687	0.803	0.805	0.797	0.790	0.789	0.778	5.80
15) T	Acrylonitrile	0.105	0.116	0.120	0.121	0.122	0.116	0.116	5.45
16) T	Acetone	0.117	0.124	0.114	0.095	0.096	0.087	0.105	13.91
17) T	Carbon Disulfide	1.516	1.705	1.731	1.667	1.625	1.566	1.635	5.06
18) T	Methyl Acetate	0.272	0.358	0.440	0.351	0.353	0.322	0.349	15.66
19) T	Methyl tert-bu...	1.173	1.398	1.396	1.435	1.460	1.405	1.378	7.50
20) T	Methylene Chlo...	0.840	0.777	0.664	0.590	0.578	0.548	0.666	17.74
21) T	trans-1,2-Dich...	0.521	0.604	0.597	0.592	0.581	0.575	0.578	5.19
22) T	Diisopropyl ether	1.460	1.762	1.779	1.789	1.804	1.778	1.729	7.65
23) T	Vinyl Acetate	0.830	0.942	0.920	1.010	1.024	1.003	0.955	7.72
24) P	1,1-Dichloroet...	0.949	1.075	1.079	1.077	1.055	1.030	1.044	4.83
25) T	2-Butanone	0.145	0.160	0.160	0.153	0.156	0.147	0.154	4.37
26) T	2,2-Dichloropr...	0.801	0.930	0.927	0.884	0.863	0.847	0.875	5.65
27) T	cis-1,2-Dichlo...	0.606	0.689	0.687	0.685	0.687	0.678	0.672	4.84
28) T	Bromochloromet...	0.437	0.431	0.437	0.459	0.443	0.427	0.439	2.59
29) T	Tetrahydrofuran	0.087	0.095	0.098	0.102	0.103	0.097	0.097	6.02
30) C	Chloroform	0.986	1.130	1.099	1.096	1.084	1.059	1.076	4.60#
31) T	Cyclohexane	0.998	1.021	0.988	0.946	0.905	0.894	0.959	5.41
32) T	1,1,1-Trichlor...	0.847	0.945	0.973	0.950	0.939	0.923	0.929	4.68
33) S	1,2-Dichloroet...	0.568	0.550	0.557	0.559	0.571	0.545	0.558	1.77
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...	0.306	0.297	0.295	0.304	0.314	0.308	0.304	2.31
36) T	1,1-Dichloropr...	0.420	0.473	0.472	0.469	0.468	0.465	0.461	4.42
37) T	Ethyl Acetate	0.181	0.198	0.198	0.206	0.210	0.200	0.199	5.09
38) T	Carbon Tetrach...	0.439	0.498	0.507	0.491	0.492	0.491	0.486	4.98
39) T	Methylcyclohexane	0.543	0.589	0.610	0.618	0.608	0.611	0.596	4.65
40) TM	Benzene	1.248	1.433	1.451	1.464	1.467	1.440	1.417	5.92
41) T	Methacrylonitrile	0.118	0.137	0.129	0.120	0.108	0.123	0.122	8.16
42) TM	1,2-Dichloroet...	0.335	0.397	0.402	0.400	0.404	0.392	0.388	6.79
43) T	Isopropyl Acetate	0.348	0.408	0.424	0.436	0.442	0.424	0.413	8.29
44) TM	Trichloroethene	0.305	0.364	0.382	0.372	0.360	0.350	0.356	7.60
45) C	1,2-Dichloropr...	0.289	0.339	0.345	0.339	0.341	0.337	0.332	6.35#
46) T	Dibromomethane	0.169	0.193	0.192	0.191	0.194	0.190	0.188	4.94
47) T	Bromodichlorom...	0.422	0.495	0.496	0.498	0.504	0.498	0.485	6.40
48) T	Methyl methacr...	0.151	0.193	0.208	0.214	0.219	0.207	0.199	12.57
49) T	1,4-Dioxane	0.002	0.002	0.002	0.002	0.002	0.002	0.002	8.51
50) S	Toluene-d8	1.182	1.148	1.186	1.215	1.262	1.247	1.207	3.57
51) T	4-Methyl-2-Pen...	0.168	0.201	0.215	0.226	0.230	0.221	0.210	10.86
52) CM	Toluene	0.747	0.873	0.908	0.926	0.955	0.954	0.894	8.76#
53) T	t-1,3-Dichloro...	0.355	0.430	0.438	0.451	0.473	0.473	0.437	10.03
54) T	cis-1,3-Dichlo...	0.412	0.503	0.523	0.524	0.540	0.538	0.506	9.55
55) T	1,1,2-Trichlor...	0.207	0.249	0.249	0.253	0.255	0.250	0.244	7.40
56) T	Ethyl methacry...	0.258	0.296	0.315	0.354	0.372	0.374	0.328	14.17


Method Path : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\
 Method File : 82Y062325S.M

57)	T	1,3-Dichloropr...	0.371	0.428	0.439	0.438	0.442	0.433	0.425	6.39
58)	T	2-Chloroethyl ...	0.044	0.145	0.154	0.169	0.181	0.179	0.145	35.44
59)	T	2-Hexanone	0.115	0.140	0.145	0.151	0.157	0.149	0.143	10.48
60)	T	Dibromochlorom...	0.260	0.315	0.321	0.329	0.336	0.329	0.315	8.82
61)	T	1,2-Dibromoethane	0.193	0.231	0.229	0.237	0.244	0.236	0.228	7.81
62)	S	4-Bromofluorob...	0.368	0.362	0.370	0.385	0.423	0.421	0.388	7.02
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.399	0.465	0.535	0.515	0.473	0.446	0.472	10.30
65)	PM	Chlorobenzene	0.981	1.110	1.131	1.126	1.130	1.114	1.099	5.30
66)	T	1,1,1,2-Tetrac...	0.315	0.377	0.377	0.385	0.393	0.386	0.372	7.76
67)	C	Ethyl Benzene	1.644	1.881	1.971	2.029	2.040	2.018	1.930	7.88#
68)	T	m/p-Xylenes	0.624	0.722	0.759	0.782	0.800	0.791	0.746	8.85
69)	T	o-Xylene	0.578	0.674	0.708	0.734	0.759	0.765	0.703	9.99
70)	T	Styrene	0.926	1.108	1.165	1.249	1.309	1.309	1.178	12.49
71)	P	Bromoform	0.178	0.204	0.203	0.212	0.225	0.220	0.207	8.02
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.354	3.764	3.823	3.778	3.709	3.759	3.698	4.67
74)	T	N-amyl acetate	0.680	0.794	0.814	0.897	0.896	0.900	0.830	10.47
75)	P	1,1,2,2-Tetrac...	0.597	0.659	0.566	0.567	0.594	0.593	0.596	5.65
76)	T	1,2,3-Trichlor...	0.559	0.516	0.514	0.510	0.489	0.474	0.510	5.67
77)	T	Bromobenzene	0.762	0.859	0.854	0.852	0.847	0.855	0.838	4.46
78)	T	n-propylbenzene	4.147	4.548	4.635	4.562	4.453	4.444	4.465	3.84
79)	T	2-Chlorotoluene	2.325	2.536	2.601	2.576	2.556	2.541	2.522	3.95
80)	T	1,3,5-Trimethy...	2.664	2.991	3.077	3.062	3.068	3.049	2.985	5.37
81)	T	trans-1,4-Dich...	0.179	0.208	0.210	0.199	0.212	0.205	0.202	5.96
82)	T	4-Chlorotoluene	2.411	2.686	2.688	2.697	2.718	2.698	2.650	4.43
83)	T	tert-Butylbenzene	2.410	2.643	2.647	2.720	2.670	2.706	2.633	4.31
84)	T	1,2,4-Trimethy...	2.551	3.002	3.081	3.098	3.101	3.099	2.989	7.28
85)	T	sec-Butylbenzene	3.629	3.998	4.083	4.097	4.016	3.943	3.961	4.35
86)	T	p-Isopropyltol...	2.887	3.282	3.338	3.414	3.442	3.415	3.296	6.34
87)	T	1,3-Dichlorobe...	1.546	1.660	1.692	1.708	1.750	1.744	1.683	4.47
88)	T	1,4-Dichlorobe...	1.564	1.740	1.688	1.685	1.690	1.666	1.672	3.50
89)	T	n-Butylbenzene	2.814	3.124	3.194	3.204	3.160	3.107	3.101	4.69
90)	T	Hexachloroethane	0.618	0.664	0.670	0.673	0.654	0.659	0.657	3.04
91)	T	1,2-Dichlorobe...	1.395	1.488	1.502	1.499	1.515	1.502	1.483	3.00
92)	T	1,2-Dibromo-3-...	0.102	0.101	0.103	0.103	0.102	0.096	0.101	2.72
93)	T	1,2,4-Trichlor...	0.778	0.841	0.848	0.843	0.871	0.845	0.838	3.74
94)	T	Hexachlorobuta...	0.516	0.499	0.489	0.464	0.449	0.424	0.474	7.22
95)	T	Naphthalene	1.240	1.383	1.516	1.599	1.696	1.655	1.515	11.53
96)	T	1,2,3-Trichlor...	0.679	0.723	0.735	0.728	0.751	0.727	0.724	3.31

(#) = Out of Range

2-Chloroethyl Vinyl ether

Response Ratio



$$\text{Response} = 1.839\text{e-}001 * \text{Amt} - 5.758\text{e-}002$$

Coef of Det (r^2) = 0.999746 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y062325S.M

Calibration Table Last Updated: Tue Jun 24 08:29:52 2025