

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.428	0.354	17.3	76	0.00
3 P	Chloromethane	0.816	0.814	0.2	93	0.00
4 C	Vinyl Chloride	1.020	0.965	5.4#	84	0.00
5 T	Bromomethane	0.802	0.764	4.7	90	0.00
6 T	Chloroethane	0.686	0.670	2.3	87	0.00
7 T	Trichlorofluoromethane	1.129	1.053	6.7	82	0.00
8 T	Diethyl Ether	0.279	0.272	2.5	86	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.516	0.487	5.6	86	0.00
10 T	Methyl Iodide	0.562	0.544	3.2	79	0.00
11 T	Tert butyl alcohol	0.037	0.035	5.4	83	0.00
12 CM	1,1-Dichloroethene	0.506	0.476	5.9#	84	-0.01
13 T	Acrolein	0.050	0.017	66.0#	31#	0.00
14 T	Allyl chloride	0.778	0.778	0.0	88	-0.01
15 T	Acrylonitrile	0.116	0.122	-5.2	92	0.00
16 T	Acetone	0.105	0.122	-16.2	116	0.00
17 T	Carbon Disulfide	1.635	1.468	10.2	80	0.00
18 T	Methyl Acetate	0.349	0.354	-1.4	91	0.00
19 T	Methyl tert-butyl Ether	1.378	1.374	0.3	87	-0.01
20 T	Methylene Chloride	0.666	0.638	4.2	98	-0.01
21 T	trans-1,2-Dichloroethene	0.578	0.544	5.9	83	-0.01
22 T	Diisopropyl ether	1.729	1.780	-2.9	90	0.00
23 T	Vinyl Acetate	0.955	1.004	-5.1	90	-0.01
24 P	1,1-Dichloroethane	1.044	1.034	1.0	87	0.00
25 T	2-Butanone	0.154	0.165	-7.1	98	0.00
26 T	2,2-Dichloropropane	0.875	0.886	-1.3	91	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.641	4.6	85	0.00
28 T	Bromochloromethane	0.439	0.454	-3.4	90	0.00
29 T	Tetrahydrofuran	0.097	0.101	-4.1	90	0.00
30 C	Chloroform	1.076	1.046	2.8#	86	0.00
31 T	Cyclohexane	0.959	0.901	6.0	86	0.00
32 T	1,1,1-Trichloroethane	0.929	0.902	2.9	86	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.544	2.5	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.304	0.301	1.0	88	0.00
36 T	1,1-Dichloropropene	0.461	0.451	2.2	86	0.00
37 T	Ethyl Acetate	0.199	0.211	-6.0	91	-0.01
38 T	Carbon Tetrachloride	0.486	0.476	2.1	86	0.00
39 T	Methylcyclohexane	0.596	0.582	2.3	84	0.00
40 TM	Benzene	1.417	1.404	0.9	85	0.00
41 T	Methacrylonitrile	0.122	0.127	-4.1	94	-0.01
42 TM	1,2-Dichloroethane	0.388	0.393	-1.3	87	0.00
43 T	Isopropyl Acetate	0.413	0.428	-3.6	87	0.00
44 TM	Trichloroethene	0.356	0.347	2.5	83	0.00
45 C	1,2-Dichloropropane	0.332	0.337	-1.5#	88	0.00
46 T	Dibromomethane	0.188	0.184	2.1	86	0.00
47 T	Bromodichloromethane	0.485	0.482	0.6	86	0.00
48 T	Methyl methacrylate	0.199	0.211	-6.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
 Data File : VY022968.D
 Acq On : 08 Jul 2025 11:15
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	80	0.00
50 S	Toluene-d8	1.207	1.196	0.9	88	0.00
51 T	4-Methyl-2-Pentanone	0.210	0.230	-9.5	91	0.00
52 CM	Toluene	0.894	0.893	0.1#	86	0.00
53 T	t-1,3-Dichloropropene	0.437	0.444	-1.6	88	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.518	-2.4	88	0.00
55 T	1,1,2-Trichloroethane	0.244	0.245	-0.4	86	0.00
56 T	Ethyl methacrylate	0.328	0.346	-5.5	87	0.00
57 T	1,3-Dichloropropane	0.425	0.433	-1.9	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.173	-19.3	91	0.00
59 T	2-Hexanone	0.143	0.158	-10.5	93	0.00
60 T	Dibromochloromethane	0.315	0.313	0.6	85	0.00
61 T	1,2-Dibromoethane	0.228	0.231	-1.3	87	0.00
62 S	4-Bromofluorobenzene	0.388	0.385	0.8	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.472	0.449	4.9	77	0.00
65 PM	Chlorobenzene	1.099	1.090	0.8	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.362	2.7	83	0.00
67 C	Ethyl Benzene	1.930	1.941	-0.6#	85	0.00
68 T	m/p-Xylenes	0.746	0.752	-0.8	85	0.00
69 T	o-Xylene	0.703	0.706	-0.4	85	0.00
70 T	Styrene	1.178	1.195	-1.4	85	0.00
71 P	Bromoform	0.207	0.210	-1.4	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.698	3.689	0.2	85	0.00
74 T	N-amyl acetate	0.830	0.921	-11.0	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.596	0.621	-4.2	95	0.00
76 T	1,2,3-Trichloropropane	0.510	0.476	6.7	81	0.00
77 T	Bromobenzene	0.838	0.816	2.6	83	0.00
78 T	n-propylbenzene	4.465	4.541	-1.7	87	0.00
79 T	2-Chlorotoluene	2.522	2.510	0.5	85	0.00
80 T	1,3,5-Trimethylbenzene	2.985	3.011	-0.9	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.217	-7.4	95	0.00
82 T	4-Chlorotoluene	2.650	2.615	1.3	84	0.00
83 T	tert-Butylbenzene	2.633	2.723	-3.4	87	0.00
84 T	1,2,4-Trimethylbenzene	2.989	3.031	-1.4	85	0.00
85 T	sec-Butylbenzene	3.961	4.044	-2.1	86	0.00
86 T	p-Isopropyltoluene	3.296	3.367	-2.2	86	0.00
87 T	1,3-Dichlorobenzene	1.683	1.671	0.7	85	0.00
88 T	1,4-Dichlorobenzene	1.672	1.637	2.1	84	0.00
89 T	n-Butylbenzene	3.101	3.196	-3.1	87	0.00
90 T	Hexachloroethane	0.657	0.654	0.5	85	0.00
91 T	1,2-Dichlorobenzene	1.483	1.455	1.9	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.103	-2.0	87	0.00
93 T	1,2,4-Trichlorobenzene	0.838	0.807	3.7	83	0.00
94 T	Hexachlorobutadiene	0.474	0.449	5.3	84	0.00
95 T	Naphthalene	1.515	1.522	-0.5	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070825\
Data File : VY022968.D
Acq On : 08 Jul 2025 11:15
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 09 02:28:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.724	0.692	4.4	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6