

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063025\
 Data File : VW031730.D
 Acq On : 30 Jun 2025 10:15
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/01/2025
 Supervised By :Semsettin Yesilyurt 07/01/2025

Quant Time: Jul 01 02:18:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 02:17:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	216272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	387016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	345396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	170268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	31661	10.201	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery =	20.400%#		
35) Dibromofluoromethane	7.898	113	25109	9.942	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery =	19.880%#		
50) Toluene-d8	10.325	98	96344	10.251	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery =	20.500%#		
62) 4-Bromofluorobenzene	12.617	95	34948	10.105	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery =	20.200%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.046	85	16044	10.893	ug/l	94
3) Chloromethane	2.253	50	17744	9.993	ug/l	99
4) Vinyl Chloride	2.405	62	23112	9.902	ug/l	95
5) Bromomethane	2.820	94	18643	10.226	ug/l	94
6) Chloroethane	2.966	64	16074	10.139	ug/l	91
7) Trichlorofluoromethane	3.308	101	21990	10.374	ug/l	94
8) Diethyl Ether	3.722	74	17881	11.102	ug/l	89
9) 1,1,2-Trichlorotrifluo...	4.100	101	24340	10.346	ug/l	99
10) Methyl Iodide	4.313	142	36597	10.090	ug/l	99
11) Tert butyl alcohol	5.234	59	10031m	52.167	ug/l	
12) 1,1-Dichloroethene	4.076	96	27046	10.495	ug/l	95
13) Acrolein	3.929	56	18553	53.683	ug/l	100
14) Allyl chloride	4.704	41	39145	9.900	ug/l	100
15) Acrylonitrile	5.405	53	40289	50.867	ug/l	99
16) Acetone	4.161	43	41372	53.597	ug/l	97
17) Carbon Disulfide	4.417	76	68668	9.958	ug/l	98
18) Methyl Acetate	4.710	43	22628	10.321	ug/l	99
19) Methyl tert-butyl Ether	5.460	73	44040	10.069	ug/l	95
20) Methylene Chloride	4.960	84	42368	10.166	ug/l	96
21) trans-1,2-Dichloroethene	5.460	96	26955	9.844	ug/l	87
22) Diisopropyl ether	6.344	45	77462	10.045	ug/l	95
23) Vinyl Acetate	6.283	43	254825	48.380	ug/l	98
24) 1,1-Dichloroethane	6.240	63	50299	10.061	ug/l	99
25) 2-Butanone	7.197	43	48340	48.435	ug/l	97
26) 2,2-Dichloropropane	7.185	77	29549	10.130	ug/l	97
27) cis-1,2-Dichloroethene	7.191	96	30687	9.749	ug/l	99
28) Bromochloromethane	7.532	49	21655	9.815	ug/l	98
29) Tetrahydrofuran	7.557	42	34596	50.617	ug/l	98
30) Chloroform	7.691	83	53600	10.090	ug/l	99
31) Cyclohexane	7.971	56	46081	10.427	ug/l #	89
32) 1,1,1-Trichloroethane	7.886	97	42532	10.389	ug/l	98
36) 1,1-Dichloropropene	8.099	75	36621	10.021	ug/l	99
37) Ethyl Acetate	7.282	43	23126	10.032	ug/l	99
38) Carbon Tetrachloride	8.081	117	38531	10.347	ug/l	99
39) Methylcyclohexane	9.343	83	44170	9.715	ug/l	97
40) Benzene	8.337	78	107869	9.977	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063025\
 Data File : VW031730.D
 Acq On : 30 Jun 2025 10:15
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/01/2025
 Supervised By :Semsettin Yesilyurt 07/01/2025

Quant Time: Jul 01 02:18:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 02:17:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.502	41	11713	9.485	ug/l	87
42) 1,2-Dichloroethane	8.416	62	37957	10.390	ug/l	99
43) Isopropyl Acetate	8.441	43	39393	9.913	ug/l	99
44) Trichloroethene	9.099	130	26788	9.927	ug/l	98
45) 1,2-Dichloropropane	9.380	63	26605	10.181	ug/l	97
46) Dibromomethane	9.471	93	16818	9.969	ug/l	96
47) Bromodichloromethane	9.654	83	39595	10.009	ug/l	98
48) Methyl methacrylate	9.447	41	20612	10.549	ug/l	96
49) 1,4-Dioxane	9.465	88	4204m	234.072	ug/l	
51) 4-Methyl-2-Pentanone	10.215	43	116628	51.050	ug/l	99
52) Toluene	10.392	92	69506	10.041	ug/l	99
53) t-1,3-Dichloropropene	10.611	75	35578	9.585	ug/l	97
54) cis-1,3-Dichloropropene	10.081	75	40343	9.588	ug/l	94
55) 1,1,2-Trichloroethane	10.788	97	21816	9.872	ug/l #	88
56) Ethyl methacrylate	10.648	69	28317	9.445	ug/l	99
57) 1,3-Dichloropropane	10.934	76	40528	10.446	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.928	63	76896	46.578	ug/l	98
59) 2-Hexanone	10.971	43	77810	50.260	ug/l	97
60) Dibromochloromethane	11.129	129	25734	9.738	ug/l	96
61) 1,2-Dibromoethane	11.233	107	22082	9.981	ug/l	100
64) Tetrachloroethene	10.861	164	22144	9.801	ug/l	90
65) Chlorobenzene	11.654	112	77068	10.113	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.727	131	24224	9.975	ug/l	98
67) Ethyl Benzene	11.733	91	131184	9.927	ug/l	98
68) m/p-Xylenes	11.836	106	100863	19.856	ug/l	98
69) o-Xylene	12.166	106	45204	9.611	ug/l	97
70) Styrene	12.184	104	79722	9.805	ug/l	98
71) Bromoform	12.355	173	13967	9.641	ug/l #	99
73) Isopropylbenzene	12.464	105	120857	9.331	ug/l	100
74) N-amyl acetate	12.275	43	33049	9.467	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.714	83	28567	9.935	ug/l	99
76) 1,2,3-Trichloropropane	12.763	75	22177m	10.079	ug/l	
77) Bromobenzene	12.745	156	27969	9.682	ug/l	91
78) n-propylbenzene	12.800	91	150962	9.734	ug/l	98
79) 2-Chlorotoluene	12.885	91	90162	9.654	ug/l	99
80) 1,3,5-Trimethylbenzene	12.940	105	101724	9.489	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.513	75	9266	9.808	ug/l	92
82) 4-Chlorotoluene	12.989	91	95411	9.731	ug/l	99
83) tert-Butylbenzene	13.202	119	87194	9.495	ug/l	96
84) 1,2,4-Trimethylbenzene	13.245	105	104832	9.651	ug/l	100
85) sec-Butylbenzene	13.379	105	132016	9.578	ug/l	99
86) p-Isopropyltoluene	13.495	119	107805	9.490	ug/l	99
87) 1,3-Dichlorobenzene	13.495	146	58049	10.021	ug/l	99
88) 1,4-Dichlorobenzene	13.574	146	59597	10.051	ug/l	94
89) n-Butylbenzene	13.818	91	105923	9.597	ug/l	99
90) Hexachloroethane	14.086	117	19306	9.420	ug/l	93
91) 1,2-Dichlorobenzene	13.867	146	51680	9.761	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.482	75	5409	9.634	ug/l	97
93) 1,2,4-Trichlorobenzene	15.129	180	30599	9.371	ug/l	98
94) Hexachlorobutadiene	15.226	225	15982	10.136	ug/l	91
95) Naphthalene	15.360	128	71758	8.964	ug/l	98
96) 1,2,3-Trichlorobenzene	15.543	180	29473	9.823	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063025\
Data File : VW031730.D
Acq On : 30 Jun 2025 10:15
Operator : SY/MD
Sample : VSTDICC010
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/01/2025
Supervised By :Semsettin Yesilyurt 07/01/2025

Quant Time: Jul 01 02:18:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 02:17:14 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW063025\
Data File : VW031730.D
Acq On : 30 Jun 2025 10:15
Operator : SY/MD
Sample : VSTDIC010
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDIC010

Quant Time: Jul 01 02:18:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W063025S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 02:17:14 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 07/01/2025
Supervised By : Semsettin Yesilyurt 07/01/2025

