

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022780.D
 Acq On : 23 Jun 2025 15:08
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 24 02:53:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	475581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	797167	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	712427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	372652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	542659	102.235	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 204.460%#			
35) Dibromofluoromethane	7.628	113	500461	103.231	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 206.460%#			
50) Toluene-d8	10.103	98	2012122	104.594	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 209.180%#			
62) 4-Bromofluorobenzene	12.402	95	673658	108.931	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 217.860%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.861	85	383820	94.381	ug/l	97
3) Chloromethane	2.062	50	721237	92.877	ug/l	96
4) Vinyl Chloride	2.196	62	944075	97.310	ug/l	100
5) Bromomethane	2.580	94	723247	94.826	ug/l	100
6) Chloroethane	2.726	64	639835	98.113	ug/l	96
7) Trichlorofluoromethane	3.043	101	1072039	99.794	ug/l	98
8) Diethyl Ether	3.446	74	267488	100.816	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.805	101	467671	95.263	ug/l	99
10) Methyl Iodide	3.994	142	581464	108.785	ug/l	99
11) Tert butyl alcohol	4.854	59	184238	522.007	ug/l	99
12) 1,1-Dichloroethene	3.781	96	475211	98.719	ug/l	96
13) Acrolein	3.647	56	232931	486.701	ug/l	98
14) Allyl chloride	4.372	41	751070	101.438	ug/l	100
15) Acrylonitrile	5.049	53	578586	522.657	ug/l	99
16) Acetone	3.866	43	454852	453.380	ug/l	93
17) Carbon Disulfide	4.098	76	1545558	99.386	ug/l	99
18) Methyl Acetate	4.379	43	336182	101.185	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	1388952	105.967	ug/l	99
20) Methylene Chloride	4.604	84	549956	86.802	ug/l	98
21) trans-1,2-Dichloroethene	5.104	96	552497	100.424	ug/l	93
22) Diisopropyl ether	6.012	45	1716130	104.367	ug/l	98
23) Vinyl Acetate	5.951	43	4870879	536.357	ug/l	100
24) 1,1-Dichloroethane	5.909	63	1003848	101.070	ug/l	99
25) 2-Butanone	6.884	43	743716	509.343	ug/l	98
26) 2,2-Dichloropropane	6.878	77	821294	98.646	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	653863	102.279	ug/l	99
28) Bromochloromethane	7.238	49	421499	100.943	ug/l	98
29) Tetrahydrofuran	7.256	42	489018	530.386	ug/l	99
30) Chloroform	7.415	83	1031414	100.825	ug/l	99
31) Cyclohexane	7.695	56	860538	94.387	ug/l	98
32) 1,1,1-Trichloroethane	7.610	97	892886	101.003	ug/l	99
36) 1,1-Dichloropropene	7.829	75	745473	101.447	ug/l	100
37) Ethyl Acetate	6.982	43	335415	105.778	ug/l	98
38) Carbon Tetrachloride	7.811	117	784957	101.239	ug/l	99
39) Methylcyclohexane	9.103	83	969618	101.963	ug/l	97
40) Benzene	8.073	78	2338781	103.518	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022780.D
 Acq On : 23 Jun 2025 15:08
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:53:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 02:48:20 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	172503	86.232	ug/l	90
42) 1,2-Dichloroethane	8.152	62	644142	104.044	ug/l	100
43) Isopropyl Acetate	8.195	43	703942	106.813	ug/l	99
44) Trichloroethene	8.859	130	573927	101.178	ug/l	94
45) 1,2-Dichloropropane	9.140	63	543704	102.824	ug/l	100
46) Dibromomethane	9.225	93	308843	102.889	ug/l	99
47) Bromodichloromethane	9.420	83	803009	103.743	ug/l	99
48) Methyl methacrylate	9.219	41	348723	110.068	ug/l	99
49) 1,4-Dioxane	9.225	88	74032	2087.539	ug/l	95
51) 4-Methyl-2-Pentanone	9.993	43	1833298	547.404	ug/l	100
52) Toluene	10.170	92	1522485	106.816	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	754526	108.342	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	860746	106.593	ug/l	97
55) 1,1,2-Trichloroethane	10.566	97	407343	104.800	ug/l	99
56) Ethyl methacrylate	10.432	69	592646	113.331	ug/l	99
57) 1,3-Dichloropropane	10.713	76	705449	104.027	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	1439247	506.517	ug/l	99
59) 2-Hexanone	10.755	43	1250090	549.018	ug/l	99
60) Dibromochloromethane	10.908	129	535327	106.627	ug/l	99
61) 1,2-Dibromoethane	11.012	107	388631	106.848	ug/l	99
64) Tetrachloroethene	10.646	164	674360	100.198	ug/l	98
65) Chlorobenzene	11.438	112	1610361	102.870	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.511	131	559842	105.587	ug/l	99
67) Ethyl Benzene	11.511	91	2906187	105.662	ug/l	99
68) m/p-Xylenes	11.627	106	2278768	214.327	ug/l	99
69) o-Xylene	11.950	106	1082170	108.020	ug/l	99
70) Styrene	11.963	104	1865505	111.168	ug/l	99
71) Bromoform	12.127	173	321083	108.756	ug/l #	98
73) Isopropylbenzene	12.249	105	2764621	100.312	ug/l	99
74) N-amyl acetate	12.066	43	667613	107.920	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.499	83	442467	99.615	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	364588m	92.907	ug/l	
77) Bromobenzene	12.530	156	631186	101.040	ug/l	99
78) n-propylbenzene	12.590	91	3318659	99.735	ug/l	99
79) 2-Chlorotoluene	12.676	91	1904783	101.319	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	2286529	102.773	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	157671	104.703	ug/l	94
82) 4-Chlorotoluene	12.773	91	2025702	102.582	ug/l	99
83) tert-Butylbenzene	12.993	119	1990057	101.425	ug/l	100
84) 1,2,4-Trimethylbenzene	13.036	105	2310931	103.747	ug/l	99
85) sec-Butylbenzene	13.170	105	2993511	101.400	ug/l	99
86) p-Isopropyltoluene	13.285	119	2565197	104.416	ug/l	100
87) 1,3-Dichlorobenzene	13.279	146	1304649	103.989	ug/l	100
88) 1,4-Dichlorobenzene	13.359	146	1259287	101.046	ug/l	100
89) n-Butylbenzene	13.609	91	2355349	101.924	ug/l	99
90) Hexachloroethane	13.877	117	487480	99.627	ug/l	98
91) 1,2-Dichlorobenzene	13.651	146	1129471	102.157	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	75666	100.453	ug/l	99
93) 1,2,4-Trichlorobenzene	14.913	180	648932	103.955	ug/l	99
94) Hexachlorobutadiene	15.017	225	334737	94.827	ug/l	99
95) Naphthalene	15.139	128	1263808	111.955	ug/l	99
96) 1,2,3-Trichlorobenzene	15.322	180	559578	103.737	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022780.D
Acq On : 23 Jun 2025 15:08
Operator : SY/MD
Sample : VSTDICC100
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 02:53:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022780.D
Acq On : 23 Jun 2025 15:08
Operator : SY/MD
Sample : VSTDIC100
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC100

Quant Time: Jun 24 02:53:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 02:48:20 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

