

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	480076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	806385	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	709921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	359311	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	268490	50.109	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 100.220%			
35) Dibromofluoromethane	7.628	113	246693	50.304	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 100.600%			
50) Toluene-d8	10.103	98	968221	49.755	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 99.500%			
62) 4-Bromofluorobenzene	12.401	95	314924	50.342	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 100.680%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.861	85	188576	45.936	ug/l	99
3) Chloromethane	2.068	50	383362	48.905	ug/l	97
4) Vinyl Chloride	2.196	62	480207	49.034	ug/l	100
5) Bromomethane	2.586	94	372449	48.375	ug/l	98
6) Chloroethane	2.726	64	317341	48.206	ug/l	98
7) Trichlorofluoromethane	3.043	101	519156	47.875	ug/l	99
8) Diethyl Ether	3.452	74	128819	48.097	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.805	101	234148	47.248	ug/l	98
10) Methyl Iodide	3.994	142	252766	46.847	ug/l	100
11) Tert butyl alcohol	4.860	59	90461	253.906	ug/l #	86
12) 1,1-Dichloroethene	3.781	96	232605	47.868	ug/l	94
13) Acrolein	3.647	56	99713	206.396	ug/l	100
14) Allyl chloride	4.378	41	357454	47.825	ug/l	99
15) Acrylonitrile	5.049	53	278053	248.823	ug/l	99
16) Acetone	3.860	43	217427	214.694	ug/l	95
17) Carbon Disulfide	4.098	76	753545	48.003	ug/l	99
18) Methyl Acetate	4.385	43	169460	50.527	ug/l	99
19) Methyl tert-butyl Ether	5.110	73	652685	49.329	ug/l	99
20) Methylene Chloride	4.610	84	279076	43.635	ug/l	97
21) trans-1,2-Dichloroethene	5.110	96	266943	48.066	ug/l	98
22) Diisopropyl ether	6.012	45	800669	48.237	ug/l	93
23) Vinyl Acetate	5.951	43	2264377	247.008	ug/l	99
24) 1,1-Dichloroethane	5.909	63	486998	48.573	ug/l	99
25) 2-Butanone	6.890	43	353339	239.723	ug/l	98
26) 2,2-Dichloropropane	6.878	77	388040	46.171	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	308904	47.867	ug/l	100
28) Bromochloromethane	7.244	49	211120	50.087	ug/l	100
29) Tetrahydrofuran	7.256	42	235186	252.693	ug/l	100
30) Chloroform	7.414	83	500741	48.491	ug/l	97
31) Cyclohexane	7.695	56	422671	45.926	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	431799	48.388	ug/l	99
36) 1,1-Dichloropropene	7.829	75	354829	47.735	ug/l	100
37) Ethyl Acetate	6.982	43	160039	49.894	ug/l	99
38) Carbon Tetrachloride	7.817	117	373583	47.632	ug/l	98
39) Methylcyclohexane	9.103	83	463905	48.226	ug/l	98
40) Benzene	8.073	78	1105593	48.376	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	106081m	53.720	ug/l	
42) 1,2-Dichloroethane	8.152	62	307548	49.108	ug/l	100
43) Isopropyl Acetate	8.195	43	333219	49.983	ug/l	99
44) Trichloroethene	8.859	130	274874	47.904	ug/l	98
45) 1,2-Dichloropropane	9.140	63	259472	48.510	ug/l	100
46) Dibromomethane	9.231	93	145961	48.070	ug/l	98
47) Bromodichloromethane	9.420	83	382665	48.872	ug/l	99
48) Methyl methacrylate	9.219	41	164362	51.285	ug/l	99
49) 1,4-Dioxane	9.225	88	35276	983.335	ug/l	96
51) 4-Methyl-2-Pentanone	9.993	43	867628	256.103	ug/l	99
52) Toluene	10.170	92	704960	48.894	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	348211	49.428	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	404721	49.547	ug/l	99
55) 1,1,2-Trichloroethane	10.566	97	194151	49.380	ug/l	98
56) Ethyl methacrylate	10.432	69	270067	51.054	ug/l	99
57) 1,3-Dichloropropane	10.713	76	340025	49.568	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	680973	245.249	ug/l	100
59) 2-Hexanone	10.755	43	580425	251.999	ug/l	99
60) Dibromochloromethane	10.908	129	253700	49.955	ug/l	100
61) 1,2-Dibromoethane	11.011	107	181754	49.399	ug/l	99
64) Tetrachloroethene	10.646	164	303522	45.257	ug/l	98
65) Chlorobenzene	11.438	112	750243	48.095	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.511	131	253468	47.973	ug/l	99
67) Ethyl Benzene	11.517	91	1351655	49.316	ug/l	100
68) m/p-Xylenes	11.621	106	1043651	98.506	ug/l	100
69) o-Xylene	11.950	106	495462	49.631	ug/l	99
70) Styrene	11.962	104	834381	49.897	ug/l	99
71) Bromoform	12.127	173	142330	48.380	ug/l	99
73) Isopropylbenzene	12.249	105	1264391	47.581	ug/l	100
74) N-amyl acetate	12.066	43	299778	50.258	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.499	83	213316	49.808	ug/l	99
76) 1,2,3-Trichloropropane	12.548	75	171584m	46.775	ug/l	
77) Bromobenzene	12.529	156	287890	47.796	ug/l	100
78) n-propylbenzene	12.590	91	1543476	48.108	ug/l	100
79) 2-Chlorotoluene	12.676	91	871527	48.080	ug/l	99
80) 1,3,5-Trimethylbenzene	12.731	105	1052077	49.044	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	66143	45.554	ug/l	99
82) 4-Chlorotoluene	12.773	91	919177	48.276	ug/l	99
83) tert-Butylbenzene	12.993	119	917192	48.481	ug/l	99
84) 1,2,4-Trimethylbenzene	13.035	105	1058356	49.278	ug/l	99
85) sec-Butylbenzene	13.170	105	1377847	48.405	ug/l	100
86) p-Isopropyltoluene	13.285	119	1157677	48.873	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	584110	48.286	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	574913	47.844	ug/l	100
89) n-Butylbenzene	13.615	91	1083397	48.623	ug/l	100
90) Hexachloroethane	13.877	117	228507	48.434	ug/l	100
91) 1,2-Dichlorobenzene	13.657	146	512081	48.036	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	35693	49.145	ug/l	99
93) 1,2,4-Trichlorobenzene	14.913	180	293980	48.842	ug/l	99
94) Hexachlorobutadiene	15.017	225	157173	46.178	ug/l	98
95) Naphthalene	15.139	128	558879	51.347	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	252267	48.503	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
Data File : VY022783.D
Acq On : 23 Jun 2025 16:17
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY062325

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 03:08:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062325\
 Data File : VY022783.D
 Acq On : 23 Jun 2025 16:17
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY062325

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 06/24/2025
 Supervised By :Semsettin Yesilyurt 06/24/2025

Quant Time: Jun 24 03:36:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 03:08:29 2025
 Response via : Initial Calibration

