

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW063025

Review By	Mahesh Dadoda	Review On	7/1/2025 12:21:51 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/1/2025 12:24:32 PM		
SubDirectory	VW063025	HP Acquire Method	MSVOA_W	HP Processing Method	82w063025s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134578
Initial Calibration Stds	VP134581,VP134582,VP134583,VP134584,VP134585,VP134586
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031728.D	30 Jun 2025 08:56		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VW031729.D	30 Jun 2025 09:54		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VW031730.D	30 Jun 2025 10:15	8260-soil-method	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VW031731.D	30 Jun 2025 10:58		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VW031732.D	30 Jun 2025 11:21	Comp. #20 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VW031733.D	30 Jun 2025 12:34		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VW031734.D	30 Jun 2025 12:55		SY/MD	Ok,M
8	VIBLK	VIBLK	VW031735.D	30 Jun 2025 14:11		SY/MD	Ok
9	VSTDICV050	ICVVW063025	VW031736.D	30 Jun 2025 15:23		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW070925

Review By	Mahesh Dadoda	Review On	7/14/2025 7:37:29 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/14/2025 7:50:58 AM
SubDirectory	VW070925	HP Acquire Method	MSVOA_W
		HP Processing Method	82w063025s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134687
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134688,VP134689 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031766.D	09 Jul 2025 08:51		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW031767.D	09 Jul 2025 10:00		SY/MD	Ok,M
3	VW0709SBL01	VW0709SBL01	VW031768.D	09 Jul 2025 10:58		SY/MD	Ok
4	VW0709SBS01	VW0709SBS01	VW031769.D	09 Jul 2025 11:34	BS Failed High For Com.#20	SY/MD	Ok,M
5	VW0709SBSD01	VW0709SBSD01	VW031770.D	09 Jul 2025 11:56	BSD Failed High For Comp.#20	SY/MD	Ok,M
6	VW0709SBS02	VW0709SBS02	VW031771.D	09 Jul 2025 13:29	Not Required	SY/MD	Not Ok
7	Q2519-03	TP-15 VOC	VW031772.D	09 Jul 2025 13:51	vial-A	SY/MD	Ok
8	Q2527-01	AR520-0015	VW031773.D	09 Jul 2025 14:13	vial-A	SY/MD	Ok
9	Q2526-04	LAW-25-0102	VW031774.D	09 Jul 2025 14:34	vial-A	SY/MD	Ok
10	Q2514-06	TP-103	VW031775.D	09 Jul 2025 14:56	vial-B	SY/MD	Ok
11	Q2480-06	GPX6	VW031776.D	09 Jul 2025 15:18	Bad purge	SY/MD	Not Ok
12	Q2480-07	GPX7	VW031777.D	09 Jul 2025 15:40	vial-A Surrogate fail;Need MeOH	SY/MD	Dilution
13	Q2529-01	TP-91	VW031778.D	09 Jul 2025 16:02	vial-A	SY/MD	Ok
14	Q2529-02	TP-80	VW031779.D	09 Jul 2025 16:24	vial-A	SY/MD	Ok
15	Q2529-03	TP-79	VW031780.D	09 Jul 2025 16:46	vial-A	SY/MD	Ok
16	Q2529-04	TP-95	VW031781.D	09 Jul 2025 17:08	vial-A	SY/MD	Ok
17	Q2529-05	TP-98	VW031782.D	09 Jul 2025 17:30	vial-A BS,BSD Failed High	SY/MD	ReRun
18	Q2529-06	TP-102	VW031783.D	09 Jul 2025 17:52	vial-A BS,BSD Failed High	SY/MD	ReRun

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW070925

Review By	Maresh Dadoda	Review On	7/14/2025 7:37:29 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/14/2025 7:50:58 AM
SubDirectory	VW070925	HP Acquire Method	MSVOA_W HP Processing Method 82w063025s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134687		
CCC	VP134688,VP134689		
Internal Standard/PEM	VP133934		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2529-07	TP-101	VW031784.D	09 Jul 2025 18:14	vial-A BS,BSD Failed High	SY/MD	ReRun
20	Q2529-08	TP-89	VW031785.D	09 Jul 2025 18:36	vial-A BS,BSD Failed High	SY/MD	ReRun
21	Q2529-09	TP-33	VW031786.D	09 Jul 2025 18:57	vial-A BS,BSD Failed High	SY/MD	ReRun
22	Q2529-10	TP-30	VW031787.D	09 Jul 2025 19:19	vial-A BS,BSD Failed High	SY/MD	ReRun
23	VIBLK	VIBLK	VW031788.D	09 Jul 2025 20:25		SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VW031789.D	09 Jul 2025 20:46		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW071025

Review By	Mahesh Dadoda	Review On	7/14/2025 7:38:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/14/2025 7:52:00 AM
SubDirectory	VW071025	HP Acquire Method	MSVOA_W HP Processing Method 82w063025s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134704
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134705,VP134706 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031790.D	10 Jul 2025 08:16		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW031791.D	10 Jul 2025 08:47		SY/MD	Ok,M
3	VW0710SBL01	VW0710SBL01	VW031792.D	10 Jul 2025 10:15		SY/MD	Ok
4	VW0710SBS01	VW0710SBS01	VW031793.D	10 Jul 2025 11:05		SY/MD	Ok,M
5	VW0710SBSD01	VW0710SBSD01	VW031794.D	10 Jul 2025 11:33		SY/MD	Ok,M
6	Q2480-03RE	GPX3RE	VW031795.D	10 Jul 2025 12:19	vial-B Internal standard fail;Surrogate fail	SY/MD	Confirms
7	Q2480-04	GPX4	VW031796.D	10 Jul 2025 12:41	vial-B Need MeOH	SY/MD	Dilution
8	Q2480-06	GPX6	VW031797.D	10 Jul 2025 13:03	vial-B Surrogate fail;Need MeOH	SY/MD	Dilution
9	Q2529-05	TP-98	VW031798.D	10 Jul 2025 13:25	vial-B	SY/MD	Ok
10	Q2529-06	TP-102	VW031799.D	10 Jul 2025 13:47	vial-B	SY/MD	Ok
11	Q2529-07	TP-101	VW031800.D	10 Jul 2025 14:09	vial-B	SY/MD	Ok
12	Q2529-08	TP-89	VW031801.D	10 Jul 2025 14:30	vial-B	SY/MD	Ok
13	Q2529-09	TP-33	VW031802.D	10 Jul 2025 14:52	vial-B	SY/MD	Ok,M
14	Q2529-10	TP-30	VW031803.D	10 Jul 2025 15:14	vial-B	SY/MD	Ok
15	Q2555-01	OU4-TS-29-070925	VW031804.D	10 Jul 2025 15:36	vial-A	SY/MD	Ok
16	Q2555-03	OU4-TS-30-070925	VW031805.D	10 Jul 2025 15:58	vial-A	SY/MD	Ok
17	Q2539-01	HACKENSACK	VW031806.D	10 Jul 2025 16:20	vial-A	SY/MD	Ok

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW071025

Review By	Mahesh Dadoda	Review On	7/14/2025 7:38:05 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/14/2025 7:52:00 AM		
SubDirectory	VW071025	HP Acquire Method	MSVOA_W	HP Processing Method	82w063025s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134704			
CCC		VP134705,VP134706			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2550-02	VNJ-203	VW031807.D	10 Jul 2025 16:42	vial-A	SY/MD	Ok
19	Q2551-02	VNJ-204	VW031808.D	10 Jul 2025 17:04	vial-A	SY/MD	Ok
20	Q2543-01	TP-41	VW031809.D	10 Jul 2025 17:26	vial-A	SY/MD	Ok
21	Q2543-02	TP-49	VW031810.D	10 Jul 2025 17:48	vial-A	SY/MD	Ok
22	Q2543-03	TP-23	VW031811.D	10 Jul 2025 18:10	vial-A	SY/MD	Ok
23	Q2543-04	TP-23-99	VW031812.D	10 Jul 2025 18:32	vial-A	SY/MD	Ok
24	Q2558-01	OU4-TS-Denali-070925	VW031813.D	10 Jul 2025 18:53	vial-A	SY/MD	Ok
25	Q2558-03	OU4-TS-Grillo-OG-0709	VW031814.D	10 Jul 2025 19:15	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VW031815.D	10 Jul 2025 19:59		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX070225

Review By	John Carlone	Review On	7/3/2025 8:44:30 AM
Supervise By	Maresh Dadoda	Supervise On	7/3/2025 3:09:59 PM
SubDirectory	VX070225	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134642 VP134633,VP134634,VP134635,VP134638,VP134639,VP134640 VP134641

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046859.D	02 Jul 2025 11:12		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX046860.D	02 Jul 2025 12:11		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX046861.D	02 Jul 2025 12:37	%D Failed for com.#13 in 5 ppb	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX046862.D	02 Jul 2025 13:18		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX046863.D	02 Jul 2025 13:39	LR-13	JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX046864.D	02 Jul 2025 14:10		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX046865.D	02 Jul 2025 14:31		JC/MD	Ok,M
8	VIBLK	VIBLK	VX046866.D	02 Jul 2025 14:53		JC/MD	Ok
9	VSTDICV050	ICVVX070225	VX046867.D	02 Jul 2025 15:14		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071025

Review By	John Carlone	Review On	7/11/2025 8:53:41 AM
Supervise By	Maresh Dadoda	Supervise On	7/14/2025 7:47:20 AM
SubDirectory	VX071025	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134710
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134714,VP134715

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046932.D	10 Jul 2025 08:43		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046933.D	10 Jul 2025 09:54		JC/MD	Ok,M
3	VX0710MBL01	VX0710MBL01	VX046934.D	10 Jul 2025 10:22		JC/MD	Ok
4	VX0710WBL01	VX0710WBL01	VX046935.D	10 Jul 2025 11:44		JC/MD	Ok
5	VX0710WBS01	VX0710WBS01	VX046936.D	10 Jul 2025 12:10		JC/MD	Ok,M
6	VX0710WBSD01	VX0710WBSD01	VX046937.D	10 Jul 2025 12:35		JC/MD	Ok,M
7	IBLK	IBLK	VX046938.D	10 Jul 2025 12:57		JC/MD	Ok
8	Q2553-06	TB	VX046939.D	10 Jul 2025 13:18	vial A pH<2 TB	JC/MD	Ok
9	Q2553-05	FB	VX046940.D	10 Jul 2025 13:39	vial A pH<2 FB;Hit of com.#17	JC/MD	ReRun
10	Q2554-02	250709059-10 Trip blan	VX046941.D	10 Jul 2025 14:00	vial A pH#6.0 TB	JC/MD	Ok
11	Q2554-04	250709059-11 Trip blan	VX046942.D	10 Jul 2025 14:21	vial A pH#6.0 TB	JC/MD	Ok
12	Q2553-01	AOC-201	VX046943.D	10 Jul 2025 14:43	vial A pH<2	JC/MD	Ok,M
13	Q2553-02	AOC-202	VX046944.D	10 Jul 2025 15:04	vial A pH<2	JC/MD	Ok,M
14	Q2553-03	AOC-203	VX046945.D	10 Jul 2025 15:26	vial A pH<2	JC/MD	Ok,M
15	Q2553-04	AOC-205	VX046946.D	10 Jul 2025 15:47	vial A pH<2	JC/MD	Ok,M
16	PB168762TB	PB168762TB	VX046947.D	10 Jul 2025 16:08		JC/MD	Ok,M
17	Q2517-04	TP-14	VX046948.D	10 Jul 2025 16:30	vial A pH#5.0	JC/MD	Ok
18	Q2519-04	TP-15	VX046949.D	10 Jul 2025 16:51	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071025

Review By	John Carlone	Review On	7/11/2025 8:53:41 AM
Supervise By	Maresh Dadoda	Supervise On	7/14/2025 7:47:20 AM
SubDirectory	VX071025	HP Acquire Method	HP Processing Method 82X070225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134710		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134714,VP134715		

19	Q2554-03	250709104-01 VOA	VX046950.D	10 Jul 2025 17:12	vial A pH#6.0	JC/MD	Ok
20	Q2554-01	250709071-01 VOA	VX046951.D	10 Jul 2025 17:34	vial A pH#6.0 Need 5X. No more vials.	JC/MD	Ok,M
21	IBLK	IBLK	VX046952.D	10 Jul 2025 17:55		JC/MD	Ok
22	VX0710MBS01	VX0710MBS01	VX046953.D	10 Jul 2025 18:16	BS Failed Low for com.#05	JC/MD	Ok,M
23	VX0710MBSD01	VX0710MBSD01	VX046954.D	10 Jul 2025 18:37	BSD Failed Low for com.#05	JC/MD	Ok,M
24	Q2480-05	GPX5	VX046955.D	10 Jul 2025 18:59		JC/MD	Ok,M
25	Q2480-06ME	GPX6ME	VX046956.D	10 Jul 2025 19:20		JC/MD	Ok,M
26	Q2480-07ME	GPX7ME	VX046957.D	10 Jul 2025 19:41		JC/MD	Ok,M
27	Q2480-08	GPX8	VX046958.D	10 Jul 2025 20:02		JC/MD	Ok,M
28	VSTDCCC050	VSTDCCC050EC	VX046959.D	10 Jul 2025 20:23		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071425

Review By	Maresh Dadoda	Review On	7/15/2025 4:54:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/15/2025 4:55:39 AM
SubDirectory	VX071425	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134751
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134752,VP134753

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046960.D	14 Jul 2025 08:10		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046961.D	14 Jul 2025 09:16		JC/MD	Ok,M
3	VX0714MBL01	VX0714MBL01	VX046962.D	14 Jul 2025 10:54		JC/MD	Ok
4	VX0714WBL01	VX0714WBL01	VX046963.D	14 Jul 2025 11:18		JC/MD	Ok
5	VX0714MBS01	VX0714MBS01	VX046964.D	14 Jul 2025 11:39		JC/MD	Ok,M
6	VX0714MBSD01	VX0714MBSD01	VX046965.D	14 Jul 2025 12:06		JC/MD	Ok,M
7	Q2480-04ME	GPX4ME	VX046966.D	14 Jul 2025 12:27		JC/MD	Ok
8	Q2480-02ME	GPX2ME	VX046967.D	14 Jul 2025 12:49		JC/MD	Ok,M
9	VX0714WBS01	VX0714WBS01	VX046968.D	14 Jul 2025 13:10		JC/MD	Ok,M
10	VX0714WBSD01	VX0714WBSD01	VX046969.D	14 Jul 2025 13:37		JC/MD	Ok,M
11	Q2523-01	MOO-25-0190	VX046970.D	14 Jul 2025 13:58	Need Soil Run	JC/MD	Not Ok
12	Q2526-02	LAW-25-0101	VX046971.D	14 Jul 2025 14:19	Need Soil Run	JC/MD	Not Ok
13	Q2565-03	MOO-25-0191	VX046972.D	14 Jul 2025 14:41	Need Soil Run	JC/MD	Not Ok
14	Q2565-04	MOO-25-0196	VX046973.D	14 Jul 2025 15:02	Need Soil Run	JC/MD	Not Ok
15	Q2565-05	MOO-25-0180	VX046974.D	14 Jul 2025 15:23	Need Soil Run	JC/MD	Not Ok
16	IBLK	IBLK	VX046975.D	14 Jul 2025 15:45		JC/MD	Ok
17	Q2578-01	WC-A5-01-G	VX046976.D	14 Jul 2025 16:06	Surrogate fail	JC/MD	ReRun
18	Q2578-05	WC-A2-09-G	VX046977.D	14 Jul 2025 16:28	Surrogate fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071425

Review By	Maresh Dadoda	Review On	7/15/2025 4:54:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/15/2025 4:55:39 AM
SubDirectory	VX071425	HP Acquire Method	HP Processing Method 82X070225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134751		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134752,VP134753		

19	Q2578-09	WC-A2-10-G	VX046978.D	14 Jul 2025 16:49	Surrogate fail	JC/MD	ReRun
20	Q2578-13	WC-A2-11-G	VX046979.D	14 Jul 2025 17:10	Surrogate fail	JC/MD	ReRun
21	Q2578-17	WC-A2-12-G	VX046980.D	14 Jul 2025 17:31	Surrogate fail	JC/MD	ReRun
22	Q2579-01	WC-A2-13-G	VX046981.D	14 Jul 2025 17:53	Surrogate fail	JC/MD	ReRun
23	Q2579-05	WC-A2-14-G	VX046982.D	14 Jul 2025 18:14	Surrogate fail	JC/MD	ReRun
24	IBLK	IBLK	VX046983.D	14 Jul 2025 18:35		JC/MD	Ok
25	Q2553-05	FB	VX046984.D	14 Jul 2025 18:56	FB	JC/MD	Ok
26	Q2595-01	CHRT-24849	VX046985.D	14 Jul 2025 19:17	Need Soil Run	JC/MD	Not Ok
27	VSTDCCC050	VSTDCCC050EC	VX046986.D	14 Jul 2025 19:38		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:26 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM		
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134461
Initial Calibration Stds	VP134462,VP134463,VP134464,VP134465,VP134466,VP134467
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134468
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134656
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134657,VP134658 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022945.D	07 Jul 2025 08:36		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022946.D	07 Jul 2025 09:08		SY/MD	Ok,M
3	VY0707SBL01	VY0707SBL01	VY022947.D	07 Jul 2025 09:44		SY/MD	Ok
4	VY0707SBS01	VY0707SBS01	VY022948.D	07 Jul 2025 10:17	BS Failed High for com.#20	SY/MD	Ok,M
5	VY0707SBSD01	VY0707SBSD01	VY022949.D	07 Jul 2025 10:40	BSD Failed High for com.#20	SY/MD	Ok,M
6	Q2507-03	SU-04-07032025-VOC	VY022950.D	07 Jul 2025 11:18	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q2513-01	HR-2-070325	VY022951.D	07 Jul 2025 11:41	vial-A	SY/MD	Ok
8	Q2513-03	HR-3-070325	VY022952.D	07 Jul 2025 12:04	vial-A	SY/MD	Ok
9	Q2504-02	VOC	VY022953.D	07 Jul 2025 12:28	vial-A Internal Standard Fail;BS,BSD Failed High	SY/MD	ReRun
10	Q2515-01	WC-1	VY022954.D	07 Jul 2025 12:51	vial-A	SY/MD	Ok,M
11	Q2487-07RE	G1(4.5)RE	VY022955.D	07 Jul 2025 13:15	vial-B Internal standard fail	SY/MD	Confirms
12	Q2487-08	G1(10)	VY022956.D	07 Jul 2025 13:38	vial-A	SY/MD	Ok
13	Q2510-02	#63025-A-VOC	VY022957.D	07 Jul 2025 14:02	vial-A	SY/MD	Ok,M
14	Q2484-01	TP-58	VY022958.D	07 Jul 2025 14:25	vial-A	SY/MD	Ok
15	Q2484-02	TP-57	VY022959.D	07 Jul 2025 14:48	vial-A	SY/MD	Ok
16	Q2484-03	TP-64	VY022960.D	07 Jul 2025 15:12	vial-A	SY/MD	Ok
17	Q2484-04	TP-107	VY022961.D	07 Jul 2025 15:35	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM		
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP134656				
CCC	VP134657,VP134658				
Internal Standard/PEM	VP133934				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2484-05	TP-106	VY022962.D	07 Jul 2025 15:59	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
19	Q2484-06	TP-104	VY022963.D	07 Jul 2025 16:22	vial-A	SY/MD	Ok
20	Q2514-01	TP-92	VY022964.D	07 Jul 2025 16:45	vial-A	SY/MD	Ok
21	Q2514-02	TP-93	VY022965.D	07 Jul 2025 17:09	vial-A	SY/MD	Ok
22	Q2480-01	GPX1	VY022966.D	07 Jul 2025 17:32	vial-A	SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070825

Review By	Maresh Dadoda	Review On	7/9/2025 4:21:01 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134673
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134675,VP134676 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022967.D	08 Jul 2025 08:30		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022968.D	08 Jul 2025 11:15		SY/MD	Ok,M
3	VY0708SBL01	VY0708SBL01	VY022969.D	08 Jul 2025 11:44		SY/MD	Ok
4	VY0708SBS01	VY0708SBS01	VY022970.D	08 Jul 2025 12:12	BS Failed High for com.#20	SY/MD	Ok,M
5	VY0708SBSD01	VY0708SBSD01	VY022971.D	08 Jul 2025 12:35	BSD Failed High for com.#20	SY/MD	Ok,M
6	Q2507-03	SU-04-07032025-VOC	VY022972.D	08 Jul 2025 13:43	vial-B	SY/MD	Ok
7	Q2504-02RE	VOCRE	VY022973.D	08 Jul 2025 14:16	vial-B Internal Standard Fail;BS,BSD Failed High	SY/MD	Confirms
8	Q2484-05	TP-106	VY022974.D	08 Jul 2025 14:40	vial-B	SY/MD	Ok
9	Q2514-03	TP-94	VY022975.D	08 Jul 2025 15:03	vial-A	SY/MD	Ok
10	Q2514-04	TP-96	VY022976.D	08 Jul 2025 15:26	vial-A	SY/MD	Ok
11	Q2514-05	TP-97	VY022977.D	08 Jul 2025 15:50	vial-A	SY/MD	Ok
12	Q2514-06	TP-103	VY022978.D	08 Jul 2025 16:13	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
13	Q2514-07	TP-36	VY022979.D	08 Jul 2025 16:37	vial-A	SY/MD	Ok
14	Q2514-08	TP-78	VY022980.D	08 Jul 2025 17:00	vial-A	SY/MD	Ok
15	Q2514-09	TP-81	VY022981.D	08 Jul 2025 17:23	vial-A	SY/MD	Ok
16	Q2514-10	TP-90	VY022982.D	08 Jul 2025 17:47	vial-A	SY/MD	Ok
17	Q2480-02	GPX2	VY022983.D	08 Jul 2025 18:10	vial-A Need MeOH;Surrogate Fail;BS,BSD Failed High	SY/MD	Dilution

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070825

Review By	Mahesh Dadoda	Review On	7/9/2025 4:21:01 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/9/2025 4:22:56 PM		
SubDirectory	VY070825	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP134673				
CCC	VP134675,VP134676				
Internal Standard/PEM	VP133934				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2480-03	GPX3	VY022984.D	08 Jul 2025 18:34	vial-A BS,BSD Failed High	SY/MD	ReRun
19	Q2480-04	GPX4	VY022985.D	08 Jul 2025 18:57	vial-A Internal Standard Fail;BS,BSD Failed High	SY/MD	ReRun
20	Q2517-03	TP-14-VOC	VY022986.D	08 Jul 2025 19:21	vial-A	SY/MD	Ok
21	VIBLK	VIBLK	VY022987.D	08 Jul 2025 19:44		SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY022988.D	08 Jul 2025 20:07		SY/MD	Ok,M

M : Manual Integration