

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2480</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VW031793.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0710SBS01	Dichlorodifluoromethane	20	18.5	ug/Kg	93			64	136	
	Chloromethane	20	19.3	ug/Kg	97			52	151	
	Vinyl chloride	20	20.1	ug/Kg	101			56	148	
	Bromomethane	20	18.8	ug/Kg	94			58	141	
	Chloroethane	20	19.6	ug/Kg	98			69	130	
	Trichlorofluoromethane	20	18.3	ug/Kg	92			69	134	
	1,1,2-Trichlorotrifluoroethane	20	19.3	ug/Kg	97			81	123	
	1,1-Dichloroethene	20	20.2	ug/Kg	101			79	121	
	Acetone	100	100	ug/Kg	100			40	171	
	Carbon disulfide	20	18.7	ug/Kg	94			59	130	
	Methyl tert-butyl Ether	20	21.2	ug/Kg	106			77	129	
	Methyl Acetate	20	17.7	ug/Kg	89			69	149	
	Methylene Chloride	20	24.3	ug/Kg	121			72	131	
	trans-1,2-Dichloroethene	20	20.1	ug/Kg	101			80	123	
	1,1-Dichloroethane	20	20.3	ug/Kg	102			82	123	
	Cyclohexane	20	20.2	ug/Kg	101			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	19.6	ug/Kg	98			76	129	
	cis-1,2-Dichloroethene	20	20.8	ug/Kg	104			82	123	
	Bromochloromethane	20	20.2	ug/Kg	101			80	127	
	Chloroform	20	21.2	ug/Kg	106			82	125	
	1,1,1-Trichloroethane	20	19.8	ug/Kg	99			80	126	
	Methylcyclohexane	20	20.4	ug/Kg	102			77	123	
	Benzene	20	20.8	ug/Kg	104			84	121	
	1,2-Dichloroethane	20	20.4	ug/Kg	102			81	126	
	Trichloroethene	20	20.4	ug/Kg	102			83	122	
	1,2-Dichloropropane	20	21.1	ug/Kg	106			83	122	
	Bromodichloromethane	20	20.3	ug/Kg	102			82	123	
	4-Methyl-2-Pentanone	100	100	ug/Kg	100			70	135	
	Toluene	20	20.9	ug/Kg	104			83	122	
	t-1,3-Dichloropropene	20	20.6	ug/Kg	103			78	124	
	cis-1,3-Dichloropropene	20	20.9	ug/Kg	104			81	122	
	1,1,2-Trichloroethane	20	20.6	ug/Kg	103			82	125	
	2-Hexanone	100	110	ug/Kg	110			66	138	
	Dibromochloromethane	20	19.4	ug/Kg	97			79	125	
	1,2-Dibromoethane	20	20.5	ug/Kg	103			80	125	
	Tetrachloroethene	20	19.5	ug/Kg	98			83	125	
	Chlorobenzene	20	21.4	ug/Kg	107			84	122	
	Ethyl Benzene	20	22.0	ug/Kg	110			82	124	
	m/p-Xylenes	40	42.3	ug/Kg	106			83	124	
	o-Xylene	20	21.8	ug/Kg	109			83	123	
	Styrene	20	21.4	ug/Kg	107			82	124	
	Bromoform	20	18.0	ug/Kg	90			75	127	
	Isopropylbenzene	20	20.4	ug/Kg	102			82	124	
	1,1,2,2-Tetrachloroethane	20	19.8	ug/Kg	99			77	127	
	1,3-Dichlorobenzene	20	19.9	ug/Kg	100			83	122	
	1,4-Dichlorobenzene	20	19.1	ug/Kg	96			84	121	
	1,2-Dichlorobenzene	20	20.0	ug/Kg	100			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2480</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VW031793.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VW0710SBS01	1,2-Dibromo-3-Chloropropane	20	18.2	ug/Kg	91			66	134	
	1,2,4-Trichlorobenzene	20	19.4	ug/Kg	97			78	127	
	1,2,3-Trichlorobenzene	20	18.5	ug/Kg	93			70	137	