

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2480
 Lab File ID: VW031767.D Date Analyzed: 07/09/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	247441	7.96	422710	8.85	377094	11.63
UPPER LIMIT	494882	8.459	845420	9.349	754188	12.129
LOWER LIMIT	123721	7.459	211355	8.349	188547	11.129
EPA SAMPLE NO.						
GPX7	202570	7.97	393419	8.86	350819	11.63
VW0709SBL01	208773	7.97	393637	8.85	349817	11.63
VW0709SBS01	187242	7.95	349121	8.85	314319	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.