

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q2480  
 Lab File ID: VX046933.D Date Analyzed: 07/10/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:54  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 309661        | 5.56  | 506196        | 6.76  | 449744        | 10.05  |
| UPPER LIMIT    | 619322        | 6.056 | 1012390       | 7.263 | 899488        | 10.549 |
| LOWER LIMIT    | 154831        | 5.056 | 253098        | 6.263 | 224872        | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| GPX5           | 264865        | 5.56  | 461340        | 6.77  | 407988        | 10.06  |
| GPX6ME         | 352552        | 5.56  | 617302        | 6.76  | 563552        | 10.06  |
| GPX7ME         | 249557        | 5.56  | 445780        | 6.77  | 406077        | 10.06  |
| GPX8           | 377228        | 5.56  | 658079        | 6.77  | 584775        | 10.06  |
| VX0710MBL01    | 220827        | 5.56  | 390309        | 6.76  | 362012        | 10.06  |
| VX0710MBS01    | 334978        | 5.56  | 561587        | 6.77  | 507345        | 10.06  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.