

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2480
 Lab File ID: VX046961.D Date Analyzed: 07/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	302863	5.56	503750	6.76	451308	10.05
UPPER LIMIT	605726	6.056	1007500	7.263	902616	10.549
LOWER LIMIT	151432	5.056	251875	6.263	225654	9.549
EPA SAMPLE NO.						
GPX2ME	302766	5.56	512075	6.76	458895	10.06
GPX4ME	315290	5.56	536279	6.76	490599	10.06
VX0714MBL01	353449	5.56	590149	6.77	542379	10.06
VX0714MBS01	309393	5.56	505368	6.77	459018	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.