

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2480</u>
Lab File ID:	<u>VY022946.D</u>	Date Analyzed:	<u>07/07/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>09:08</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	469181	7.71	754142	8.62	648632	11.41
UPPER LIMIT	938362	8.207	1508280	9.116	1297260	11.914
LOWER LIMIT	234591	7.207	377071	8.116	324316	10.914
EPA SAMPLE NO.						
GPX1	283156	7.71	548345	8.62	553452	11.41
VY0707SBL01	359794	7.71	659995	8.62	632698	11.41
VY0707SBS01	438661	7.71	730129	8.61	612594	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.