

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2480
 Lab File ID: VY022968.D Date Analyzed: 07/08/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	422213	7.71	699147	8.62	604420	11.41
UPPER LIMIT	844426	8.207	1398290	9.116	1208840	11.914
LOWER LIMIT	211107	7.207	349574	8.116	302210	10.914
EPA SAMPLE NO.						
GPX2	323213	7.71	600118	8.62	719431	11.41
GPX3	292899	7.71	553522	8.62	566452	11.41
VY0708SBL01	314896	7.71	593436	8.62	582188	11.41
VY0708SBS01	371442	7.71	616012	8.62	524718	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.