

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q2480  
 Lab File ID: VX046933.D Date Analyzed: 07/10/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:54  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 231018        | 12.018 |  |  |  |  |
| UPPER LIMIT    | 462036        | 12.518 |  |  |  |  |
| LOWER LIMIT    | 115509        | 11.518 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| GPX5           | 192639        | 12.02  |  |  |  |  |
| GPX6ME         | 282479        | 12.02  |  |  |  |  |
| GPX7ME         | 201295        | 12.02  |  |  |  |  |
| GPX8           | 281510        | 12.02  |  |  |  |  |
| VX0710MBL01    | 178141        | 12.02  |  |  |  |  |
| VX0710MBS01    | 262771        | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.