



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Environmental Restoration, LLC

Project Name: CC2-16 Analytical

Project # N/A

Order ID # Q2481

Test Name: Flash Point,pH

A. Number of Samples and Date of Receipt:

10 Water samples were received on 06/27/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Flash Point, PCB, pH, TCLP Extraction, TCLP ICP Metals and TCLP Mercury. This data package contains results for Flash Point,pH.

C. Analytical Techniques:

The analysis of Flash Point was based on method 1010B and The analysis of pH was based on method 9040C.

D. QA/ QC Samples:

The Holding Times were met for all samples except for CC0267-OXPL of pH, for CC0625-NL of pH.for CC0625-OXBL of pH.for CC0627-AL of pH.for CC0627-AOXL of pH.for CC0627-BL of pH.for CC0627-CLOXAL of pH.for CC0627-CLOXPL of pH.for CC0627-OXL of pH.for CC0627-SFBL of pH as samples were receive out of holding time.

The Duplicate analysis met criteria for all parameters.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

The temperature of the samples at the time of receipt was 21.0°C.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____