

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB136367

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/03/2025	09:10
2	CAL2	1	Water	NA	NA	20.2	7.00	07/03/2025	09:11
3	CAL3	1	Water	NA	NA	20.3	10.02	07/03/2025	09:15
4	ICV	1	Water	NA	NA	20.3	7.02	07/03/2025	09:20
5	CCV1	1	Water	NA	NA	20.3	2.01	07/03/2025	09:25
6	Q2481-01	1	Water	NA	NA	20.3	1.5	07/03/2025	09:35
7	Q2481-01DUP	1	Water	NA	NA	20.3	1.36	07/03/2025	09:36
8	Q2481-02	1	Water	NA	NA	20.5	5.02	07/03/2025	09:40
9	Q2481-03	1	Water	NA	NA	20.6	14.12	07/03/2025	10:00
10	Q2481-04	1	Water	NA	NA	20.1	1.5	07/03/2025	10:10
11	Q2481-05	1	Water	NA	NA	20.6	10.02	07/03/2025	10:25
12	Q2481-06	1	Water	NA	NA	20.1	6.02	07/03/2025	10:35
13	Q2481-07	1	Water	NA	NA	20.3	6.02	07/03/2025	10:40
14	Q2481-08	1	Water	NA	NA	20.1	5.03	07/03/2025	10:45
15	Q2481-09	1	Water	NA	NA	20.1	14.03	07/03/2025	10:50
16	CCV2	1	Water	NA	NA	20.2	12.02	07/03/2025	11:00
17	Q2481-10	1	Water	NA	NA	20.3	14.05	07/03/2025	11:15
18	CCV3	1	Water	NA	NA	20.2	2.01	07/03/2025	11:20

WORKLIST(Hardcopy Internal Chain)

UP 136357

WorkList Name : ph w q2481

WorkList ID : 190535

Department : Wet-Chemistry

Date : 07-03-2025 08:49:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2481-01	CC0627-AL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-02	CC0627-CLOXP	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-03	CC0625-OXBL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-04	CC0627-AOXL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-05	CC0625-NL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-06	CC0267-OXP	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-07	CC0627-OXL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-08	CC0627-CLOXAL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-09	CC0627-BL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C
Q2481-10	CC0627-SFBL	Water	pH	Cool 4 deg C	ENV160	A13	06/27/2025	9040C

Date/Time 07/03/25 09:00
 Raw Sample Received by: JF Wey
 Raw Sample Relinquished by: JF Wey

Date/Time 07/03/25 12:30
 Raw Sample Received by: JF Wey
 Raw Sample Relinquished by: JF Wey