

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH QC BATCHID: LB136391

	NA	NO	YES
1.Calibration Summary met criteria.	_____	_____	_____✓
2.ICP Interference Check Sample Results Summary Submitted.	_____	_____	_____✓
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	_____✓
4.Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	_____✓
5.Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓	_____
6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
8.Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	_____✓
9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓

**ADDITIONAL COMMENTS:**

Sample for Q2481-02,03,04,06,07,09 are analyzed straight as x10 dilution due to vey highly contaminated matrix and physical samples are dark and dark brown after digestion, not able to inject as straight to avoid damage to the instrument, can clog the tubes and carryover issue.

Analyst MOHAN

Date 7/9/2025 12:00:00 AM

QA REVIEW

Date