



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Environmental Restoration, LLC

SDG No.: Q2481

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.25	0.2	125	70 - 130	CV	07/08/2025	09:14	LB136391
CRI01	Arsenic	20.6	20.0	103	65 - 135	P	07/10/2025	19:01	LB136434
	Barium	96.3	100	96	65 - 135	P	07/10/2025	19:01	LB136434
	Cadmium	5.33	6.0	89	65 - 135	P	07/10/2025	19:01	LB136434
	Chromium	10.5	10.0	105	65 - 135	P	07/10/2025	19:01	LB136434
	Lead	12.4	12.0	103	65 - 135	P	07/10/2025	19:01	LB136434
	Selenium	16.7	20.0	84	65 - 135	P	07/10/2025	19:01	LB136434
	Silver	10.5	10.0	105	65 - 135	P	07/10/2025	19:01	LB136434