



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2481

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB19	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	09:04	LB136391

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2481

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	09:12	LB136391
CCB62	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	09:50	LB136391
CCB63	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	10:26	LB136391
CCB64	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	10:56	LB136391
CCB65	Mercury	0.076	+/-0.2	U	0.20	CV	07/08/2025	11:17	LB136391

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2481

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	18:55	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	18:55	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	18:55	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	18:55	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	18:55	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	18:55	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	18:55	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2481

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	19:54	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	19:54	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	19:54	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	19:54	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	19:54	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	19:54	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	19:54	LB136434
CCB02	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	21:04	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	21:04	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	21:04	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	21:04	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	21:04	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	21:04	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	21:04	LB136434
CCB03	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	22:36	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	22:36	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	22:36	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	22:36	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	22:36	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	22:36	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	22:36	LB136434
CCB04	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	23:53	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	23:53	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	23:53	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	23:53	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	23:53	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	23:53	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	23:53	LB136434
CCB05	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	00:51	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	00:51	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	00:51	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	00:51	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	00:51	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	00:51	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	00:51	LB136434
CCB06	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	02:23	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	02:23	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	02:23	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	02:23	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2481

Contract: ENVI60

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	07/11/2025	02:23	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	02:23	LB136434
	Silver	4.87	+/-5	J	10.0	P	07/11/2025	02:23	LB136434
CCB07	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	03:26	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	03:26	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	03:26	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	03:26	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	03:26	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	03:26	LB136434
	Silver	4.57	+/-5	J	10.0	P	07/11/2025	03:26	LB136434
CCB08	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	04:27	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	04:27	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	04:27	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	04:27	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	04:27	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	04:27	LB136434
	Silver	6.06	+/-5	J*	10.0	P	07/11/2025	04:27	LB136434
CCB09	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	05:24	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	05:24	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	05:24	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	05:24	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	05:24	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	05:24	LB136434
	Silver	2.97	+/-5	J	10.0	P	07/11/2025	05:24	LB136434
CCB10	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	06:21	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	06:21	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	06:21	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	06:21	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	06:21	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	06:21	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	06:21	LB136434
CCB11	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	06:47	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	06:47	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	06:47	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	06:47	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	06:47	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	06:47	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	06:47	LB136434