

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Environmental Restoration, LLC</u>	SDG No.: <u>Q2481</u>
Contract: <u>ENVI60</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P5</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	2.10			-20	20	07/10/2025	19:11	LB136434
	Barium	5.81	6.0	97	-94	106	07/10/2025	19:11	LB136434
	Cadmium	-0.49	1.0	49	-5	7	07/10/2025	19:11	LB136434
	Chromium	48.9	52.0	94	42	62	07/10/2025	19:11	LB136434
	Lead	0.44			-12	12	07/10/2025	19:11	LB136434
	Selenium	4.66			-20	20	07/10/2025	19:11	LB136434
	Silver	-1.13			-10	10	07/10/2025	19:11	LB136434
ICSAB01	Arsenic	92.1	100	92	88.4	120	07/10/2025	19:15	LB136434
	Barium	505	540	94	437	637	07/10/2025	19:15	LB136434
	Cadmium	972	970	100	826	1120	07/10/2025	19:15	LB136434
	Chromium	548	540	102	460	624	07/10/2025	19:15	LB136434
	Lead	43.4	49.0	89	37	61	07/10/2025	19:15	LB136434
	Selenium	52.3	46.0	114	26	66	07/10/2025	19:15	LB136434
	Silver	217	200	108	170	232	07/10/2025	19:15	LB136434