

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2481

Contract: ENVI60

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168712BS							
Arsenic	ug/L	4000	3540		88	80 - 120	P
Barium	ug/L	1000	1000		100	80 - 120	P
Cadmium	ug/L	1000	934		93	80 - 120	P
Chromium	ug/L	2000	2010		100	80 - 120	P
Lead	ug/L	5000	4680		94	80 - 120	P
Selenium	ug/L	10000	9760		98	80 - 120	P
Silver	ug/L	380	352		93	80 - 120	P



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals
- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client:

Environmental Restoration, LLC

SDG No.:

Q2481

Contract:

ENVI60

Lab Code:

ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168742BS Mercury	ug/L	4.0	3.81		95	80 - 120	CV