

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2481

Continuing Calib Date: 07/02/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 10:56

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.76       | 4.77      | 4.67      | 4.87 | 0.01       |
| Aroclor-1016-2       | (2) | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1016-3       | (3) | 4.84       | 4.84      | 4.74      | 4.94 | 0.00       |
| Aroclor-1016-4       | (4) | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1016-5       | (5) | 5.22       | 5.22      | 5.12      | 5.32 | 0.00       |
| Aroclor-1260-1       | (1) | 6.25       | 6.26      | 6.16      | 6.36 | 0.01       |
| Aroclor-1260-2       | (2) | 6.44       | 6.45      | 6.35      | 6.55 | 0.01       |
| Aroclor-1260-3       | (3) | 6.81       | 6.81      | 6.71      | 6.91 | 0.00       |
| Aroclor-1260-4       | (4) | 7.07       | 7.07      | 6.97      | 7.17 | 0.00       |
| Aroclor-1260-5       | (5) | 7.31       | 7.32      | 7.22      | 7.42 | 0.01       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58      | 3.78 | 0.01       |
| Decachlorobiphenyl   |     | 8.71       | 8.71      | 8.61      | 8.81 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2481

Continuing Calib Date: 07/02/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 10:56

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.75       | 4.75      | 4.65      | 4.85 | 0.00       |
| Aroclor-1016-2       | (2) | 4.77       | 4.77      | 4.67      | 4.87 | 0.01       |
| Aroclor-1016-3       | (3) | 4.94       | 4.94      | 4.84      | 5.04 | 0.00       |
| Aroclor-1016-4       | (4) | 4.98       | 4.99      | 4.89      | 5.09 | 0.01       |
| Aroclor-1016-5       | (5) | 5.19       | 5.20      | 5.10      | 5.30 | 0.01       |
| Aroclor-1260-1       | (1) | 6.23       | 6.23      | 6.13      | 6.33 | 0.01       |
| Aroclor-1260-2       | (2) | 6.41       | 6.42      | 6.32      | 6.52 | 0.01       |
| Aroclor-1260-3       | (3) | 6.56       | 6.57      | 6.47      | 6.67 | 0.01       |
| Aroclor-1260-4       | (4) | 7.03       | 7.04      | 6.94      | 7.14 | 0.01       |
| Aroclor-1260-5       | (5) | 7.28       | 7.28      | 7.18      | 7.38 | 0.00       |
| Tetrachloro-m-xylene |     | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   |     | 8.66       | 8.66      | 8.56      | 8.76 | 0.00       |