

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2481

Continuing Calib Date: 07/03/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 15:14

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1248-2 (2)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-3 (3)	5.22	5.22	5.12	5.32	0.00
Aroclor-1248-4 (4)	5.57	5.57	5.47	5.67	0.00
Aroclor-1248-5 (5)	5.61	5.61	5.51	5.71	0.00
Tetrachloro-m-xylene	3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl	8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2481

Continuing Calib Date: 07/03/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 15:14

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1248-2 (2)	4.98	4.99	4.89	5.09	0.01
Aroclor-1248-3 (3)	5.02	5.03	4.93	5.13	0.01
Aroclor-1248-4 (4)	5.20	5.20	5.10	5.30	0.00
Aroclor-1248-5 (5)	5.59	5.59	5.49	5.69	0.00
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl	8.66	8.66	8.56	8.76	0.00