

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112016.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	488.970	500.000	-2.2
Aroclor-1016-2	4.781	4.684	4.884	505.170	500.000	1.0
Aroclor-1016-3	4.838	4.741	4.941	495.500	500.000	-0.9
Aroclor-1016-4	4.957	4.861	5.061	492.240	500.000	-1.6
Aroclor-1016-5	5.215	5.118	5.318	510.920	500.000	2.2
Aroclor-1260-1	6.252	6.157	6.357	477.850	500.000	-4.4
Aroclor-1260-2	6.442	6.346	6.546	554.830	500.000	11.0
Aroclor-1260-3	6.808	6.713	6.913	529.250	500.000	5.9
Aroclor-1260-4	7.068	6.972	7.172	502.020	500.000	0.4
Aroclor-1260-5	7.311	7.216	7.416	507.690	500.000	1.5
Decachlorobiphenyl	8.706	8.612	8.812	43.030	50.000	-13.9
Tetrachloro-m-xylene	3.674	3.577	3.777	55.600	50.000	11.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112016.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.746	4.651	4.851	510.860	500.000	2.2
Aroclor-1016-2	4.765	4.669	4.869	526.510	500.000	5.3
Aroclor-1016-3	4.940	4.844	5.044	504.240	500.000	0.8
Aroclor-1016-4	4.982	4.887	5.087	502.650	500.000	0.5
Aroclor-1016-5	5.194	5.099	5.299	495.030	500.000	-1.0
Aroclor-1260-1	6.224	6.129	6.329	505.430	500.000	1.1
Aroclor-1260-2	6.412	6.316	6.516	557.890	500.000	11.6
Aroclor-1260-3	6.564	6.469	6.669	491.940	500.000	-1.6
Aroclor-1260-4	7.035	6.939	7.139	471.020	500.000	-5.8
Aroclor-1260-5	7.276	7.181	7.381	494.560	500.000	-1.1
Decachlorobiphenyl	8.655	8.561	8.761	46.510	50.000	-7.0
Tetrachloro-m-xylene	3.670	3.573	3.773	52.030	50.000	4.1