

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112022.D **Time Analyzed:** 08:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.761	4.666	4.866	477.850	500.000	-4.4
Aroclor-1016-2	4.780	4.684	4.884	499.450	500.000	-0.1
Aroclor-1016-3	4.837	4.741	4.941	470.620	500.000	-5.9
Aroclor-1016-4	4.957	4.861	5.061	480.320	500.000	-3.9
Aroclor-1016-5	5.213	5.118	5.318	488.070	500.000	-2.4
Aroclor-1260-1	6.251	6.157	6.357	472.510	500.000	-5.5
Aroclor-1260-2	6.441	6.346	6.546	510.920	500.000	2.2
Aroclor-1260-3	6.808	6.713	6.913	520.960	500.000	4.2
Aroclor-1260-4	7.067	6.972	7.172	494.100	500.000	-1.2
Aroclor-1260-5	7.309	7.216	7.416	498.970	500.000	-0.2
Decachlorobiphenyl	8.705	8.612	8.812	42.460	50.000	-15.1
Tetrachloro-m-xylene	3.674	3.577	3.777	56.790	50.000	13.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112022.D **Time Analyzed:** 08:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.651	4.851	498.230	500.000	-0.4
Aroclor-1016-2	4.762	4.669	4.869	511.900	500.000	2.4
Aroclor-1016-3	4.937	4.844	5.044	494.400	500.000	-1.1
Aroclor-1016-4	4.980	4.887	5.087	495.680	500.000	-0.9
Aroclor-1016-5	5.191	5.099	5.299	479.500	500.000	-4.1
Aroclor-1260-1	6.220	6.129	6.329	488.810	500.000	-2.2
Aroclor-1260-2	6.409	6.316	6.516	512.810	500.000	2.6
Aroclor-1260-3	6.561	6.469	6.669	486.240	500.000	-2.8
Aroclor-1260-4	7.031	6.939	7.139	466.890	500.000	-6.6
Aroclor-1260-5	7.273	7.181	7.381	483.760	500.000	-3.2
Decachlorobiphenyl	8.651	8.561	8.761	44.900	50.000	-10.2
Tetrachloro-m-xylene	3.668	3.573	3.773	51.020	50.000	2.0