

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112037.D **Time Analyzed:** 14:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	467.580	500.000	-6.5
Aroclor-1016-2	4.781	4.684	4.884	485.270	500.000	-2.9
Aroclor-1016-3	4.838	4.741	4.941	462.880	500.000	-7.4
Aroclor-1016-4	4.957	4.861	5.061	468.950	500.000	-6.2
Aroclor-1016-5	5.214	5.118	5.318	471.980	500.000	-5.6
Aroclor-1260-1	6.252	6.157	6.357	475.360	500.000	-4.9
Aroclor-1260-2	6.441	6.346	6.546	522.660	500.000	4.5
Aroclor-1260-3	6.808	6.713	6.913	536.400	500.000	7.3
Aroclor-1260-4	7.067	6.972	7.172	507.900	500.000	1.6
Aroclor-1260-5	7.311	7.216	7.416	517.010	500.000	3.4
Decachlorobiphenyl	8.705	8.612	8.812	43.440	50.000	-13.1
Tetrachloro-m-xylene	3.675	3.577	3.777	52.370	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112037.D **Time Analyzed:** 14:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.745	4.651	4.851	508.790	500.000	1.8
Aroclor-1016-2	4.763	4.669	4.869	513.600	500.000	2.7
Aroclor-1016-3	4.938	4.844	5.044	505.810	500.000	1.2
Aroclor-1016-4	4.981	4.887	5.087	509.740	500.000	1.9
Aroclor-1016-5	5.193	5.099	5.299	491.140	500.000	-1.8
Aroclor-1260-1	6.222	6.129	6.329	503.560	500.000	0.7
Aroclor-1260-2	6.410	6.316	6.516	523.380	500.000	4.7
Aroclor-1260-3	6.562	6.469	6.669	491.770	500.000	-1.6
Aroclor-1260-4	7.031	6.939	7.139	477.980	500.000	-4.4
Aroclor-1260-5	7.273	7.181	7.381	485.650	500.000	-2.9
Decachlorobiphenyl	8.652	8.561	8.761	45.890	50.000	-8.2
Tetrachloro-m-xylene	3.669	3.573	3.773	50.580	50.000	1.2