

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111585.D	11 Jun 2025 10:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO111586.D	11 Jun 2025 10:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO111587.D	11 Jun 2025 10:40		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO111588.D	11 Jun 2025 10:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO111589.D	11 Jun 2025 11:17		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO111590.D	11 Jun 2025 11:35		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO111591.D	11 Jun 2025 11:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO111592.D	11 Jun 2025 12:12		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO111593.D	11 Jun 2025 12:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO111594.D	11 Jun 2025 12:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO111595.D	11 Jun 2025 13:07		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO111596.D	11 Jun 2025 13:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO111597.D	11 Jun 2025 13:44		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO111598.D	11 Jun 2025 14:02		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO111599.D	11 Jun 2025 14:20		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO111600.D	11 Jun 2025 14:39		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO111601.D	11 Jun 2025 15:14		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO111602.D	11 Jun 2025 15:32		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM	
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM	
SubDirectory	PO061125	HP Acquire Method	HP Processing Method	PO061125
STD. NAME	STD REF.#			
Tune/Reschk	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24332,PP24347,PP24352,PP24357 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387			
Initial Calibration Stds				
CCC				
Internal Standard/PEM				
ICV/I.BLK				
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

19	AR1248ICC050	AR1248ICC050	PO111603.D	11 Jun 2025 15:49		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO111604.D	11 Jun 2025 16:06		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO111605.D	11 Jun 2025 16:25		YP/AJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PO111606.D	11 Jun 2025 16:43		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO111607.D	11 Jun 2025 17:00		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO111608.D	11 Jun 2025 17:18		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO111609.D	11 Jun 2025 17:36		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO111610.D	11 Jun 2025 17:55		YP/AJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PO111611.D	11 Jun 2025 18:13		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO111612.D	11 Jun 2025 18:31		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO111613.D	11 Jun 2025 18:50		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO111614.D	11 Jun 2025 19:07		YP/AJ	Ok
31	PO061125ICV500	ICVPO061125	PO111615.D	11 Jun 2025 19:25		YP/AJ	Ok
32	AR1242ICV500	ICVPO061125AR1242	PO111616.D	11 Jun 2025 20:02		YP/AJ	Ok
33	AR1248ICV500	ICVPO061125AR1248	PO111617.D	11 Jun 2025 20:39		YP/AJ	Ok
34	AR1254ICV500	ICVPO061125AR1254	PO111618.D	11 Jun 2025 21:16		YP/AJ	Ok
35	AR1268ICV500	ICVPO061125AR1268	PO111619.D	11 Jun 2025 21:52		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070225

Review By	yogesh	Review On	7/2/2025 11:58:42 AM
Supervise By	mohammad	Supervise On	7/4/2025 4:31:44 AM
SubDirectory	PO070225	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111966.D	02 Jul 2025 10:38		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO111967.D	02 Jul 2025 10:56		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO111968.D	02 Jul 2025 11:15		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO111969.D	02 Jul 2025 11:32		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO111970.D	02 Jul 2025 11:50		YP/AJ	Ok
6	I.BLK	I.BLK	PO111971.D	02 Jul 2025 12:08		YP/AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PO111972.D	02 Jul 2025 12:26		YP/AJ	Ok
8	PB168671BL	PB168671BL	PO111973.D	02 Jul 2025 12:45	Not used	YP/AJ	Not Ok
9	PB168691BL	PB168691BL	PO111974.D	02 Jul 2025 13:02		YP/AJ	Ok
10	PB168691BS	PB168691BS	PO111975.D	02 Jul 2025 13:19		YP/AJ	Ok
11	Q2458-09	TP-59	PO111976.D	02 Jul 2025 13:37		YP/AJ	Ok
12	Q2467-01	1A-1B-1C-Fire Stopher	PO111977.D	02 Jul 2025 13:54		YP/AJ	Ok
13	Q2462-04	60425-AB	PO111978.D	02 Jul 2025 14:12		YP/AJ	Ok
14	Q2475-01	SOIL-PILE	PO111979.D	02 Jul 2025 15:05		YP/AJ	Ok,M
15	Q2478-01	WC-1	PO111980.D	02 Jul 2025 15:23		YP/AJ	Ok
16	AR1660CCC500	AR1660CCC500	PO111981.D	02 Jul 2025 16:19		YP/AJ	Ok
17	AR1242CCC500	AR1242CCC500	PO111982.D	02 Jul 2025 16:36		YP/AJ	Ok,M
18	AR1248CCC500	AR1248CCC500	PO111983.D	02 Jul 2025 16:53		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070225

Review By	yogesh	Review On	7/2/2025 11:58:42 AM
Supervise By	mohammad	Supervise On	7/4/2025 4:31:44 AM
SubDirectory	PO070225	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1254CCC500	AR1254CCC500	PO111984.D	02 Jul 2025 17:11		YP/AJ	Ok
20	I.BLK	I.BLK	PO111985.D	02 Jul 2025 17:28		YP/AJ	Ok
21	PB168704BL	PB168704BL	PO111986.D	02 Jul 2025 17:45		YP/AJ	Ok
22	PB168704BS	PB168704BS	PO111987.D	02 Jul 2025 18:03		YP/AJ	Ok
23	PB168704BSD	PB168704BSD	PO111988.D	02 Jul 2025 18:20		YP/AJ	Ok
24	Q2462-01	40425	PO111989.D	02 Jul 2025 18:39		YP/AJ	Ok
25	Q2463-01	TW-WTS-11	PO111990.D	02 Jul 2025 18:56	Need straight	YP/AJ	Not Ok
26	Q2465-04	62025-ABC	PO111991.D	02 Jul 2025 19:14	TCMX high in 1st column	YP/AJ	Ok,M
27	Q2481-02	CC0627-CLOXPL	PO111992.D	02 Jul 2025 19:31	Need to run with lower dilution 10X	YP/AJ	Not Ok
28	Q2481-03	CC0625-OXBL	PO111993.D	02 Jul 2025 19:48	Need to run with lower dilution 10X	YP/AJ	Not Ok
29	Q2481-05	CC0625-NL	PO111994.D	02 Jul 2025 20:07	Need to run with lower dilution 10X	YP/AJ	Not Ok
30	AR1660CCC500	AR1660CCC500	PO111995.D	02 Jul 2025 21:16		YP/AJ	Ok
31	AR1242CCC500	AR1242CCC500	PO111996.D	02 Jul 2025 22:11		YP/AJ	Ok
32	AR1248CCC500	AR1248CCC500	PO111997.D	02 Jul 2025 22:28		YP/AJ	Ok
33	AR1254CCC500	AR1254CCC500	PO111998.D	02 Jul 2025 22:46		YP/AJ	Ok
34	I.BLK	I.BLK	PO111999.D	02 Jul 2025 23:05		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070325

Review By	yogesh	Review On	7/3/2025 12:01:52 PM
Supervise By	mohammad	Supervise On	7/8/2025 9:14:04 AM
SubDirectory	PO070325	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112000.D	03 Jul 2025 08:12		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112001.D	03 Jul 2025 08:31		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO112002.D	03 Jul 2025 08:49		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO112003.D	03 Jul 2025 09:08		YP/AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PO112004.D	03 Jul 2025 09:26		YP/AJ	Ok
6	I.BLK	I.BLK	PO112005.D	03 Jul 2025 09:44		YP/AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PO112006.D	03 Jul 2025 10:02		YP/AJ	Ok
8	Q2481-08	CC0627-CLOXAL	PO112007.D	03 Jul 2025 10:44		YP/AJ	Ok,M
9	Q2481-10	CC0627-SFBL	PO112008.D	03 Jul 2025 11:03	need clean up	YP/AJ	Not Ok
10	Q2463-01	TW-WTS-11	PO112009.D	03 Jul 2025 11:21		YP/AJ	Ok
11	Q2481-02	CC0627-CLOXPL	PO112010.D	03 Jul 2025 11:39		YP/AJ	Ok,M
12	Q2481-03	CC0625-OXBL	PO112011.D	03 Jul 2025 11:56		YP/AJ	Ok,M
13	Q2481-05	CC0625-NL	PO112012.D	03 Jul 2025 12:14		YP/AJ	Ok,M
14	Q2481-06	CC0267-OXPL	PO112013.D	03 Jul 2025 12:32		YP/AJ	Ok
15	Q2481-07	CC0627-OXL	PO112014.D	03 Jul 2025 12:49	TCMX high in 2nd column	YP/AJ	Ok,M
16	Q2481-09	CC0627-BL	PO112015.D	03 Jul 2025 13:07		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO112016.D	03 Jul 2025 14:19		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO112017.D	03 Jul 2025 14:55		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070325

Review By	yogesh	Review On	7/3/2025 12:01:52 PM
Supervise By	mohammad	Supervise On	7/8/2025 9:14:04 AM
SubDirectory	PO070325	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO112018.D	03 Jul 2025 15:14		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO112019.D	03 Jul 2025 15:32		YP/AJ	Ok
21	I.BLK	I.BLK	PO112020.D	03 Jul 2025 15:51		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:03 AM
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112021.D	07 Jul 2025 08:40		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112022.D	07 Jul 2025 08:58		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO112023.D	07 Jul 2025 09:17		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO112024.D	07 Jul 2025 09:34		YP/AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PO112025.D	07 Jul 2025 09:53		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO112026.D	07 Jul 2025 10:10		YP/AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PO112027.D	07 Jul 2025 10:27		YP/AJ	Ok
8	PB168704BL	PB168704BL	PO112028.D	07 Jul 2025 10:46	Not used	YP/AJ	Not Ok
9	PB168710BL	PB168710BL	PO112029.D	07 Jul 2025 11:04		YP/AJ	Ok,M
10	PB168710BS	PB168710BS	PO112030.D	07 Jul 2025 11:22		YP/AJ	Ok
11	Q2466-01	61825	PO112031.D	07 Jul 2025 11:40		YP/AJ	Ok
12	Q2466-02	62625	PO112032.D	07 Jul 2025 11:59	AR1016 Hit	YP/AJ	Ok,M
13	Q2492-01	VNJ-254-1	PO112033.D	07 Jul 2025 12:17	AR1248 Hit	YP/AJ	Ok,M
14	Q2492-02	VNJ-254-2	PO112034.D	07 Jul 2025 12:34	AR1248 Hit	YP/AJ	Ok,M
15	Q2476-01	50731	PO112035.D	07 Jul 2025 12:53	AR1242 Hit , need 5x dilution	YP/AJ	Not Ok
16	Q2481-10	CC0627-SFBL	PO112036.D	07 Jul 2025 13:10		YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO112037.D	07 Jul 2025 14:47		YP/AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PO112038.D	07 Jul 2025 15:22		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:03 AM
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO112039.D	07 Jul 2025 15:40		YP/AJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PO112040.D	07 Jul 2025 15:58		YP/AJ	Ok
21	I.BLK	I.BLK	PO112041.D	07 Jul 2025 16:16		YP/AJ	Ok,M
22	PB168734BL	PB168734BL	PO112042.D	07 Jul 2025 16:34		YP/AJ	Ok
23	PB168734BS	PB168734BS	PO112043.D	07 Jul 2025 16:51		YP/AJ	Ok
24	Q2505-01	#62825	PO112044.D	07 Jul 2025 17:09	AR1242 Hit , Need cleanup ,	YP/AJ	Not Ok
25	Q2505-02	#62525	PO112045.D	07 Jul 2025 17:26	DCB high in 2nd column , 1242 HIT , need cleanup	YP/AJ	Not Ok
26	Q2505-03	#2008	PO112046.D	07 Jul 2025 17:44	AR1242	YP/AJ	Ok
27	Q2508-01	AUD-25-0105	PO112047.D	07 Jul 2025 18:03		YP/AJ	Ok,M
28	Q2508-02	AUD-25-0106	PO112048.D	07 Jul 2025 18:20	need cleanup	YP/AJ	Not Ok
29	Q2508-03	AUD-25-0107	PO112049.D	07 Jul 2025 18:39		YP/AJ	Ok
30	Q2516-01	CHATHAM HEATER IN	PO112050.D	07 Jul 2025 18:56		YP/AJ	Ok,M
31	Q2509-02	AUD-25-0112	PO112051.D	07 Jul 2025 19:14	need cleanup	YP/AJ	Not Ok
32	AR1660CCC500	AR1660CCC500	PO112052.D	07 Jul 2025 20:15		YP/AJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PO112053.D	07 Jul 2025 20:51		YP/AJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PO112054.D	07 Jul 2025 21:08		YP/AJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PO112055.D	07 Jul 2025 21:25		YP/AJ	Ok,M
36	I.BLK	I.BLK	PO112056.D	07 Jul 2025 21:44		YP/AJ	Ok
37	PB168735BL	PB168735BL	PO112057.D	07 Jul 2025 22:01		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:03 AM
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB168735BS	PB168735BS	PO112058.D	07 Jul 2025 22:20		YP/AJ	Ok,M
39	Q2507-01	SU-04-7.3-2025	PO112059.D	07 Jul 2025 22:38		YP/AJ	Ok,M
40	Q2510-01	#63025-A	PO112060.D	07 Jul 2025 22:57	AR1260 Hit	YP/AJ	Ok,M
41	Q2513-01	HR-2-070325	PO112061.D	07 Jul 2025 23:15		YP/AJ	Ok,M
42	Q2513-03	HR-3-070325	PO112062.D	07 Jul 2025 23:33		YP/AJ	Ok,M
43	Q2513-03MS	HR-3-070325MS	PO112063.D	07 Jul 2025 23:52		YP/AJ	Ok
44	Q2513-03MSD	HR-3-070325MSD	PO112064.D	08 Jul 2025 00:10		YP/AJ	Ok
45	Q2514-01	TP-92	PO112065.D	08 Jul 2025 00:29		YP/AJ	Ok,M
46	Q2514-02	TP-93	PO112066.D	08 Jul 2025 00:46		YP/AJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PO112067.D	08 Jul 2025 03:54		YP/AJ	Ok,M
48	AR1242CCC500	AR1242CCC500	PO112068.D	08 Jul 2025 04:12	DCB Low in 1st column	YP/AJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PO112069.D	08 Jul 2025 04:30		YP/AJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PO112070.D	08 Jul 2025 04:49		YP/AJ	Ok,M
51	AR1268CCC500	AR1268CCC500	PO112071.D	08 Jul 2025 05:07	Wrong vial	YP/AJ	Not Ok
52	I.BLK	I.BLK	PO112072.D	08 Jul 2025 05:26		YP/AJ	Ok,M
53	Q2514-03	TP-94	PO112073.D	08 Jul 2025 05:44		YP/AJ	Ok,M
54	Q2514-04	TP-96	PO112074.D	08 Jul 2025 06:02		YP/AJ	Ok,M
55	Q2514-05	TP-97	PO112075.D	08 Jul 2025 06:21		YP/AJ	Ok,M
56	Q2514-06	TP-103	PO112076.D	08 Jul 2025 06:39		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:03 AM
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	Q2514-07	TP-36	PO112077.D	08 Jul 2025 06:58		YP/AJ	Ok,M
58	Q2514-08	TP-78	PO112078.D	08 Jul 2025 07:16		YP/AJ	Ok,M
59	Q2514-09	TP-81	PO112079.D	08 Jul 2025 07:35		YP/AJ	Ok,M
60	Q2514-10	TP-90	PO112080.D	08 Jul 2025 07:53		YP/AJ	Ok,M
61	Q2515-01	wc-1	PO112081.D	08 Jul 2025 08:11		YP/AJ	Ok,M
62	Q2487-15	G1(0-6)	PO112082.D	08 Jul 2025 08:30	AR1268+1260+1254 Hit , AR1268 CCC missing	YP/AJ	Not Ok
63	AR1660CCC500	AR1660CCC500	PO112083.D	08 Jul 2025 10:09		YP/AJ	Ok,M
64	AR1242CCC500	AR1242CCC500	PO112084.D	08 Jul 2025 10:46		YP/AJ	Not Ok
65	I.BLK	I.BLK	PO112085.D	08 Jul 2025 11:26		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073411.D	01 Jul 2025 13:30		YP\AJ	Ok
2	I.BLK	I.BLK	PP073412.D	01 Jul 2025 13:46		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP073413.D	01 Jul 2025 14:04		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP073414.D	01 Jul 2025 14:21		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP073415.D	01 Jul 2025 14:37		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP073416.D	01 Jul 2025 14:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP073417.D	01 Jul 2025 15:10		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP073418.D	01 Jul 2025 15:26		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP073419.D	01 Jul 2025 15:43		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP073420.D	01 Jul 2025 15:59		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP073421.D	01 Jul 2025 16:16		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP073422.D	01 Jul 2025 16:33		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP073423.D	01 Jul 2025 16:49		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP073424.D	01 Jul 2025 17:05		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP073425.D	01 Jul 2025 17:22		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP073426.D	01 Jul 2025 17:39		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP073427.D	01 Jul 2025 17:55		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP073428.D	01 Jul 2025 18:12		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP073429.D	01 Jul 2025 18:28		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP073430.D	01 Jul 2025 18:45		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP073431.D	01 Jul 2025 19:01		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PP073432.D	01 Jul 2025 19:18		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP073433.D	01 Jul 2025 19:35		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP073434.D	01 Jul 2025 19:51		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP073435.D	01 Jul 2025 20:08		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP073436.D	01 Jul 2025 20:24		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP073437.D	01 Jul 2025 20:41		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP073438.D	01 Jul 2025 20:57		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP073439.D	01 Jul 2025 21:14		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP073440.D	01 Jul 2025 21:30		YPIAJ	Ok,M
31	PP070125ICV500	ICVPP070125	PP073441.D	01 Jul 2025 21:47		YPIAJ	Ok
32	AR1242ICV500	ICVPP070125AR1242	PP073442.D	01 Jul 2025 22:20		YPIAJ	Ok
33	AR1248ICV500	ICVPP070125AR1248	PP073443.D	01 Jul 2025 22:53		YPIAJ	Ok
34	AR1254ICV500	ICVPP070125AR1254	PP073444.D	01 Jul 2025 23:26		YPIAJ	Ok
35	AR1268ICV500	ICVPP070125AR1268	PP073445.D	02 Jul 2025 00:00		YPIAJ	Ok
36	DDT ANALOGUE	DDT ANALOGUE	PP073446.D	02 Jul 2025 07:34		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By	mohammad	Supervise On	7/4/2025 4:31:38 AM
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073447.D	02 Jul 2025 08:57		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP073448.D	02 Jul 2025 10:07		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP073449.D	02 Jul 2025 10:27		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP073450.D	02 Jul 2025 10:44		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP073451.D	02 Jul 2025 11:00		YP\AJ	Ok
6	I.BLK	I.BLK	PP073452.D	02 Jul 2025 11:17		YP\AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PP073453.D	02 Jul 2025 11:34		YP\AJ	Ok
8	PB168692BL	PB168692BL	PP073454.D	02 Jul 2025 15:14		YP\AJ	Ok
9	PB168692BS	PB168692BS	PP073455.D	02 Jul 2025 15:30		YP\AJ	Ok
10	Q2472-01	PN#1	PP073456.D	02 Jul 2025 15:46		YP\AJ	Ok
11	Q2473-01	PIT#1	PP073457.D	02 Jul 2025 16:03		YP\AJ	Ok,M
12	Q2473-02	PIT#2	PP073458.D	02 Jul 2025 16:19		YP\AJ	Ok,M
13	Q2473-03	PIT#3	PP073459.D	02 Jul 2025 16:36		YP\AJ	Ok,M
14	Q2473-04	PIT#4	PP073460.D	02 Jul 2025 16:53		YP\AJ	Ok,M
15	Q2473-04MS	PIT#4MS	PP073461.D	02 Jul 2025 17:09		YP\AJ	Ok,M
16	Q2473-04MSD	PIT#4MSD	PP073462.D	02 Jul 2025 17:25		YP\AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP073463.D	02 Jul 2025 18:47		YP\AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PP073464.D	02 Jul 2025 19:04		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By	mohammad	Supervise On	7/4/2025 4:31:38 AM
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP073465.D	02 Jul 2025 19:20		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP073466.D	02 Jul 2025 19:36		YPIAJ	Ok
21	I.BLK	I.BLK	PP073467.D	02 Jul 2025 19:53		YPIAJ	Ok
22	Q2484-01	TP-58	PP073468.D	02 Jul 2025 20:09		YPIAJ	Ok,M
23	Q2484-02	TP-57	PP073469.D	02 Jul 2025 20:26		YPIAJ	Ok,M
24	Q2484-03	TP-64	PP073470.D	02 Jul 2025 20:42		YPIAJ	Ok
25	Q2484-04	TP-107	PP073471.D	02 Jul 2025 20:58		YPIAJ	Ok
26	Q2484-05	TP-106	PP073472.D	02 Jul 2025 21:15		YPIAJ	Ok
27	Q2484-06	TP-104	PP073473.D	02 Jul 2025 21:31		YPIAJ	Ok
28	Q2458-10	FB-06272025	PP073474.D	02 Jul 2025 21:47		YPIAJ	Ok
29	Q2477-01	50728	PP073475.D	02 Jul 2025 22:04		YPIAJ	Ok,M
30	Q2481-01	CC0627-AL	PP073476.D	02 Jul 2025 22:20		YPIAJ	Ok,M
31	Q2481-04	CC0627-AOXL	PP073477.D	02 Jul 2025 22:36		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP073478.D	02 Jul 2025 23:58		YPIAJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PP073479.D	03 Jul 2025 00:15	failing low in 1st column	YPIAJ	Not Ok
34	AR1248CCC500	AR1248CCC500	PP073480.D	03 Jul 2025 00:31		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PP073481.D	03 Jul 2025 00:47	failing low in 1st column	YPIAJ	Not Ok
36	I.BLK	I.BLK	PP073482.D	03 Jul 2025 01:04		YPIAJ	Ok

M : Manual Integration