

Surrogate Summary

SDG No.: Q2481

Client: Environmental Restoration, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO111586.D	PIBLK-PO111586.D	Tetrachloro-m-xyl	1	20	20.3	102		60	140
		Decachlorobiphen	1	20	20.7	104		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	20.5	103		60	140
I.BLK-PO111985.D	PIBLK-PO111985.D	Tetrachloro-m-xyl	1	20	20.3	102		60	140
		Decachlorobiphen	1	20	17.7	89		60	140
		Tetrachloro-m-xyl	2	20	19.2	96		60	140
		Decachlorobiphen	2	20	20.2	101		60	140
PB168704BL	PB168704BL	Tetrachloro-m-xyl	1	20	19.7	98		30	173
		Decachlorobiphen	1	20	17.4	87		10	173
		Tetrachloro-m-xyl	2	20	18.6	93		30	173
		Decachlorobiphen	2	20	19.8	99		10	173
PB168704BS	PB168704BS	Tetrachloro-m-xyl	1	20	19.7	99		30	173
		Decachlorobiphen	1	20	18.1	91		10	173
		Tetrachloro-m-xyl	2	20	18.4	92		30	173
		Decachlorobiphen	2	20	20.7	104		10	173
PB168704BSD	PB168704BSD	Tetrachloro-m-xyl	1	20	19.6	98		30	173
		Decachlorobiphen	1	20	18.0	90		10	173
		Tetrachloro-m-xyl	2	20	18.3	91		30	173
		Decachlorobiphen	2	20	20.6	103		10	173
I.BLK-PO111999.D	PIBLK-PO111999.D	Tetrachloro-m-xyl	1	20	20.6	103		60	140
		Decachlorobiphen	1	20	17.1	85		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	19.0	95		60	140
I.BLK-PO112005.D	PIBLK-PO112005.D	Tetrachloro-m-xyl	1	20	20.9	105		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	19.2	96		60	140
Q2481-08	CC0627-CLOXAL	Tetrachloro-m-xyl	1	20	32.7	163		30	173
		Decachlorobiphen	1	20	11.5	58		10	173
		Tetrachloro-m-xyl	2	20	20.1	100		30	173
		Decachlorobiphen	2	20	11.3	56		10	173
Q2481-02	CC0627-CLOXPL	Tetrachloro-m-xyl	1	20	17.9	90		30	173
		Decachlorobiphen	1	20	9.90	49		10	173
		Tetrachloro-m-xyl	2	20	14.7	74		30	173
		Decachlorobiphen	2	20	10.5	52		10	173
Q2481-03	CC0625-0XBL	Tetrachloro-m-xyl	1	20	8.30	42		30	173
		Decachlorobiphen	1	20	6.80	34		10	173
		Tetrachloro-m-xyl	2	20	15.1	76		30	173
		Decachlorobiphen	2	20	7.00	35		10	173
Q2481-05	CC0625-NL	Tetrachloro-m-xyl	1	20	24.9	125		30	173

Surrogate Summary

SDG No.: Q2481

Client: Environmental Restoration, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2481-05	CC0625-NL	Decachlorobiphen	1	20	9.90	49		10	173
		Tetrachloro-m-xyl	2	20	23.5	117		30	173
Q2481-06	CC0267-OXPL	Decachlorobiphen	2	20	6.00	30		10	173
		Tetrachloro-m-xyl	1	20	12.1	61		30	173
		Decachlorobiphen	1	20	10.9	54		10	173
		Tetrachloro-m-xyl	2	20	10.9	54		30	173
		Decachlorobiphen	2	20	10.7	53		10	173
Q2481-07	CC0627-OXL	Tetrachloro-m-xyl	1	20	6.90	35		30	173
		Decachlorobiphen	1	20	5.70	29		10	173
		Tetrachloro-m-xyl	2	20	5.00	25	*	30	173
		Decachlorobiphen	2	20	5.70	29		10	173
Q2481-09	CC0627-BL	Tetrachloro-m-xyl	1	20	25.2	126		30	173
		Decachlorobiphen	1	20	21.2	106		10	173
		Tetrachloro-m-xyl	2	20	25.2	126		30	173
		Decachlorobiphen	2	20	18.8	94		10	173
I.BLK-PO112020.D	PIBLK-PO112020.D	Tetrachloro-m-xyl	1	20	20.6	103		60	140
		Decachlorobiphen	1	20	16.9	85		60	140
		Tetrachloro-m-xyl	2	20	18.5	93		60	140
		Decachlorobiphen	2	20	17.8	89		60	140
I.BLK-PO112026.D	PIBLK-PO112026.D	Tetrachloro-m-xyl	1	20	20.1	101		60	140
		Decachlorobiphen	1	20	16.5	83		60	140
		Tetrachloro-m-xyl	2	20	18.1	91		60	140
		Decachlorobiphen	2	20	17.0	85		60	140
Q2481-10	CC0627-SFBL	Tetrachloro-m-xyl	1	20	30.4	152		30	173
		Decachlorobiphen	1	20	25.5	127		10	173
		Tetrachloro-m-xyl	2	20	27.2	136		30	173
		Decachlorobiphen	2	20	27.4	137		10	173
I.BLK-PO112041.D	PIBLK-PO112041.D	Tetrachloro-m-xyl	1	20	19.6	98		60	140
		Decachlorobiphen	1	20	16.9	85		60	140
		Tetrachloro-m-xyl	2	20	17.6	88		60	140
		Decachlorobiphen	2	20	17.9	89		60	140
I.BLK-PP073412.D	PIBLK-PP073412.D	Tetrachloro-m-xyl	1	20	16.6	83		60	140
		Decachlorobiphen	1	20	15.7	78		60	140
		Tetrachloro-m-xyl	2	20	16.5	83		60	140
		Decachlorobiphen	2	20	16.5	82		60	140
I.BLK-PP073467.D	PIBLK-PP073467.D	Tetrachloro-m-xyl	1	20	16.7	84		60	140
		Decachlorobiphen	1	20	15.8	79		60	140
		Tetrachloro-m-xyl	2	20	18.1	90		60	140
		Decachlorobiphen	2	20	18.3	91		60	140
Q2481-01	CC0627-AL	Tetrachloro-m-xyl	1	20	20.4	102		30	173
		Decachlorobiphen	1	20	19.5	97		10	173

Surrogate Summary

SDG No.: Q2481

Client: Environmental Restoration, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2481-01	CC0627-AL	Tetrachloro-m-xyl	2	20	24.8	124		30	173
		Decachlorobiphen	2	20	24.5	123		10	173
Q2481-04	CC0627-AOXL	Tetrachloro-m-xyl	1	20	18.8	94		30	173
		Decachlorobiphen	1	20	12.3	62		10	173
		Tetrachloro-m-xyl	2	20	18.3	92		30	173
		Decachlorobiphen	2	20	19.4	97		10	173
I.BLK-PP073482.D	PIBLK-PP073482.D	Tetrachloro-m-xyl	1	20	15.8	79		60	140
		Decachlorobiphen	1	20	14.4	72		60	140
		Tetrachloro-m-xyl	2	20	17.8	89		60	140
		Decachlorobiphen	2	20	18.7	93		60	140