



Lab Name:	<u>Alliance</u>	Contract:	<u>ENV160</u>	
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2481</u>	
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):	<u>07/29/2025</u>	<u>07/29/2025</u>
		Calibration Times:	<u>16:30</u>	<u>18:07</u>

LAB FILE ID:	RT 200 =	<u>PS031275.D</u>	RT 500 =	<u>PS031276.D</u>
RT 750 =	<u>PS031277.D</u>	RT 1000 =	<u>PS031278.D</u>	RT 1500 = <u>PS031279.D</u>

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ENVI60</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2481</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):	<u>07/29/2025</u> <u>07/29/2025</u>
		Calibration Times:	<u>16:30</u> <u>18:07</u>

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS031275.D</u>	RT 500 = <u>PS031276.D</u>
RT 750 = <u>PS031277.D</u>	RT 1000 = <u>PS031278.D</u>	RT 1500 = <u>PS031279.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.93	9.92	9.92	9.92	9.92	9.92	9.82	10.02
2,4-D	9.03	9.03	9.03	9.02	9.02	9.03	8.93	9.13
2,4-DCAA	7.77	7.77	7.77	7.76	7.76	7.77	7.67	7.87