

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/25/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031260.D **Time Analyzed:** 07:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.337	9.242	9.442	720.060	712.500	1.1
2,4-D	8.452	8.356	8.556	561.210	705.000	-20.4
2,4-DCAA	7.321	7.225	7.425	474.150	750.000	-36.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/25/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031260.D **Time Analyzed:** 07:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.928	9.829	10.029	680.560	712.500	-4.5
2,4-D	9.024	8.924	9.124	294.610	705.000	-58.2
2,4-DCAA	7.766	7.666	7.866	408.080	750.000	-45.6