

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031304.D **Time Analyzed:** 09:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.321	9.227	9.427	724.450	712.500	1.7
2,4-D	8.440	8.346	8.546	700.000	705.000	-0.7
2,4-DCAA	7.311	7.215	7.415	770.290	750.000	2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031304.D **Time Analyzed:** 09:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.920	9.824	10.024	704.490	712.500	-1.1
2,4-D	9.022	8.925	9.125	682.020	705.000	-3.3
2,4-DCAA	7.762	7.666	7.866	724.830	750.000	-3.4